

FACTORY CALIBRATION PROCEDURE

CONTENTS:

General	C-805
Factory circuit specifications	C-807
Factory calibration procedure	C-809
Special test equipment	C-821

INTRODUCTION:

This isn't a field recalibration procedure as is the procedure in your instruction manual. This is a guide in calibrating brand-new instruments, just assembled instruments that have never been turned on before. Therefore it calls out many procedures and adjustments that are rarely required for subsequent recalibration.

Even though we wrote this procedure primarily for our own factory test department, it's valuable to others also if used with some caution:

1. Special test equipment, if mentioned, is not available from Tektronix unless it's listed also in our current catalog. This special equipment is used in our test department to speed calibration. Usually you can either duplicate its function with standard equipment in your facility, devise alternate approaches, or build the special test equipment yourself.
2. Factory circuit specifications are not guaranteed unless they also appear as catalog or instruction manual specifications. Factory circuit specs usually are tighter than advertised specs. This helps insure the instrument will meet or exceed advertised specs after shipment and during subsequent field recalibrations over several years of use. Your instrument may not meet factory circuit specs but should meet catalog or instruction manual specs.
3. Presetting internal adjustments, if mentioned, usually is unnecessary. This is helpful for "first-time" calibration only. If internal adjustments are preset, you'll have to perform a 100% recalibration. So don't preset them unless you're certain a "start-from-scratch" policy is the best.

In this procedure, all front panel controls for the instrument under test are in capital letters (SENSITIVITY) and internal adjustments are capitalized only (Gain Adj).

561

Publication:
061-210
April 1963



For 561 and RM561, all serial numbers, not for 561A or RM561A.



ABBREVIATIONS:

a	amp	mid r	midrange or centered
ac	alternating current	min	minimum
approx	approximately	mm	millimeter
b	base	mpt	metalized, paper tubular (capacitor)
bulb	light, lamp, etc.	msec	millisecond
c	collector	mt	mylar, tubular (capacitor)
ccw	counterclockwise or full counterclockwise	mv	millivolt
cer	ceramic	μ	micro (10^{-6})
cm	centimeter	μ f	microfarad
comp	composition (resistor)	μ h	microhenry
cps	cycles per second	μ sec	microsecond
crt	cathode ray tube	n	nano (10^{-9})
cw	clockwise or full clockwise	nsec	nanosecond
db	decibel	Ω	ohm
dc	direct current	p	pico (10^{-12})
div	division	pbt	paper, "bathtub" (capacitor)
e	emitter	pcc	paper covered can (capacitor)
emc	electrolytic, metal cased (capacitor)	pf	picofarad ($\mu\mu$ f)
emt	electrolytic, metal tubular	piv	peak inverse voltage
fil	filament	pmc	paper, metal cased (capacitor)
freq	frequency	poly	polystyrene
gmV	guaranteed minimum value (capacitor)	pot	potentiometer
gnd	chassis ground	prec	precision (resistor)
h	henry	pt	paper, tubular (capacitor)
hv	high voltage	ptm	paper, tubular molded (capacitor)
inf	infinity	ptp	peak-to-peak
int	internal	sec	second
k	kilo (10^3)	sn	serial number
k	kilohm	term	terminal
m	milli (10^{-3})	tub	tubular (capacitor)
ma	milliamp	unreg	unregulated
max	maximum	v	volt
mc	megacycle	var	variable
meg	megohm	w	watt
mh	millihenry	WW	wire wound
		x-former	transformer

FACTORY CIRCUIT SPECIFICATIONS

SPEC QUALIFICATION

Factory circuit specifications are qualified by the conditions specified in the main body of the calibration procedure. The numbers listed beside the specs correspond to the factory calibration procedure steps where the check or adjustment is made. Instruments may not meet factory circuit specs if calibration or check-out methods and test equipment differ substantially from those in this procedure.

NOT INTENDED FOR INCOMING INSPECTION

We initially calibrate the instrument to factory circuit specifications. These specs usually are tighter than advertised specs, thus helping to insure the instrument will meet or be within advertised specs after shipment and during subsequent recalibrations. Instruments that have left our factory may not meet factory circuit specs but should meet catalog or instruction manual specs.

1. EQUIPMENT REQUIRED

2. PRELIMINARY INSPECTION

- 2d. Crt face plate tilt: 1/32", max.
- 2d. Crt face plate concavity: 1/32", max.
- 2d. Crt face plate convexity: 1/32", max.

3. 561 PRESETS

4. RESISTANCE CHECKS

5. POWER SUPPLIES

5b. Value:

supply	561 tolerance	RM561 tolerance
-100 v	±2 %	±2 %
+125 v	±2 %	±2 %
+300 v	±2.5%	±2 %
- 12 v	±2.5%	±2.5%

The -12 v supply must read between 11.9 v and 12.5 v under min load (150 ma) and when line voltage is varied between 105 and 125 v ac.

The -100 v, +125 v, +300 v supplies must hold regulation when line voltage is varied between 105 and 125 v ac under full load and no load conditions.

5c. Ripple maximums:

supply	561		RM561	
	full load	no load	full load	no load
-100 v	20 mv	5 mv	20 mv	5 mv
+125 v	10 mv	5 mv	10 mv	5 mv
+300 v	80 mv	5 mv	80 mv	10 mv
- 12 v	45 mv	5 mv*	5 mv	5 mv*

*Min loading, 150 ma.

6. HIGH VOLTAGE

6a. Regulation 105 to 125 v ac: ±2%, max.

7. CALIBRATOR

- 7b. Accuracy: ±2%, max.
- 7b. 100 mv: ±2%, max. (RM561: sn 430-up) (561: sn 1590 up)
- 7c. Symmetry: ±20%, max.

8. ALTERNATE SWEEP

9. DUAL-TRACE BLANKING

10. SCALE ILLUM, ALIGN CRT

11. COMPRESSION, EXPANSION

11b. Total: 1 mm, max.

12. CALIBRATOR WAVEFORM, CRT, INT TRIGGER

13. GEOMETRY

- 13a. Horiz geometry: 1 mm, max total.
- 13b. Vert geometry: 1 mm, max.

14. VERT SENSITIVITY, CRT ELECTRICAL CENTER
 - 14a. 561: 22.0 to 24.4 v/div.
RM561: 20.5 to 24.1 v/div.
 - 14b. Crt electrical center: .6 major div, max.

15. VERT COMPENSATION, INT TRIGGER

16. HORIZ SENSITIVITY, CRT ELECTRICAL CENTER
 - 16a. 561: 18.0 to 19.8 v/div.
RM561: 17.7 to 19.5 v/div.
 - 16c. Crt electrical center: .8 major div, max.

17. HORIZ COMPENSATION, INT TRIGGER

18. FREQ RESPONSE
 - 18b. 1 mc, min at -3 db point.

19. LINE TRIGGER

20. CRT GRID

21. THE END.

FACTORY CALIBRATION PROCEDURE

CALIBRATION

NOTES

1. EQUIPMENT REQUIRED

a. Test scope

- 1 530 series Tektronix type scope
- 1 H Tektronix type wide-band, high gain plug-in unit
- 1 10X probe Tektronix type probe

b. Test equipment

- 1 2A60 Tektronix type amplifier plug-in unit
- 1 2B67 Tektronix type time-base plug-in unit
- 1 105 Tektronix type square-wave generator
- 1 180A Tektronix type time-mark generator
- 1 190B Tektronix type constant-amplitude sinewave generator
- 2 TU-4 Tektronix type test loads

c. Test accessories

- 1 011-045 50 Ω termination
- 1 011-032 50 Ω 5:1 attenuator
- 3 012-001 52 Ω 42" cables, uhf connectors

d. Miscellaneous equipment

- 1 630 Triplet meter, 20,000 Ω /v dc
- or 262 Simpson meter, 20,000 Ω /v dc
- 1 -- Variable line voltage source with meter

- 1 special Crt capacitance standardizer
- 1 special Standard calibrator

1b. Test equipment

- (1) TU-40 may be substituted for 105 and 190B.
- (2) TU-50 may be substituted for 105, 180A and 190B.

2. PRELIMINARY INSPECTION

a. Check for unsoldered joints, rosin joints, lead dress and long leads. Check for loose hardware and protruding parts. Check controls for smooth mechanical operation, proper indexing, and knob spacing from front panel.

b. RM561: Check rack securing knobs for correct direction of travel during rotation. Check that fan screen is securely mounted against back panel.

c. Fuse

Fuse-RM561

117 v operation :	159-005	3	a	mdx	slo-blo
234 v operation :	159-041	1.25	a	mdx	slo-blo
F720 (internal) :	159-023	2	a	mdx	slo-blo

Fuse - 561

117 v operation :	159-023	2	a	mdl	slo-blo
234 v operation :	159-019	1	a	mdl	slo-blo
F720 (internal) :	159-023	2	a	mdx	slo-blo

d. Crt face plate tilt

Check crt face plate tilt relative to front panel (keep 561 black light shield in place). Push crt forward to a straight edge firmly against the front panel, across a diameter of the crt. Check gap within phosphor area with rule or crt face tilt checker (special): 1/32", max. If necessary, adjust crt clamp bracket to bring face tilt within specs. Tighten crt clamp.

e. HV shield(s)

Install HV shield(s).

3. 561 PRESETS

a. External controls

FOCUS	ccw
INTENSITY	ccw
ASTIG	mid r
SCALE ILLUM	mid r
POWER	off
CALIBRATOR	OFF

b. Internal adjustments

High Voltage R841	ccw
-100 Volts R616	mid r
Cal Ampl R871	mid r
RM561: Crt Beam Rotator R860	mid r

3b. Presetting internal adjustments

- (1) Presetting internal adjustments is helpful for "first-time" calibration but is usually unnecessary for recalibration. If you preset, you'll have to perform a 100% recalibration. Don't preset them unless you're certain a "start-from-scratch" policy is the best.

c. Leave controls and adjustments, for any step, as they were in the step preceding unless noted otherwise.

4. RESISTANCE CHECKS

a. Check resistance to ground.

supply	approx resistance
117 v ac (power switch on)	inf
-100 v	15 k
- 12 v	135 Ω (common lead to gnd)
+125 v	33 k
+300 v	55 k
+ 6 v unreg	95 Ω (common lead to gnd)
+ 65 v unreg*	50 k
+190 v unreg*	8 k
+380 v unreg*	65 k

*Voltage may be labeled differently on some schematics.

b. Terminal 18-19 continuity

Connect an ohmmeter between terminals 18 on both Amphenol connecting plugs and check continuity. Repeat procedure for terminals number 19. Check for inf resistance between terminals 18 and 19.

5. POWER SUPPLIES

a. Setup

Insert both TU-4's in 561 and connect a 52Ω cable between one TU-4's ripple and dc error connector and the H unit input. Set both TU-4's to 561 indicator; -100 v, full load.

Set test scope to .01 v/div, dc, 5 msec/cm, +line, auto. Connect 561 to variable line voltage, set source to 117 v and turn 561 POWER ON.

b. -100 v Adj R616

Set line voltage to 117 v. Set -100 v Adj R616 for zero error on test scope. Turn TU-4 supply switch to other supplies and check value.

supply	561 tolerance	RM561 tolerance
+125 v	±2 %	±2 %
+300 v	±2.5%	±2 %
- 12 v*	±2.5%	±2.5%

*Nominal value of -12 v supply is 12.21 v. The test loads are designed with this in mind.

The -12 v supply must read between 11.9 v and 12.5 v under min load (150 ma) and when line voltage is varied between 105 and 125 v ac.

The -100 v, +125 v, +300 v supplies must hold regulation when line voltage is varied between 105 and 125 v ac under full load and no load conditions.

c. Check ripple to following maximums.

supply	561		RM561	
	full load	no load	full load	no load
-100 v	20 mv	5 mv	20 mv	5 mv
+125 v	10 mv	5 mv	10 mv	5 mv
+300 v	80 mv	5 mv	80 mv	10 mv
- 12 v	45 mv	5 mv*	5 mv	5 mv*

*Min loading, 150 ma.

d. RM561-T601 check

This step applies to RM561 only. Short terminals A and D of T601 together. Connect ac voltmeter between terminals B and C. The meter should read approx 14 v. Remove meter.

5a. Test-load functions

- (1) Each TU-4 is capable of half loading the supplies when set to full load. To check the supplies under full load conditions set both TU-4's to full load. Conversely, for low load conditions set both TU-4's to no load.
- (2) The push to remove ripple button removes ripple so that a more accurate reading may be obtained.
- (3) The push for gnd ref button will give a zero reference on test scope.

5b. Percent error

- (1) Each div on the test scope (supply level compared to zero reference) represents a 1% error in supply voltage when H is at .01 volts/cm.

5c. Ripple

When measuring ripple, release the push to remove ripple button and read ripple on test scope: 1 cm = 10 mv with .01 v/cm test scope sensitivity setting.

5c. -12 v loading

In order to check the -12 v supply under min load conditions both TU-4's should be set to no load and an external resistor connected between -12 v supply and gnd. The resistor should be picked so as to draw -150 ma.

- e. Line polarity neons, 117 v ac only

Upper neon on and lower neon off indicates correct line polarity. If both neons are on line polarity is reversed.

6. HIGH VOLTAGE

- a. HV Adj R841 regulation: $\pm 2\%$, max

Connect voltmeter from the intensity control end of R852 to gnd. Adjust HV Adj R841 for -3300 v. Vary line voltage from 105 to 125 v ac. HV must regulate within 2% at these line voltage extremes and as INTENSITY is rotated through full range. Remove meter. Leave at 117 v ac.

7. CALIBRATOR

a. Setup

accurate +100 v--52 Ω cable--cal in, standard cal test scope, vert input--52 Ω cable--output, standard cal

561 CAL OUT--52 Ω cable--unk-in, standard cal

Set standard calibrator to 100 v, mixed.

Ground V884, pin 8.

b. Accuracy

±2%, max

Check CALIBRATOR accuracy as follows:

561 CALIBRATOR VOLTS	stand cal volts	test scope sensitivity v/cm, ac	deflection max
100	100	.01	adjust to 0 with Cal Adj R871
50	50	.5	2 cm
20	20	.2	2 cm
10	10	.1	2 cm
5	5	.05	2 cm
2	2	.02	2 cm
1	1	.01	2 cm
.5	.5	.005	2 cm
.2	.2	.005	.8 cm
.1	.1	.005	.4 cm

Check for 100 mv signal into 50 Ω load with CALIBRATOR set at .5 VOLTS (sn 1590 up).

.5	.1	.005	.4 cm
----	----	------	-------

Remove V884, pin 8 gnd strap. Remove calibrator signal.

7a. Known accurate +100 v

- (1) A good source is the test scope:
- (2) Connect the standard calibrator cal in connector to the test scope cal out connector.
- (3) Connect the standard calibrator output to the test scope input.
- (4) Remove the output section of the test scope amplitude calibrator's multivibrator and set the amplitude calibrator control to 100 volts.
- (5) Connect an accurate voltmeter (John Fluke type 803 differential voltmeter) to the cal out connector and adjust the Cal Adj for exactly +100 v on the voltmeter.
- (6) Remove the meter.

7b. Interpreting display

- (1) The test scope display is a 60 cps square wave: one half of each cycle is the standard calibrator dc reference (accurate), the other half is the 561 calibrator dc reference (unknown accuracy).
- (2) The amplitude of the display is the voltage difference between an accurate dc reference and a dc reference of unknown accuracy (the 561 calibrator, whose accuracy we're checking).

(3) $561 \text{ cal } \% \text{ error} = \frac{\text{voltage difference} \times 100}{561 \text{ CALIBRATOR setting}}$

(4) Example:

	case 1	case 2
Standard calibrator:	100 volts	20 volts
561 CALIBRATOR setting:	100 VOLTS	20 VOLTS
Test scope vert sensitivity:	5 v/cm	.1 v/cm
Test scope vert deflection:	1 cm	2 cm

Case 1: $\% \text{ error} = \frac{5 \times 100}{100} = 5\%$

Case 2: $\% \text{ error} = \frac{.2 \times 100}{20} = 1\%$

- (5) In the table of step 7b, we've worked out the setting so that the deflection listed is the maximum allowable to remain within test specifications.

RM561 CALIBRATOR	stand cal	test scope	
		sensitivity v/cm, ac	deflection max
100 VOLTS	100 v	.01	adjust to 0 with Cal Adj R871
10 VOLTS	10 v	.1	2 cm
1 VOLTS	1 v	.01	2 cm
.1 VOLTS	.1 v	.005	.4 cm
10 mVOLTS	10 mv	.005	min*
1 mVOLTS	1 mv	.005	min*

*For greater accuracy use a 2A63 to check these positions.

Check for 100 mv signal into 50 Ω load with CALIBRATOR at 1 VOLTS (sn 430 up).

1 VOLTS	.1	.005	.4 cm
---------	----	------	-------

Remove V884, pin 8 gnd strap. Remove calibrator signal.

c. Cal symmetry ±20%, max

Connect voltmeter between V884, pin 7 and gnd. Note voltage as CALIBRATOR is turned through all of the steps: 40 to 60 v. Remove meter.

8. ALTERNATE SWEEP

a. Alternate sweep

Remove the left TU-4 and insert 2B67. Set TU-4 to dual trace. Set 2B67 for free-run. Check for a displayed two traces. Interchange TU-4 and 2B67 and again check for two traces displayed. Reset TU-4 to normal.

9. DUAL-TRACE BLANKING

a. Setup

105--50 Ω term--52 Ω cable--signal input, TU-4
or TU-50, 105 gen--special atten head--signal input, TU-4

b. Dual-trace blanking

Adjust 105 for 1 cm display of 100 kc signal. Adjust INTENSITY for normal intensity. Connect a jumper from the TU-4 signal input to the TU-4 Z axis input. Upper portion of the crt display must disappear and the lower portion will get brighter. Remove 105 signal and TU-4 jumper.

10. SCALE ILLUM, ALIGN CRT**a. SCALE ILLUM**

Rotate SCALE ILLUM through its range. Check for open spots and for brightest graticule lights when SCALE ILLUM is cw.

b. Align crt

Set 2B67 to 1 msec/cm, free-run. Adjust FOCUS and ASTIG for well-defined trace. Align trace with the center horiz graticule line. Use adjustment knob on 561, Crt Beam Rotator on RM561.

11. COMPRESSION, EXPANSION**a. Setup**

105--50 Ω term--52 Ω cable--signal input TU-4
or TU-50--special atten head--signal input TU-4

b. Compression, expansion total: 1 mm, max

Adjust 105 for exactly 2 cm display at graticule center. Position display to top then to bottom graticule lines. Note compression or expansion; total must not exceed 1 mm. Remove 105 signal.

12. CALIBRATOR WAVEFORM, CRT, INT TRIGGER**a. CALIBRATOR waveform**

Remove TU-4 and insert 2A60. Connect a jumper between CAL OUT and 2A60 input. Check for good waveform throughout all CALIBRATOR positions.

b. Int trigger

Trigger on 4 cm CALIBRATOR waveform. Remove signal, trace should disappear. Exchange plug-ins and recheck trigger. Exchange plug-ins once again.

c. Crt

Check crt for double-peaking, phosphor spots, cathode interface.

13. GEOMETRY

- a. Horiz geometry 1 mm, max total

Set 2B67 to 1 msec/div, free-run. Recheck crt alignment. Position the trace from top to bottom of graticule area and check for horizontal bowing: 1 mm, max total.

- b. Vert geometry 1 mm, max

Connect 180A 1000 μ sec and 100 μ sec markers to 2A60 input. Adjust 2B67 triggering level for stable display and timing, if needed. Adjust 2A60 for markers reaching from bottom to top of graticule. Adjust ASTIG and FOCUS for well-defined trace. Max deviation of vertical trace from vertical graticule line, top to bottom, is 1 mm. Remove 180A signal.

14. VERT SENSITIVITY, CRT ELECTRICAL CENTER

- a. Vert sensitivity 561: 22.0 to 24.4 v/div
RM561: 20.5 to 24.1 v/div

Connect a meter across vertical deflection plates. Set trace to top graticule line, note meter reading. Set trace to bottom graticule line, note meter reading. Divide the algebraic difference in the meter readings by 8. This figure is the vert sensitivity. For the 561 it must be between 22.0 and 24.4 v/div. For the RM561 it must be between 20.5 and 24.1 v/div.

- b. Crt electrical center .6 major div, max

Short the vertical crt plates (use non-magnetic metal). Note distance of trace from crt graticule center: .6 major div, max.

15. VERT COMPENSATION, INT TRIGGER

a. Setup

Install a crt capacitance standardizer (special) in the left compartment, behind the 2A60 plug-in.

Don't use 47 pf input time-constant standardizer.

105--50 Ω term--52 Ω cable--2A60--standardizer
--560
or TU-50, 105 gen--special atten head--2A60--
standardizer--560

2B67 presets: +slope, ac slow coupling, int source.

Set 2A60 to .05 v/div.

b. Crt Plate Compensation C760

Set 105 for 10 kc signal. Adjust 105 and 2A60 for 6 major div of display at normal intensity. Adjust 2B67 triggering level for a stable display. Adjust Crt Plate Compensation C760 and vertical crt plate leads for best square wave.

c. Int trigger

Check internal trigger lead by changing 2A60 input to gnd. The trace should disappear. Reset to dc.

Remove 105 signal.

16. HORIZ SENSITIVITY, ELECTRICAL CENTER

a. Setup

Interchange 2B67 and 2A60. Remove the crt capacitance standardizer at this time.

b. Horiz sensitivity 561: 18.0 to 19.8 v/div
 RM561: 17.7 to 19.5 v/div

Connect a meter across horizontal deflection plates. Set trace to left vert graticule line, note meter reading. Set trace to right line, note meter reading. Divide the algebraic difference in the meter readings by 10. This figure is the horiz sensitivity. For the 561 it must be between 18.0 and 19.8 v/div. For the RM561 it must be between 17.7 and 19.5 v/div.

c. Crt electrical center .8 major div, max

Short the horiz crt plates (use non-magnetic metal). Note distance of the trace from crt graticule center: .8 major div, max.

17. HORIZ COMPENSATION, INT TRIGGER

a. Setup

Install a crt capacitance standardizer (special) in the right compartment, behind the 2A60 plug-in.

Don't use 47 pf input time-constant standardizer.

105--50 Ω term--52 Ω cable--2A60--standardizer
--560
or TU-50, 105 gen--special atten head--2A60--
standardizer--560

2B67 presets: +slope, ac slow coupling, int source.

Keep 2A60 to .05 v/div.

b. Crt Plate Compensation C761

Set 105 for 10kc signal. Adjust 105 and 2A60 for 6 major div of display at normal intensity. Adjust 2B67 triggering level for a stable display. Adjust Crt Plate Compensation C761 for best square wave.

c. Int trigger

Check internal trigger lead by changing 2A60 input to gnd. The trace should disappear. Reset to dc.

Remove 105 signal.

18. FREQ RESPONSE

a. Setup

Remove crt capacitance standardizer and replace 2A60 into right-hand compartment. Set 2A60 to .05 v/div.

190B--50 Ω term--2A60 input

b. Freq response 1 mc, min at -3 db point

Set 190B to 50 kc. Adjust 190B for 4 major div deflection. Set 190B to 1 mc. Note deflection: 2.8 major div, min. Reverse plug-ins and repeat freq response check. Remove 190B signal.

19. LINE TRIGGER**a. Line trigger**

Connect the uhf end of 10X probe to 2A60 input. Connect the probe to the ac line at the fuse holder. Check for proper 2B67 \pm trigger phasing. Reverse plug-ins and repeat check. Remove probe.

20. CRT GRID**a. EXTERNAL INPUT--CRT GRID**

Remove crt grid ground strap (scope rear). Apply 10v of test scope calibrator signal to CRT GRID. With normal trace intensity, trace should be modulated.

21. THE END.

SPECIAL TEST EQUIPMENT

Special test equipment, if mentioned, is not available from Tektronix unless it's listed also in our current catalog. This special equipment is used in our test department to speed calibration.

Usually you can either duplicate its function with standard equipment in your facility, devise alternate approaches, or build the special test equipment yourself.

USE OF STANDARD CALIBRATOR

The standard calibrator, when calibrated, is traceable to NBS and is used to guarantee tolerances of vertical amplifiers and calibrators of Tektronix oscilloscopes.

monitoring the output with an ac coupled scope.

The circuit consists of a chopper and a divider network of 0.1% accurate resistors. The divider network provides a standard voltage output when loaded with 1 meg and when an accurate +100 v is applied to the input. The chopper allows the voltage output of the standard calibrator to switch between a known voltage and an unknown voltage. The difference between these voltages may then be determined by

You must take the hum level of the standard calibrator into account when checking divider accuracy at low levels (.1 v and below). Measure the error introduced by hum level by turning both the standard calibrator and the calibrator of the scope under test to off. Observe the vertical displacement (hum level) and subtract this, when appreciable, from other readings.

Leave the standard calibrator in NORMAL when not in use.

STANDARD CALIBRATOR:

Dwgs 600-B, 7-10-61 (front and rear panels); 601-B, 7-10-61 (schematic); 918-A, (parts).



