

11A71 Amplifier


Serial Number _____

Please check for CHANGE INFORMATION at the rear of this manual.

INSTRUMENT SERIAL NUMBERS

Each instrument has a serial number on a panel insert, tag, or stamped on the chassis. The first number or letter designates the country of manufacture. The last five digits of the serial number are assigned sequentially and are unique to each instrument. Those manufactured in the United States have six unique digits. The country of manufacture is identified as follows:

B000000	Tektronix, Inc., Beaverton, Oregon, USA
100000	Tektronix Guernsey, Ltd., Channel Islands
200000	Tektronix United Kingdom, Ltd., London
300000	Sony/Tektronix, Japan
700000	Tektronix Holland, NV, Heerenveen, The Netherlands

Copyright© Tektronix, Inc., 1987–1988. All rights reserved. Tektronix products are covered by U.S. and foreign patents, issued and pending. TEKTRONIX, INC. SCOPEMOBILE and  are registered trademarks. Printed in U.S.A.

11A71 Features and Functions

Features

- Up to 1000 MHz bandwidth (in 11402)
- Single trace
- Calibrated sensitivities from 10 mV to 1 V/division
- 50 Ω input impedance
- High-resolution, calibrated DC offset (0.25 division/increment, coarse; 0.025 division/increment, fine)

Functions

Signals applied to the CH 1 input connectors can be displayed or removed from the display by pressing the display on/off buttons adjacent to the input connector.

All other 11A71 functions are controlled through the host mainframe. Such mainframe-controlled functions are:

- Sensitivity over a range of 10 mV to 1 V/division in a 1, 2, 5 sequence.
- Vertical Offset over a range of ± 10 divisions.
- Coupling: AC, DC, or Off.

Contents

Section 1—General Information

Contents of this Supplement	1-1
11A71 Service Manual	1-2
Plug-in to Mainframe Compatibility	1-2
Initial Inspection	1-3
Operating Temperature	1-3
Installing and Removing the 11A71	1-4
Instrument Options	1-6
Packaging for Shipment	1-6

Section 2—Operating the 11A71 in 11300-series Mainframes

Display On/Off	2-1
Selecting Coupling	2-1
Other Functions	2-2
Position and Offset	2-2
DC Circuit Loading	2-3
Adding Waveforms	2-4
50 Ω Overload	2-5
Active Probes	2-5
Probe ID	2-5
Front-Panel Error Messages	2-6
GPIB and RS-232-C Commands and Syntax	2-7

Section 3—Operating the 11A71 in 11400-series Mainframes

Display On/Off	3-1
Other Functions	3-1
Vertical Offset	3-2
DC Circuit Loading	3-2
Adding and Subtracting Waveforms	3-4
50 Ω Overload	3-5
Active Probes	3-5
Probe ID	3-5
Front-Panel Error Messages	3-7
GPIB and RS-232-C Commands and Syntax	3-8

Section 4—Specification

Performance Conditions	4-1
Electrical Characteristics	4-1
Environmental Characteristics	4-5
Physical Characteristics	4-6
Recommended Probes	4-6

Appendix A—Glossary

Appendix B—Instrument Documentation

Illustrations

1-1	Installing a plug-in unit in a mainframe oscilloscope	1-5
2-1	Location of AC coupling capacitors in plug-in amplifiers	2-3
3-1	Location of AC coupling capacitors in plug-in amplifiers	3-3
3-2	Sequence of decisions used to select a new waveform in response to probe ID button	3-6
4-1	Dimensions of 11A71	4-7

Tables

2-1	Functions and Where They are Described	2-2
2-2	Channel Parameter Commands (for 11300-series oscilloscopes)	2-7
3-1	Functions and Where They are Described	3-1
3-2	Channel Parameter Commands (for 11400-series oscilloscopes)	3-8
4-1	Electrical Characteristics	4-1
4-2	Environmental Characteristics	4-5
4-3	Physical Characteristics	4-6

Section 1

General Information

This Supplement to the mainframe User Reference manual and the *Functional Test Procedure* manual are standard accessories to the 11A71. For a list of related manuals, see Appendix B.

Contents of this Supplement

This Supplement contains the following four sections:

Section 1—GENERAL INFORMATION, describes mainframe to plug-in unit compatibility, explains how to install and remove the 11A71 from a mainframe, outlines any options available for the 11A71, and gives details about packaging for shipment.

Most 11A71 functions, and their use, are described in the *Operating Information* section of the mainframe User Reference manuals. These functions are those common to all amplifier plug-in units. Examples of common functions are offset, sensitivity, input impedance (where selectable), coupling, and bandwidth limit. Only functions unique to the 11A71-11300-series or 11A71-11400-series combinations are explained in the following sections.

Section 2—OPERATING THE 11A71 IN 11300-SERIES MAINFRAMES, explains how to operate those 11A71 functions not covered in the *11301 and 11302 User Reference Manual*, or the *11301A and 11302A User Reference Manual*.

Section 3—OPERATING THE 11A71 IN 11400-SERIES MAINFRAMES, explains how to operate those 11A71 functions not covered in the *11401 and 11402 User Reference Manual*.

Section 4—SPECIFICATION, gives detailed specifications of all 11A71-mainframe oscilloscope combinations.

11A71 Service Manual

WARNING

The 11A71 Service manual is for use by qualified service personnel only. To avoid personal injury, do not perform any service other than that contained in the Operator's manual unless you are qualified to do so. Refer to the Operators Safety Summary and Service Safety Summary before performing any service.

The 11A71 Service Manual contains the following information:

Section 1—GENERAL INFORMATION.

Section 2—THEORY OF OPERATION.

Section 3—MAINTENANCE AND DIAGNOSTICS.

Section 4—PERFORMANCE VERIFICATION AND ADJUSTMENT.

Section 5—INSTRUMENT OPTIONS.

Section 6—REPLACEABLE ELECTRICAL PARTS.

Section 7—SCHEMATIC DIAGRAMS & CIRCUIT BOARD ILLUSTRATIONS.

Section 8—REPLACEABLE MECHANICAL PARTS.

Plug-in to Mainframe Compatibility

The 11A71 is designed for use in the 11301, 11301A, 11302, 11302A, 11401, 11402 and future 11000-series plug-in mainframes. 11A71 bandwidth varies depending on the host mainframe. Details about bandwidth are included in Section 4, *Specification*, of this supplement, and in the Tektronix Corporate Catalog. Refer to the Tektronix Corporate Catalog for complete compatibility information.

Initial Inspection

This instrument was inspected mechanically and electrically before shipment, and should meet all electrical specifications. First, inspect the 11A71 for physical damage incurred in transit. Secondly, to verify that the instrument is functioning properly, perform the *Functional Test Procedure*, shipped with the instrument. This brief procedure verifies most instrument functions and checks the internal auto-calibration references. If you find instrument damage or deficiency, contact your local Tektronix Field Office or representative.

Operating Temperature

The 11A71 can be operated where the ambient air temperature is between 0° and +50° C and can be stored in ambient temperatures from -40° to +75° C. After storage at temperatures outside the operating limits, allow the chassis to reach operating temperature limits before applying power.

Installing and Removing the 11A71



To avoid instrument damage set the mainframe ON/STANDBY switch to STANDBY before installing or removing the 11A71.

If the green indicator light remains ON, when the STANDBY position is selected, the switch has been internally disabled.

To remove or reinstall plug-in units, use the mainframe PRINCIPAL POWER SWITCH (rear panel) to shut OFF the power. This will prevent damage to either the plug-in or the mainframe.

To once again enable the ON/STANDBY switch, refer the instrument to qualified service personnel.

When installed in any 11401 or 11402 plug-in compartment, or in the 11301, 11301A, 11302 or 11302A left or center plug-in compartment, the 11A71 will provide a conventional display.

When installed in the center or right plug-in compartment of the 11301, 11301A, 11302 or 11302A, the 11A71 will also provide the X (horizontal) part of an X-Y display, or provide a trigger signal for the mainframe time base.

To install the 11A71 in any 11000-series oscilloscope mainframe, set the mainframe ON/STANDBY switch to STANDBY. Align the grooves in the top and bottom of the 11A71 with the guides in the mainframe plug-in compartment, then insert the 11A71 into the mainframe until its front panel is flush with the front panel of the mainframe.

To remove the 11A71 from a mainframe, set the mainframe ON-STANDBY switch to STANDBY. Then pull the release latch (see Figure 1-1) to disengage the unit from the mainframe, and pull the 11A71 straight out of the plug-in compartment.




Figure 1-1. Installing a plug-in unit in a mainframe oscilloscope.

Instrument Options

Option 26 includes one P6231 probe.

Packaging for Shipment

If this instrument is to be shipped by commercial transportation, we recommend that it be packaged in the original manner. The original carton and packaging material should be saved and reused for this purpose.

NOTE

Package and ship plug-in units and mainframes separately.

If the 11A71 is to be shipped to a Tektronix Service Center for service or repair, attach a tag to the instrument. On the tag, include the following information:

- Name and address of the instrument owner;
- Name of a person at your firm who can be contacted about the instrument;
- Complete instrument type and serial number; and
- A description of the service required.

If the original package is not available or is not fit for use, package the 11A71 as follows:

1. Obtain a corrugated cardboard carton with inside dimensions at least six inches greater than the instrument dimensions. Use a carton with a test strength of at least 200 pounds.
2. Fully wrap the 11A71 with anti-static sheeting, or its equivalent, to protect the finish.
3. Cushion the 11A71 on all sides by tightly packing dunnage or urethane foam between the carton and the instrument. Allow three inches of packing on each side.
4. Seal the carton with shipping tape or with industrial staples.
5. Mark the address of the Tektronix Service Center and your return address on the carton in one or more prominent places.

Section 2

Operating the 11A71 in 11300-series Mainframes

Display On/Off

The 11A71 has only one front-panel control—the display on/off button. Pressing a display on/off button will cause channel 1, and its signal, to be displayed or removed from the display (the function toggles). When channel 1 is used for display, its label (CH 1) will be lighted.

Selecting Coupling

To select the desired setting for Coupling, proceed as follows:

- 1a. If no trace is displayed, press the plug-in display on/off button to create a trace.
- 1b. If several traces are displayed, select a one by touching the desired trace description at the top of the CRT.
2. On the 11301 and 11302 oscilloscopes press any VERTICAL button once—OFFSET, SIZE, or POS. On the 11301A and 11302A oscilloscopes press any VERTICAL button twice. This will cause the Control Menu to be displayed.
3. Touch the coupling label. Successive touches will change its state.

Other Functions

Other 11A71 functions are controlled within the 11300-series mainframe, and their operation is described in detail in Section 2, *Operating Information*, of the 11301 and 11302 *User Reference Manual*, and the 11301A and 11302A *User Reference Manual*. Table 2-1 shows where to find information about operating the 11A71.

TABLE 2-1
Functions and Where They are Described

Function	Described Under Heading
Offset, Size (V/div)	Waveform Acquisition: Vertical Menu; or for X-Y Display: Horizontal Menu.
Display Select Ch 1	Waveform Acquisition: Vertical Menu. or for X-Y Display: Horizontal Menu.
Trigger Selection	Trigger Source Major Menu.

Position and Offset

All 11300-series oscilloscopes assign position control to the Left Control knob by pressing the VERTICAL POSition button once. The Vertical Position control moves the trace as a user convenience. For example, when displaying multiple channels it may be desirable to set ground references for each trace on separate graticule lines. Position is a screen-related function; its units are divisions.

The offset function, accessed by pressing the VERTICAL OFFSET button, subtracts a precision voltage from the input signal. Changing offset moves the trace just as does the position control. However, if the deflection factor is subsequently changed, the effect is different.

Changing the sensitivity will increase or decrease the size of the display around some screen level (e.g., two divisions above graticule center) set by the Position control. When using a sensitivity that makes the displayed waveform larger than the screen, the Offset control is used to bring the waveform area of interest to the screen location established by the Position control.

The Position control has a range of plus and minus four divisions from graticule center, while the Offset control has a range of ± 10 divisions.

DC Circuit Loading

AC coupling capacitors are connected differently in the 11A52 and 11A71 than in the 11A32 and 11A34. Figure 2-1 shows this difference.




Figure 2-1. Location of AC coupling capacitors in plug-in amplifiers.

In the 11A52 and 11A71, the coupling capacitor isolates both the input termination and the amplifier from external DC voltages. Avoid inadvertently selecting DC Coupling if your circuit will be damaged by the resulting 50 Ω termination to ground. See Caution below. The coupling capacitor in the 11A32 and 11A34 does not isolate the input termination from external DC voltages.

When AC Coupling is selected, the 11A52 coupling capacitor allows the user to test low-impedance circuits that are elevated in potential by up to 100 volts DC without loading the circuit at DC. Passive attenuating probes such as the P6056 or P6057 eliminate the DC loading and lower the low-frequency -3 dB point.

In the Off coupling mode the AC coupling capacitor precharges in about three seconds to the average DC voltage of the circuit being tested. Always use the Off mode when connecting the 11A52 input to a circuit that may have more than 25 volts (the 11A52 peak input voltage rating) present. This will precharge the coupling capacitor. Wait a few seconds before selecting AC coupling.

The coupling capacitor is discharged upon removal of the external circuit.



Always use caution when working with voltages in excess of 25 volts.

When the input coupling is set to DC a 50 Ω termination resistance is connected directly from the 11A71 input connector to ground. Take care that the circuit connected to the 11A71 input will not be damaged by the 50 Ω load.

Two ways of unintentionally invoking DC Coupling in the 11A71 are:

- 1. Invoking Autoset, because the Autoset process starts by searching for a DC voltage, and*
- 2. Recalling a stored control setting that dictates DC Coupling.*

Switching coupling to DC when more than 25 V is present at the 11A71 input will exceed the peak input voltage specification, and may damage the 11A71 input relay. A damaged relay could cause an error in calibration.

Adding Waveforms

NOTE

Before adding waveforms, check that each channel's display is independently on screen.

The 11300-series oscilloscopes allow the addition of any two channels. That is, one channel may be added to a channel from another plug-in unit. A simple restriction applies. Each channel must be in its linear operating range. This is assured if each channel separately is within the screen area before addition.

Those portions of a trace which are off screen will not be valid when brought back on screen using another channel's input signal or Offset control. This general restriction applies to any dual-channel oscilloscope.

50 Ω Overload

When the input voltage substantially exceeds 5 Vrms, the 11A71 will sense the overload and disconnect the 50 Ω input termination. The mainframe will display the message:

Input channel 1 overload on LEFT/CENTER/RIGHT plug-in

To reconnect the 50 Ω input termination, first correct the overload condition, then select AC or DC from the Coupling menu.

Active Probes

Using an active probe, such as the P6231, eliminates the possibility of AC Coupling. Coupling options available when using such a probe are Off and DC.

Probe ID

The Probe ID section of the Utility Menu is the means of selecting how any 11300-series oscilloscope responds to an ID button-push of recommended probes. All, or some combination of the following operations may be set to start in response to probe ID buttons. To display the Utility Menu, press the front-panel UTILITY button (once on 11301 or 11302 oscilloscopes, twice on 11301A or 11302A oscilloscopes). For details, see the *Probe ID* section of the *Waveform Acquisition* subsection in the mainframe User Reference manual.

Pressing the probe ID button can initiate one or more of the following operations:

1. **Present a new display, or if that channel is already displayed, select the existing trace.** Pressing the ID button of a probe connected to an undisplayed left or center compartment channel will cause that channel to be displayed, unless doing so would exceed the maximum number of traces. Unlike the 11A71 display on/off button, pressing the probe ID button a second time will not remove the display.

Probe ID button-presses for displayed channels will do two things: a) select the trace(s) using this channel, and b) momentarily brighten all traces using this channel.

2. **Stored settings can be sequentially recalled.** Pressing the probe ID button can cause a sequential recall of stored settings. The stored settings feature must be enabled using the Probe ID Utility menu. Settings must be stored as explained under *STORE/RECALL Major Menu* in the *11301 and 11302 User Reference Manual*, or the *11301A and 11302A User Reference Manual*.
3. **The mainframe can Autoset to accommodate the input signal introduced by the probe.** The Autoset feature can be enabled or disabled using the Probe ID Utility menu. Autoset is the automatic setup of vertical deflection factor, triggering, and sweep speed to produce a meaningful display, e.g., two to five divisions of vertical deflection and two to five repetitions of the input signal. For more information, refer to *Autoset* in the *11301 and 11302 User Reference Manual*, or the *11301A and 11302A User Reference Manual*.
4. **The mainframe can automatically measure the selected trace.** Automatic measurements of the selected trace can be initiated by pressing a probe ID button. The Automatic Measurements feature can be enabled or disabled using the Probe ID Utility menu. Such measurements are peak-to-peak, maximum, middle, and minimum voltages; frequency, period, pulse width, and duty cycle. For full information about automatic measurements, see *Measure* in the *11301 and 11302 User Reference Manual*, or the *11301A and 11302A User Reference Manual*.
5. **An interrupt to the GPIB and RS-232-C can be generated.** Pressing a probe ID button will cause the mainframe to produce an SRQ to the GPIB and RS-232-C. For more information, refer to the *GPIB/RS-232-C* section of the *11301 and 11302 User Reference Manual*, or the *11301A and 11302A User Reference Manual*.

Front-Panel Error Messages

- | | |
|-----------------|---|
| Message: | Internal DAC overflow on channel 1 of LEFT, CENTER, or RIGHT plug-in unit. |
| Cause: | Some plug-in unit detected that a requested setting overflowed an internal DAC. Such overflow usually indicates defective hardware. In this situation, the plug-in unit sets the DAC to the limit nearer the requested setting. |
| Message: | Bad Level 2 probe checksum on channel 1 of LEFT, CENTER, or RIGHT plug-in unit. |
| Cause: | Some plug-in unit detected that a Level 2 TEKPROBE had failed or been improperly connected. |

GPIB and RS-232-C Commands and Syntax

The following commands set parameters of a specified channel.

TABLE 2-2
Channel Parameter Commands
(for 11300-series Oscilloscopes)

Header	Link	Argument		
CH<L1, C1, or R1>	COUPLING:	AC DC OFF		
	OFFSET: ¹	<NRx>		
	Volts/div	Offset Range	Resolution via RS-232-C, GPIB, or Numeric Entry	Step Size via front-panel Control knob
	10 mV 20 mV 50 mV 100 mV 200 mV 500 mV 1 V	±100 mV ±200 mV ±500 mV ±1 V ±2 V ±5 V ±10 V	250E-6 500E-6 1.25E-3 2.5E-3 5.0E-3 12.5E-3 25E-3	Coarse: 0.25 div. Fine: 0.025 div.
	BWHI:	<NRx>	HF Limit 400E6, 11301 500E6, 11302	
	IMPEDANCE:	<NRx>	Impedance 50	
	SENSITIVITY: ¹	Sensitivity 10E-3 20E-3 50E-3 100E-3 200E-3 500E-3 1		
	UNITS:	<qstring> (query only) The 11A71 answers a Units query with a units status message, which indicates the units of conversion of a probe connected to its Ch 1 input.		
PROBE:	<qstring> (query only) This query-only link returns a quoted string indicating what type of probe is connected to the input. If a Level 1 TEKPROBE is connected, the query response is "Level 1." If a Level 2 TEKPROBE is connected, the query response is "Level 2/<probe_type>/<serial_number>." When neither a Level 1 nor Level 2 TEKPROBE is connected, the query response is "NONE."			

¹ The numbers listed are those available at the input connectors. Connecting an attenuating probe will change the value by the probe attenuating factor (e.g., a 10X probe will change the value ±10 to ±100).

Legend	L1, C1, or R1>	L, C, and R mean Left, Center, and Right plug-in compartments; 1 means Channel 1.
	COUPLING	Sets the specified channel input coupling.
	OFFSET	Sets the specified channel offset.
	BWHI	Sets the HF Limit (bandwidth) of the specified channel.
	NRx	Numeric representation.
	IMPEDANCE	Sets the input impedance of the specified channel.
	SENSITIVITY	Sets the deflection factor of the specified channel. Sensitivity is a channel-specific command which does not apply to compound waveforms.
	qstring	Quoted string data.

Section 3

Operating the 11A71 in 11400-series Mainframes

Display On/Off

The 11A71 has only one front-panel control—the display on/off button. Pressing the display on/off button will cause channel 1 and its signal to be displayed or removed from the display (the function toggles). When channel 1 is used for display, its label (CH 1) will be lighted.

Other Functions

All other 11A71 functions are controlled within the 11401 or 11402 oscilloscope, and their operation is described in the *11401 and 11402 User Reference manual*. Table 3-1 shows where to find information about operating the 11A71.

TABLE 3-1
Functions and Where They are Described

Function	Described under Heading ¹
Position	Vertical Size and Position.
Vertical Size (V/div)	Vertical Size and Position.
Coupling (AC, DC, Off)	Selecting Channel Coupling, under Waveform Control.
Select 11A71 Channel	Waveform Control (an alternative to pressing the 11A71 display on/off button).
Offset	Vertical Position.

¹Under Waveform Control in Section 2, *Operating Information*.

Vertical Offset

The 11401 and 11402 oscilloscopes use offset to position the trace on the screen. In YT (signal vs. time) mode offset determines vertical position, but when a plug-in unit provides the X component of an XY display, offset controls horizontal position.

Offset is an input-signal-related-control, and its units are those of the input signal—usually volts.

Offset subtracts a precision voltage from the input signal. To access the offset function, touch the double-ended arrow at the left edge of the screen. Control of offset is then assigned to the Lower Control knob.

In the 11401 and 11402, changing the vertical deflection factor magnifies or compresses the displayed signal about screen center. Therefore, to examine a portion of the input signal in greater detail, use Vert Offset (the Lower Control knob) to position the area of interest to screen center. Next, use the Vert Size (the Upper Control knob) to adjust the display to the size needed for detailed examination.

DC Circuit Loading

AC coupling capacitors are connected differently in the 11A52 and 11A71 than in the 11A32 and 11A34. Figure 3-1 shows this difference.

In the 11A52 and 11A71, the coupling capacitor isolates both the input termination and the amplifier from external DC voltages. Avoid inadvertently selecting DC coupling if your circuit will be damaged by the resulting 50 Ω termination to ground. See Caution below. The coupling capacitor in the 11A32 and 11A34 does not isolate the input termination from external DC voltages.

When AC Coupling is selected, the 11A52 coupling capacitor allows the user to test low-impedance circuits that are elevated in potential by up to 100 volts DC without loading the circuit at DC. Passive attenuating probes such as the P6056, or P6057 eliminate the DC loading and lower the low-frequency -3 dB point.

In the Off Coupling mode the AC coupling capacitor precharges in about three seconds to the average DC voltage of the circuit being tested. Always use the Off mode when connecting the 11A52 input to a circuit that may have more than 25 volts (the 11A52 peak input voltage rating) present. This will precharge the coupling capacitor. Wait a few seconds before selecting AC Coupling.

The coupling capacitor is discharged upon removal of the external circuit.



Always use caution when working with voltages in excess of 25 volts.

When the input coupling is set to DC, a 50 Ω termination resistance is connected directly from the 11A71 input connector to ground. Take care that the circuit connected to the 11A71 input will not be damaged by the 50 Ω load.

Two ways of unintentionally invoking DC coupling in the 11A71 are:

- 1. Invoking Autoset, because the Autoset process starts by searching for a DC voltage, and*
- 2. Recalling a stored control setting that dictates DC Coupling.*

Switching coupling to DC when more than 25 V is present at the 11A71 input will exceed the peak input voltage specification, and may damage the 11A71 input relay. A damaged relay could cause an error in calibration.




Figure 3-1. Location of AC coupling capacitors in plug-in amplifiers.

Adding and Subtracting Waveforms

The 11401 and 11402 mainframes do not allow the addition or subtraction of live waveforms without restriction. The reason for this lies in the manner in which samples are taken and used by the mainframe. Each plug-in compartment has an associated sampler in the 11401 or 11402. Samples are taken simultaneously by each sampler. A channel from one plug-in unit may be added to or subtracted from that of another plug-in unit accurately by virtue of this simultaneity. Thus noise and other signals unrelated to the trigger will be canceled when channels in different plug-in units are subtracted.

Signals from different channels within a plug-in unit are sampled at different times. Signals that are entirely repetitive and related to the trigger signal may be successfully added or subtracted in spite of this. However, signals that are not repetitive or related to the trigger will not be sampled at the same time, and may not be displayed correctly. Therefore it is preferable that signals to be added or subtracted be selected from different plug-in compartments.

The best common-mode rejection is obtained with a differential amplifier plug-in unit because its hardware subtracts the signals at its two inputs, passing only the result to the mainframe.

Even when using channels from different plug-in compartments, the following two problems limit the accuracy of sums or differences:

1. Each waveform in memory is constantly being updated, and therefore consists of some points that are older than others. Although samples may be taken simultaneously, there is a chance that a correct sample may be replaced in memory with a newer value before the waveform is processed and displayed.
2. Delay differences between channels of different plug-in units due to probe cable length differences are normally removed by deskewing the probes and the associated amplifier channel. Samples are still taken simultaneously, but the display is manipulated to remove time differences. Signals correlated to the trigger signal are correctly restored to their proper time relationship, but there is no way for signals unrelated to the trigger to be corrected with this deskewing technique. High-frequency rejection of signals unrelated to the trigger is not improved by deskewing.

The solution to both problems is to invoke averaging. The averaging process suppresses all signals and noise unrelated to the trigger.

50 Ω Overload

When the input voltage substantially exceeds 5 Vrms, the 11A71 will sense the overload and disconnect the 50 Ω input termination. The mainframe will display the message:

Input channel 1 overload on LEFT/CENTER/RIGHT plug-in

To reconnect the 50 Ω input termination, first correct the overload condition, then select AC or DC from the Coupling menu.

Active Probes

Using an active probe, such as the P6231, eliminates the possibility of AC Coupling. Coupling options available when using such a probe are Off and DC.

Probe ID

The **Probe ID function** is the means of selecting how the 11401 or 11402 responds to an ID button-push of recommended probes. (The **Probe ID Function** is part of the **Probes** pop-up menu, which is selected from the Utilities major menu.) All, or some combination of the following operations may be set to start in response to probe ID buttons. To display the Utility major menu, press the front-panel UTILITY button. For details, see the *Probe Calibration and ID Function* section of the Utilities subsection in the *11401 and 11402 User Reference Manual*.

Pressing the probe ID button can initiate one of the following operations:

1. If the **Probe ID function** (in the **Probes** pop-up menu of the Utility major menu) is set to **Sequence Settings**, the next front-panel setting will be recalled. If the **Probe ID Function** is not set to **Sequence Settings**, then operation 2 or 3, which involve a new waveform and waveform selection, will occur.

If the Sequence Settings option is enabled, front-panel settings can be sequentially recalled via the Probe ID function. The settings must be stored as explained under *Storing and Recalling Waveforms* in the *11401 and 11402 User Reference Manual*.

2. If the **Probe ID Function** (in the **Probes** pop-up menu of the Utility major menu) is set to **Waveform Select/New Wfm**, a waveform will be selected. Figure 3-2 shows the sequence of decisions used to select a new waveform.

3. If the Probe ID Function is set to **Wfm Select/New Wfm & AutoSet**, then the decisions shown in Figure 3-2 will be made, and the resulting selected waveform is autoset.

Events 1, 2, and 3 are mutually exclusive. Event 4 may be combined with any one of them.

4. If the IDPROBE SRQMASK is enabled, an appropriate event will be returned to the GPIB and RS-232-C ports.




Figure 3-2. Sequence of decisions used to select a new waveform in response to probe ID button.

Front-Panel Error Messages

Message: Internal DAC overflow on channel 1 of LEFT, CENTER, or RIGHT plug-in unit.

Cause: Some plug-in unit detected that a requested setting overflowed an internal DAC. Such overflow usually indicates defective hardware. In this situation, the plug-in unit sets the DAC to the limit nearer the requested setting.

Message: Bad Level 2 probe checksum on channel 1 of LEFT, CENTER, or RIGHT plug-in unit.

Cause: Some plug-in unit detected that a Level 2 TEKPROBE had failed or been improperly connected.

GPIB and RS-232-C Commands and Syntax

The following commands set parameters of a specified channel.

Table 3-2
Channel Parameter Commands
(for 11400-series Oscilloscopes)

Header	Link	Argument		
CH<L1, C2, or R1,>	COUPLING:	AC DC OFF		
	OFFSET: ¹	<NRx>		
	Volts/div	Offset Range	Resolution via RS-232-C, GPIB, or Numeric Entry	Step Size via front-panel Control knob
	10 mV 20 mV 50 mV 100 mV 200 mV 500 mV 1 V	±100 mV ±200 mV ±500 mV ±1 V ±2 V ±5 V ±10 V	250E-6 500E-6 1.25E-3 2.5E-3 5.0E-3 12.5E-3 25E-3	Coarse: 0.25 div. Fine: 0.025 div.
	BW:	<NRx>		Bandwidth 500E6, 11401 1000E6, 11402
	IMPEDANCE:	<NRx>		Impedance 50
	SENSITIVITY: ¹	Sensitivity 10E-3 20E-3 50E-3 100E-3 200E-3 500E-3 1		
	UNITS:	<qstring> (query only) The 11A71 answers a Units query with a units status message, which indicates the units of conversion of a probe connected to its Ch N input.		
PROBE:	<qstring> (query only) This query-only link returns a quoted string indicating what type of probe is connected to the input. If a Level 1 TEKPROBE is connected, the query response is "Level 1." If a Level 2 TEKPROBE is connected, the query response is "Level 2/<probe_type>/ <serial_number>." When neither Level 1 nor Level 2 TEKPROBE is connected, the query response is "NONE."			

¹The numbers listed are those available at the input connectors. Connecting an attenuating probe will change the value by the probe attenuating factor (e.g., a 10X probe will change the value ±10 to ±100).

Legend	CH<L1, C1 & R1> L, C, and R mean Left, Center, and Right plug-in compartments; 1 means Channel 1.
COUPLING	Sets the specified channel input coupling.
OFFSET	Sets the specified channel offset.
BW	Sets the HF Limit (bandwidth) of the specified channel.
NRx	Numeric representation.
IMPEDANCE	Sets the input impedance of the specified channel.
SENSITIVITY	Sets the deflection factor of the specified channel. Sensitivity is a channel-specific command which does not apply to compound waveforms.
qstring	Quoted string data.

Section 4

Specification

Performance Conditions

The specifications that follow apply when the instrument is in the condition of Enhanced Accuracy. Enhanced Accuracy is initiated by pushing the E A button on the specific host mainframe after the system has reached thermal equilibrium, which requires 20 minutes warmup. Enhanced Accuracy is indicated on the CRT display and remains in effect as long as the mainframe internal temperature change is less than 5° C from the temperature at which the calibration was performed. When the 5° C change does occur the accuracy condition becomes Not-Enhanced. In the Not-Enhanced condition those Characteristics that are temperature sensitive may not remain within the limits of these specifications. The instrument specifications are valid at an ambient temperature of 0° to 50° C, unless otherwise stated.

TABLE 4-1
Electrical Characteristics

Circuit	Characteristic	Performance Requirement			
Display Deflection Factor (Sensitivity)	Calibrated Range	10 mV to 1 V/div. in seven steps in a 1, 2, 5 sequence			
	Enhanced DC Accuracy¹ , of 11A71 in 11300-series Mainframes, with on-screen cursors readings with Position set to 0.00.				
	Without Probes				
	Volts/Div	ΔV DC Acc.	DC Balance	DC Offset Acc.	
	10 mV to 1 V	±(0.9% + 0.05 div)	±(0.2 div)	±(0.4% + 0.01 mV)	
	With P6231 Probe calibrated from 11300-series Calibrator output				
Volts/Div	ΔV DC Acc.	DC Balance	DC Offset Acc.		
100 mV to 1 V	±(1.6% + 0.05 div)	±(0.2 div)	±(0.15% + 2 mV)		
Probe Tip TC term	100 ppm/° C				

¹ For absolute DC accuracy of single-point measurements using Offset, and the DC Offset Accuracy, DC Balance and ΔV DC Accuracy terms. Apply the ΔV DC Accuracy only to the difference between the Vertical Position setting and the measurement point.

TABLE 4-1 (cont.)
Electrical Characteristics

Circuit	Characteristic	Performance Requirement			
Display Deflection Factor (Sensitivity) (Cont)	Enhanced DC Accuracy ¹ , of 11A71 in 11401/11402 Mainframes.				
	Without Probes				
	Volts/Div	ΔV DC Acc.	DC Balance	DC Offset Acc.	
	10 mV to 1 V	±(0.7% + 0.03 div)	±(0.2 div)	±(0.20% + 0.01 div)	
	With P6231 Probe calibrated from 11401/11402 Calibrator output				
	Volts/Div	ΔV DC Acc.	DC Balance	DC Offset Acc.	
	100 mV to 1 V	±(1.4% + 0.03 div)	±(0.2 div)	±(0.15% + 2 mV)	
	Probe Tip TC term	100 ppm/° C			
	Coarse Resolution		1, 2, 5 sequence.		
	Display Offset	Accuracy	See <i>Deflection Factor Accuracy</i> , which precedes this characteristic.		
Range, Resolution		±10 div. Coarse and fine resolution are 0.25 and 0.025 div., respectively.			
Display Frequency Response	High Frequency Limit, (–3 dB point) and Calculated Rise Time (20° to + 30° C)	Type of Mainframe			
		Volts/Division	11301(A)	11302(A)	11401
	All	400 MHz 0.9 ns	500 MHz 0.7 ns	500 MHz 0.7 ns	1000 MHz 0.35 ns
	Low Frequency –3 dB point, AC coupled Display and Trigger signals.	1 kHz maximum from 50 Ω source			

¹ For absolute DC accuracy of single-point measurements using Offset, and the DC Offset Accuracy, DC Balance and ΔV DC Accuracy terms. Apply the ΔV DC Accuracy only to the difference between the Vertical Position setting and the measurement point.

TABLE 4-1 (cont.)
Electrical Characteristics


Circuit	Characteristic	Performance Requirement
Input	Maximum Input Voltage, AC or DC Coupled	5 Vrms (0.5 W) or 0.25 watt-second pulses not exceeding 25 V peak.
	Maximum DC Input Voltage, AC Coupled	≤100 V (DC+peak AC).
	<div style="text-align: center;">  <p>CAUTION</p> <p><i>Signals of more than 25 V peak amplitude must be connected with the input coupling set to OFF so that the input coupling capacitor is precharged.</i></p> </div>	
	Input Disconnect Threshold	+5 VDC ±1 V. –5 VDC ±1 V. 5 Vrms up to 100 MHz.
	Input Protection Disconnect Time	30 ms or less for 25 VDC or either polarity. 40 ms to 150 ms for 12.5 VDC of either polarity.
	Power-Down Condition	50 Ω terminations disconnect when 11A71 is powered down.
	Impedance (DC coupled)	50 Ω within 2%, VSWR < 1.45:1 at 10 mV/div., DC to 1 GHz; VSWR < 1.25:1 at 20 mV to 1 V/div., DC to 1 GHz.
	Impedance (AC coupled)	50 Ω within 2% in series with at least 2.2 μF. VSWR ≤ 1.45:1 at 10 mV/div., DC to 1 GHz; VSWR ≤ 1.25:1 at 20 mV to 1 V/div., DC to 1 GHz.
	Reflection Coefficient	±10% or less, total not to exceed 20% pk-pk in a 1 GHz time-domain reflectometer system.
	Input Bias Current (+20° to +30° C)	50 μA maximum.

TABLE 4-1 (cont.)
Electrical Characteristics

Circuit	Characteristic	Performance Requirement
Miscellaneous	Probe Compatibility	The 11A71 is compatible with Level 1 and 2 TEKPROBES. ¹

¹TEKPROBE is Tektronix' name for the interface used with probes designed for the 11000-series of oscilloscopes and plug-in units. TEKPROBES have output connectors with one or more spring loaded coding pins. Two categories of TEKPROBES are:

Level 1

A level 1 probe uses analog encoding to indicate the probe's scale factor to the plug-in unit.

Level 2

A level 2 probe uses an EEPROM to store data about the the probe's transfer units, scale factor, and output voltage scale factor. Such data are serially encoded, then stored in the EEPROM. The probe data is intended to be read once at instrument power-up or when the probe is first connected to a plug-in unit (that is, at probe power-up).

TABLE 4-2
Environmental Characteristics

Characteristic	Information
Ambient temperature (External to main frame)	
Operating within specs.	0° to +50° C., mainframe ambient.
Nonoperating	–40° to +75° C.
Humidity, Operating and Nonoperating	Five days, per MIL-T-28800C. Type III, class 5 as described in 3.9.2.2 and 4.5.5.1.2.2.
Altitude	
Operating	To 4,570 m (15,000 ft.).
Nonoperating	To 15,200 m (50,000 ft.).
Vibration	
Operating, installed on Flexible Extender	MIL-T-28800C, Sec. 4.5.5.3.1, type III, class 5.
Shock, Nonoperating (not installed in mainframe)	MIL-T-28800C, Sec. 4.5.5.4.1, type III, class 5.
Bench Handling (operating and nonoperating)	MIL-T-28800C, Sec. 4.5.5.4.3, type III, class 5.
Packaged Product Vibration and Shock	
Vibration and Bounce of Packaged Product	Meets ASTM D999-75, Method A (NSTA Project 1A-B-1).
Drop of Packaged Product	Meets ASTM D775-61, Paragraph 5 (NSTA Project 1A-B-2).
Electromagnetic Compatibility	MIL. STD. 461B. FCC Part 15, Subpart J, Class A. VDE 0871/6.78, Class B.

Table 4-3
Physical Characteristics

Characteristic	Information
Weight (max)	2 lb. 0 oz. (908 grams).
Weight of Packaged Product (max)	4 lb. 14 oz. (2213 grams).
Dimensions (max)	Refer to Figure 4-1.

Recommended Probes

Tektronix recommends the P6231 Probe for use with the 11A71. It has the following characteristics:

P6231 Low Impedance Subminiature 10X Active Probe. The P6231 has a bandwidth of 1.5 GHz, input impedance of 450 Ω , and DC offset of ± 5 V controlled by the 11A71. The P6231 is a Level 2 probe.




Figure 4-1. Dimensions of 11A71.

Appendix A—Glossary

Deflection factor	The ratio of input signal to response in the 11000 system. The reciprocal of sensitivity.
Display on/off button	The front-panel, plug-in button that designates a channel for display, or removes a channel from the display.
Mainframe	Any 11000-series oscilloscope exclusive of plug-in units.
Offset	A control that allows the user to subtract a precision voltage from the input signal to vary the position of the displayed signal.
ON/STANDBY	The front-panel power control on 11000-series mainframes. When set to ON, applies operating power to all circuits. When set to STANDBY, the mainframe DC power supply supplies power to the counter-timer crystal oven (Option 1T).
Overdrive	A condition in which amplifiers are driven into a non-linear operating range. Typically many divisions off screen.
Overload	The condition existing when a potentially damaging voltage is applied to the input connector.
Toggle	To switch alternately between two functions (e.g., on and off).

Appendix B—Instrument Documentation

- *Introducing the 11301 and 11302 Programmable Oscilloscopes.* This book helps the new user learn the concepts and operation of the 11301 and 11302 programmable oscilloscopes.

or

Introducing the 11301A and 11302A Counter/Timer Oscilloscopes. This book helps the new user learn the concepts and operation of the 11301A and 11302A counter/timer oscilloscopes.

- *Introducing the 11401 and 11402 Digitizing Oscilloscopes.* Intended for new users, this book explains the concepts and operation of the 11401 and 11402 digitizing oscilloscopes.
- The *11301 and 11302 User Reference Manual* contains detailed information about measurements the instruments can make, and instructions about how to set the oscilloscope to perform such measurements. Also, installation instructions, specifications, description of options, and details about using the 11301 and 11302 with the GPIB and RS-232-C interfaces are included here.

or

The *11301A and 11302A User Reference Manual* contains detailed information about measurements the instruments can make, and instructions about how to set the oscilloscope to perform such measurements. Also, installation instructions, specifications, description of options, and details about using the 11301A and 11302A with the GPIB and RS-232-C interfaces are included here.

- The *11401 and 11402 User Reference Manual* contains detailed information about measurements the instruments can make, and instructions about how to set the oscilloscope to perform such measurements. Also, installation instructions, specifications, description of options, and details about using the 11401 and 11402 with the GPIB and RS-232-C interfaces are included here.
- The *11301 and 11302 Pocket Reference* is a short-form reference. The Pocket Reference gives menus and a brief description of functions selected. External interface commands are listed with their basic syntax, and a list of error and warning messages is included.

or

The *11301A and 11302A Pocket Reference* is a short-form reference. The Pocket Reference gives menus and a brief description of functions selected. External interface commands are listed with their basic syntax, and a list of error and warning messages is included.

- The *11A71 Service Manual* includes information needed to check and adjust, perform diagnostic troubleshooting, repair, and otherwise service your instrument.
- The *11A71 Functional Test Procedure* manual verifies that the instrument is functioning properly.

MANUAL CHANGE INFORMATION

At Tektronix, we continually strive to keep up with latest electronic developments by adding circuit and component improvements to our instruments as soon as they are developed and tested.

Sometimes, due to printing and shipping requirements, we can't get these changes immediately into printed manuals. Hence, your manual may contain new change information on following pages.

A single change may affect several sections. Since the change information sheets are carried in the manual until all changes are permanently entered, some duplication may occur. If no such change pages appear following this page, your manual is correct as printed.