



ADAPTER, LED TEST
(Part No. 014-0047-00)



With reference plane of LED against LED holder, emitting plane of LED should be set to 3.60 inches from probe filter surface.



NOTE: A reading of 0.01 foot candle is equivalent to 1 millicandella.

Fig. & Index No.	Tektronix Part No.	Serial/Model No. Eff Disc	Q	t					Description	
				y	1	2	3	4		5
	014-0047-00 ✓		1							ADAPTER, LED test
	-----		-							adapter includes:
1	214-1861-00 N ✓		1							GASKET, filter
2	220-0685-00 N ✓	rown ✓	1							NUT, knurled, 1.125-18 x 1.40 inch OD
3	380-0319-01 N ✓		1							HOUSING, filter
4	378-0814-00 N ✓		4							BAFFLE, light
5	166-0563-00 N ✓		1							TUBE, LED
6	352-0358-00 N ✓		1							HOLDER, LED, 0.21 inch diameter
	352-0361-00 N ✓		1							HOLDER, LED, 0.085 inch diameter
	352-0367-00 N ✓		1							HOLDER, LED, 0.125 inch diameter

DATA SHEET	
NO.	062-1545-00 ✓
DATE	DEC 1972
<small> COPYRIGHT © 1972 TEKTRONIX INC. ALL RIGHTS RESERVED </small>	

014-0047-00