

T Unit Calibration Outline

- | | |
|----------------------------------|---|
| 1. Position Range Adjust (Horiz) | Horiz. Pos. at 12:00
Adjust to start trace
at left age of graticule |
| 2. | |
| 2. Trigger | 200mv external |
| 3. Mag Gain | .1 & 1ms |
| 4. Mag Centering (Front Panel) | |
| 5. Sweep Cal | .1 & 1ms |
| 6. Fast & Slow Speed Timing | |
| 7. HF Sync | 20mc 2cm of signal |

For excessive jitter -- Look for dirty capacitor grounding bands
 Try disconnecting electrolytic cap leads

