

# 2236A

## PORTABLE OSCILLOSCOPE INSTRUCTION


**WARNING**

THE FOLLOWING SERVICING INSTRUCTIONS ARE FOR USE BY QUALIFIED PERSONNEL ONLY. TO AVOID PERSONAL INJURY, DO NOT PERFORM ANY SERVICING OTHER THAN THAT CONTAINED IN OPERATING INSTRUCTIONS UNLESS YOU ARE QUALIFIED TO DO SO. REFER TO OPERATORS SAFETY SUMMARY AND SERVICE SAFETY SUMMARY PRIOR TO PERFORMING ANY SERVICE.

*Please Check for  
**CHANGE INFORMATION** at  
the Rear of This Manual*

Copyright © 1989 Tektronix, Inc. All rights reserved. Contents of this publication may not be reproduced in any form without the written permission of Tektronix, Inc.

Products of Tektronix, Inc. and its subsidiaries are covered by U.S. and foreign patents issued and pending.

TEKTRONIX, TEK, SCOPE-MOBILE, and  are registered trademarks of Tektronix, Inc.

Printed in U.S.A. Specification and price change privileges are reserved.

## INSTRUMENT SERIAL NUMBERS

Each instrument has a serial number on a panel insert, tag, or stamped on the chassis. The first two digits designate the country of manufacture. The last five digits of the serial number are unique to each instrument. Those manufactured in the United States have six unique digits. The country of manufacture is identified as follows:

- B000000 Tektronix, Inc., Beaverton, Oregon, U.S.A.
- HK00000 Tektronix, Inc., Hong Kong
- G100000 Tektronix Guernsey, Ltd., Channel Islands
- E200000 Tektronix United Kingdom, Ltd., Marlow
- J300000 Sony/Tektronix, Japan
- H700000 Tektronix Holland, NV, Heerenveen,  
The Netherlands

# OPERATORS SAFETY SUMMARY

The safety information in this summary is for operating personnel. Warnings and cautions will also be found throughout the manual where they apply.

## Terms in This Manual

CAUTION statements identify conditions or practices that could result in damage to the equipment or other property.

WARNING statements identify conditions or practices that could result in personal injury or loss of life.

## Terms as Marked on Equipment

CAUTION indicates a personal injury hazard not immediately accessible as one reads the markings, or a hazard to property, including the equipment itself.

DANGER indicates a personal injury hazard immediately accessible as one reads the marking.

## Symbols in This Manual



This symbol indicates where applicable cautionary or other information is to be found. For maximum input voltage see Table 1-1.

## Symbols as Marked on Equipment



DANGER – High voltage.



Protective ground (earth) terminal.



ATTENTION – Refer to manual.

## Power Source

This product is intended to operate from a power source that does not apply more than 250 V rms between the supply conductors or between either supply conductor and ground. A protective ground connection, by way of the grounding conductor in the power cord, is essential for safe operation.

## Grounding the Product

This product is grounded through the grounding conductor of the power cord. To avoid electrical shock, plug the power cord into a properly wired receptacle before making any connections to the product input or output terminals. A protective ground connection, by way of the grounding conductor in the power cord, is essential for safe operation.

## Danger Arising From Loss of Ground

Upon loss of the protective-ground connection, all accessible conductive parts, including knobs and controls that may appear to be insulating, can render an electric shock.

## Use the Proper Power Cord

Use only the power cord and connector specified for your product.

Use only a power cord that is in good condition.

For detailed information on power cords and connectors, see Figure 2-2.

## Use the Proper Fuse

To avoid fire hazard, use only a fuse of the correct type, voltage rating and current rating as specified in the parts list for your product.


## Do Not Operate in an Explosive Atmosphere

To avoid explosion, do not operate this instrument in an explosive atmosphere.

## Do Not Remove Covers or Panels

To avoid personal injury, do not remove the product covers or panels. Do not operate the product without the covers and panels properly installed.





The 2236A Oscilloscope.

7685-50



# SPECIFICATION

## INTRODUCTION

The TEKTRONIX 2236A oscilloscope is a lightweight 100-MHz instrument with a Counter, Timer, and Multimeter (CTM). Measurement values are displayed on the Vacuum Fluorescent nine-digit readout.

The dual vertical channel system supplies calibrated deflection factors from 2 mV per division to 5 V per division. Trigger circuits enable stable triggering over the full bandwidth of the vertical deflection system. The horizontal system provides calibrated sweep speeds from 0.5 s per division to 50 ns per division, along with delayed-sweep features. A X10 magnifier circuit extends the maximum sweep speed to 5 ns per division when the A and B SEC/DIV switch is set to 0.05  $\mu$ s per division.

The Counter Timer section of the CTM measures frequency, period, width, and totalizes, and provides a digital readout of oscilloscope delay time and delta time. It also provides frequency, period, and width measurements and totalizes on portions of a waveform selected by the intensified zone (gated measurements). Counting and timing measurements are made through the A and B Trigger system. All Counter Timer measurement values are displayed on the readout.

The Multimeter section of the CTM measures voltages and resistance through the floating input connectors located on the right side of the instrument. Voltages can also be measured through the CH 1 OR X & DMM connector. Temperature measurements can be made by connecting the optional temperature probe to the Multimeter input connectors. The Multimeter measurement values are displayed with a 3 3/4-digit (5000 count) format, on the digital readout.

## ACCESSORIES

The instrument is shipped with the following standard accessories:

- 1 Instruction manual
- 2 Multimeter test lead clips
- 2 Multimeter test leads
- 2 10X probes with accessories
- 1 Power cord
- 1 Power cord clamp
- 1 Flat washer
- 1 Self-tapping screw
- 1 Fuse

For part numbers and information about both standard and optional accessories, refer to the Accessories page at the back of this manual. Your Tektronix representative, local Tektronix Field Office, or Tektronix product catalog can also provide accessories information.

## PERFORMANCE CONDITIONS

The following electrical characteristics (Table 1-1) are valid for the instrument when it has been adjusted at an ambient temperature between +20°C and +30°C, has had a warm-up period of at least 30 minutes, and is operating at an ambient temperature between 0°C and +50°C (unless otherwise noted).

Items listed in the Performance Requirements column are verifiable qualitative or quantitative limits that define the measurement capabilities of the instrument.

Environmental characteristics are given in Table 1-2. The 2236A meets the requirements of MIL-T-28800C, paragraphs 4.5.5.1.3, 4.5.5.1.4, and 4.5.5.1.2.2 for Type III, Class 5 equipment, except otherwise noted.


Physical characteristics of the instrument are listed in Table 1-3.

Table 1-1  
Electrical Characteristics

Characteristics	Performance Requirements
<b>VERTICAL DEFLECTION SYSTEM</b>	
Deflection Factor	
Range	2 mV per division to 5 V per division in a 1-2-5 sequence.
Accuracy	
+ 15°C to + 35°	±2%.
0°C to + 50°C	±3%. <sup>a</sup>
	For 5 mV per division to 5 V per division VOLTS/DIV switch settings, the gain is set at a VOLTS/DIV switch setting of 10 mV per division.
	2 mV per division gain is set with the VOLTS/DIV switch set to 2 mV per division.
Range of VOLTS/DIV Variable Control	Continuously variable between settings. Increases deflection factor by at least 2.5 to 1.
Step Response	
Rise Time	
+ 10°C to + 35°	
5 mV per Division to 5 V per Division	3.5 ns or less. <sup>a</sup>
2 mV per Division	3.9 ns or less. <sup>a</sup>
+ 35°C to + 50°	
5 mV per Division to 5 V per Division	3.9 ns or less. <sup>a</sup>
2 mV per Division	4.4 ns or less. <sup>a</sup>
	Rise time is calculated from the formula: $\frac{0.35}{\text{Bandwidth (-3 dB)}}$
Aberrations	
Positive-Going Step	
2 mV per Division to 0.5 V per Division	+ 4%, -4%, 4% p-p. <sup>a</sup>
1 V per Division to 5 V per Division	+ 12%, -12%, 12% p-p. <sup>a</sup>
	Measured with 5-division reference signal, centered vertically, from a 50 Ω source driving a 50 Ω coaxial cable terminated in 50 Ω at the input connector with the VOLTS/DIV Variable control in the CAL detent. Vertically center the top of the reference signal.


<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements
Bandwidth (-3 dB)	
+0°C to +35°	
5 mV per Division to 5 V per Division	Dc to at least 100 MHz.
2 mV per Division	Dc to at least 90 MHz.
+35°C to +50°	
5 mV per Division to 5 V per Division	Dc to at least 90 MHz. <sup>a</sup>
2 mV per Division	Dc to at least 80 MHz. <sup>a</sup> Measured with a vertically centered 6-division reference signal from a 50 Ω source driving a 50 Ω coaxial cable that is terminated in 50 Ω, both at the input connector and at the probe input, with the VOLTS/DIV Variable control in the CAL detent.
AC Coupled Lower Limit	10 Hz or less at -3 dB. <sup>a</sup>
Bandwidth Limiter	Upper limits (-3 dB) bandpass at 20 MHz ±10%.
AC Coupled Lower Cutoff Frequency	10 Hz or less at -3 dB. <sup>a</sup>
Chop Mode Switching Rate	500 kHz ±30%. <sup>a</sup>
Input Characteristics	
Resistance	1 MΩ ±2%. <sup>a</sup>
Capacitance	20 pF ±2 pF. <sup>a</sup>
Maximum Safe Input Voltage 	See Figure 1-1 for derating curve.
DC Coupled	400 V (dc + peak ac) or 800 V ac p-p to 10 kHz or less. <sup>a</sup>
AC Coupled	400 V (dc + peak ac) or 800 V ac p-p to 10 kHz or less. <sup>a</sup>
Common-Mode Rejection Ratio (CMRR)	At least 10 to 1 at 80 MHz. Checked at 10 mV per division for common-mode signals of 6 divisions or less with VOLTS/DIV Variable control adjusted for best CMRR at 50 kHz.
Input Current	1 nA or less (0.5 division or less trace shift when switching between DC and GND input coupling with the VOLTS/DIV switch at 2 mV per division). <sup>a</sup>
Trace Shift with Volts/Div Switch Rotation	0.75 division or less. <sup>a</sup> VOLTS/DIV Variable control in CAL detent.

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements		
Trace Shift as VOLTS/DIV Variable Control is Rotated	1.0 division or less. <sup>a</sup>		
Trace Shift with Invert	1.5 division or less. <sup>a</sup>		
Channel Isolation	Greater than 100 to 1 at 50 MHz.		
POSITION Control Range	At least ±11 division from graticule center.		
<b>TRIGGER SYSTEM</b>			
A TRIGGER Sensitivity (P-P AUTO and NORM)	<b>10 MHz</b>	<b>60 MHz</b>	<b>100 MHz</b>
A Sweep			
Internal	0.35 div	1.0 div	1.5 div
External	35 mV	120 mV	150 mV
Counter			
Internal	0.5 div	1.5 div	2.0 div
External	50 mV	160 mV	300 mV
HF REJ	Reduces trigger signal amplitude at high frequencies by about 20 dB with rolloff beginning 40 kHz ± 25%. <sup>a</sup>		
LF REJ	Attenuates signals below 40 kHz (-3 dB point at 40 kHz ± 25%). <sup>a</sup>		
Lowest Usable Frequency in P-P AUTO Mode	20 Hz with 1.0 division internal or 100 mV external. <sup>a</sup>		
TV LINE (Sync Amplitude)	<u>Internal<sup>a</sup></u>	<u>External<sup>a</sup></u>	
	0.35 div	35 mV p-p	
TV FIELD Mode	≥ 1 division of composite sync. <sup>a</sup>		
B TRIGGER Sensitivity (Internal Only)	<b>10 MHz</b>	<b>60 MHz</b>	<b>100 MHz</b>
B Sweep			
Counter	0.4 div	1.2 div	1.5 div
Counter	0.5 div	1.5 div	2.0 div
EXT INPUT			
Maximum Input Voltage 	400 V (dc + peak ac) or 800 V ac p-p at 10 kHz or less. <sup>a</sup> See Figure 1-1 for derating curve.		
Input Resistance	1 MΩ ± 2%. <sup>a</sup>		
Input Capacitance	20 pF ± 2.5 pF. <sup>a</sup>		
AC Coupled Lower Cutoff Frequency	10 Hz or less at lower -3 dB point. <sup>a</sup>		
Offset	25 mV or less.		

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements	
LEVEL Control Range		
A TRIGGER (NORM)		
INT	Can be set to any point of the trace that can be displayed. <sup>a</sup>	
EXT, DC	At least ±1.6 V, 3.2 V p-p.	
EXT, DC ÷ 10	At least ±16 V, 32 V p-p. <sup>a</sup>	
B TRIGGER (Internal)	Can be set to any point of the trace that can be displayed. <sup>a</sup>	
VAR HOLDOFF Control	Increases A Sweep holdoff time by at least a factor of 10. <sup>a</sup>	
Trigger View System		
Deflection Factor		
Internal	Same as vertical.	
External		
AC and DC	100 mV per division.	
DC ÷ 10	1 V per division.	
Accuracy	±20%.	
Delay Difference Between EXT INPUT and Either Vertical Channel	Less than 3.0 ns. <sup>a</sup>	
<b>HORIZONTAL DEFLECTION SYSTEM</b>		
Sweep Rate		
Calibrated Range		
A Sweep	0.5 s per division to 0.05 μs per division in a 1-2-5 sequence of 22 steps. X10 magnifier extends maximum sweep speed to 5 ns per division.	
B Sweep	50 ms per division to 0.05 μs per division in a 1-2-5 sequence of 19 steps. X10 magnifier extends maximum sweep speed to 5 ns per division.	
Accuracy	<b>Unmagnified</b>	<b>Magnified</b>
+ 10°C to +35°C	±2%	±3%
0°C to +50°C	±3%	±4%
	Sweep accuracy applies over the center 8 divisions. Exclude the first 25 ns of the sweep for the X10 magnified sweep speeds and anything beyond the 100th magnified division.	
POSITION Control Range	Start of sweep to 10th division will position past the center vertical graticule line in X1 or 100th division in X10.	

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements
Sweep Linearity	± 5%. Linearity measured over any 2 of the center 8 divisions. With magnifier in X10, exclude the first 25 ns and anything past the 100th division of the X10 magnified sweeps.
Variable Control Range	Continuously variable between calibrated settings of the SEC/DIV control. Extends the A and the B Sweep speeds by at least a factor of 2.5 times over the calibrated SEC/DIV switch settings.
Sweep Length	Greater than 10 division.
A/B SWP SEP Range	± 3.5 divisions or greater.
Delay Time	Applies to 0.5 μs per division and slower. Delay time is functional, but not calibrated at A sweep speeds above 0.5 μs per division.
Delay POSITION Range	Less than (0.5 div + 300 ns) to greater than 10 divisions.
Jitter	One part or less in 10,000 (0.01%) of the maximum available delay time.

**X-Y OPERATION (X1 MAGNIFICATION)**

Deflection Factors	Same as Vertical Deflection System with the VOLTS/DIV Variable controls in CAL detent positions.
Accuracy	
X-Axis	
+ 15°C to 35°C	±3%.
0°C to + 50°C	±4%. <sup>a</sup>
	Measured with a dc-coupled, 5-division reference signal.
Y-Axis	Same as Vertical Deflection System. <sup>a</sup>
Bandwidth (-3 dB)	
X-Axis	Dc to a least 2.5 MHz.
Y-Axis	Same as Vertical Deflection System. <sup>a</sup>
Phase Difference Between X- and Y-Axis Amplifiers	±3° from dc to 150 kHz. Vertical Input Coupling set to DC.

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements																
<b>PROBE ADJUST</b>																	
Output Voltage of PROBE ADJUST Jack	0.5 V ±2%.																
Repetition Rate	1 kHz ±5%. <sup>a</sup>																
<b>Z-AXIS INPUT</b>																	
Sensitivity	5 V causes noticeable modulation. Positive-going input decreases intensity.  Usable: frequency range is dc to 20 MHz.																
Maximum Safe Input Voltage	30 V (dc + peak ac) or 30 VC p-p ac at 1 kHz or less. <sup>a</sup>																
Input Resistance	10 kΩ ±10%. <sup>a</sup>																
<b>COUNTER-TIMER</b>																	
Frequency	<table border="1"> <thead> <tr> <th>Ranges</th> <th>Maximum Resolution<sup>a</sup></th> </tr> </thead> <tbody> <tr> <td>100 Hz</td> <td>0.00001 Hz</td> </tr> <tr> <td>1 kHz</td> <td>0.0001 Hz</td> </tr> <tr> <td>10 kHz</td> <td>0.001 Hz</td> </tr> <tr> <td>100 kHz</td> <td>0.01 Hz</td> </tr> <tr> <td>1 MHz</td> <td>0.1 Hz</td> </tr> <tr> <td>10 MHz</td> <td>1 Hz</td> </tr> <tr> <td>100 MHz</td> <td>10 Hz</td> </tr> </tbody> </table>	Ranges	Maximum Resolution <sup>a</sup>	100 Hz	0.00001 Hz	1 kHz	0.0001 Hz	10 kHz	0.001 Hz	100 kHz	0.01 Hz	1 MHz	0.1 Hz	10 MHz	1 Hz	100 MHz	10 Hz
	Ranges	Maximum Resolution <sup>a</sup>															
100 Hz	0.00001 Hz																
1 kHz	0.0001 Hz																
10 kHz	0.001 Hz																
100 kHz	0.01 Hz																
1 MHz	0.1 Hz																
10 MHz	1 Hz																
100 MHz	10 Hz																
Maximum Input Frequency	Same as trigger specifications. <sup>a</sup>																
Minimum Displayable Frequency	0.20000 Hz. <sup>a</sup>																
Time Base Error																	
Standard	±1 X 10 <sup>-5</sup> (10 parts per million). Less than 5 X 10 <sup>-6</sup> change per year. <sup>a</sup>																
Option 14	±5 X 10 <sup>-7</sup> (0.5 parts per million). Less than 1 X 10 <sup>-7</sup> change per month. <sup>a</sup>																
Nongated Mode (Hz)																	
Resolution Error	$\left[ \pm \left( \frac{1.4 \text{ TJE}}{N} \right) F^2 \pm \text{LSD} \right]_{a, b}$																
Accuracy	Resolution Error ± (TBE)F.																

TJE = Trigger jitter error (seconds).

F = Frequency of input (Hz).

N = Number of input samples accumulated = F(0.25 second ±0.01 second) ≥ 1.

LSD = One count in least significant digit.

TBE = Time base error (fractional).

<sup>a</sup> Performance Requirement not checked in Service Manual.

<sup>b</sup> Reduces to ± LSD for pulse inputs with transition times ≤20 ns.

Table 1-1 (cont)

Characteristics	Performance Requirements	
Frequency (cont)		
Gated Mode (Hz)		
Resolution Error	$\left[ \pm \left( \frac{1.4 \text{ TJE}}{N_g \sqrt{G}} \right) F^2 \pm \text{LSD} \right]_{a,b}$	
Accuracy	Resolution Error $\pm$ Frequency Gating Error $\pm$ (TBE)F.	
Gating Error	$\left( \frac{2 \text{ ns}}{N_g} \right) F^2{}^a$	
Period	Ranges	Maximum Resolution <sup>a</sup>
	5 s	1 $\mu$ s
	1 s	100 ns
	100 ms	10 ns
	10 ms	1 ns
	1 ms	100 ps
	100 $\mu$ s	10 ps
Minimum Input Period	See trigger specifications.	
Maximum Displayable Period	5 seconds. <sup>a</sup>	
Nongated Mode (seconds)		
Resolution Error	$\left[ \pm \left( \frac{1.4 \text{ TJE}}{N} \right) \pm \text{LSD} \right]_{a,b}$	
Accuracy	Resolution Error $\pm$ (TBE)P. <sup>a</sup>	
Time Interval Gating Error	$\frac{2 \text{ ns}}{N_g}$	

F = Frequency of input (Hz).

G = Number of gate intervals in one measurement = (A sweep repetition rate) (0.25 second  $\pm$  0.01 second)  $\geq$  1.

LSD = One count in least significant digit.

N = Number of input samples accumulated = F(0.25 second  $\pm$  0.01 second)  $\geq$  1.

N<sub>g</sub> = Number of samples inside one gate interval  $\geq$  1.

P = Period of input (seconds).

TBE = Time base error (fractional).

TJE = Trigger jitter error (seconds).

<sup>a</sup> Performance Requirement not checked in Service Manual.

<sup>b</sup> Reduces to  $\pm$  LSD for pulse inputs with transition times  $\leq$  20 ns.

Table 1-1 (cont)

Characteristics	Performance Requirements	
Period (cont)		
Gated Mode (seconds)		
Resolution Error	$\left[ \pm \left( \frac{1.4 \text{ TJE}}{\text{Ng} \sqrt{G}} \right) \pm \text{LSD} \right]^{a, b}$	
Accuracy	Resolution Error $\pm$ Time Interval Gating Error $\pm$ (TBE)P.	
Time Interval Gating Error	$\frac{2 \text{ ns}}{\text{Ng}}$	
Width	Ranges	Maximum Resolution <sup>a</sup>
	5 s	1 $\mu$ s
	1 s	100 $\mu$ s
	100 ms	10 ns
	10 ms	1 ns
	1 ms	100 ps
	100 $\mu$ s	10 ps
Minimum Input Width	5 ns. <sup>a</sup>	
Maximum Displayable Width	5 seconds. <sup>a</sup>	
Nongated Mode (seconds)		
Resolution Error	$\left[ \pm \frac{1}{\sqrt{N}} (\text{TJE of leading edge} \pm \text{TJE of trailing edge}) \pm \frac{10 \text{ ns}}{\sqrt{N}} \right]^{a, c}$	
Accuracy	Resolution Error $\pm$ (TBE)W $\pm$ Hysteresis Error $\pm$ 10 ns. Checked in A TRIGGER NORM Mode.	
Gated Mode Accuracy (seconds)	Resolution Error <sup>d</sup> $\pm$ (TBE)W $\pm$ Hysteresis Error $\pm$ Time Interval Gating Error $\pm$ 10 ns.	

G = Number of gate intervals in one measurement = (A sweep repetition rate) (0.25 second  $\pm$  0.01 second)  $\geq$  1.

LSD = One count in least significant digit.

N = Number of input samples accumulated = F(0.25 second  $\pm$  0.01 second)  $\geq$  1.

Ng = Number of samples inside one gate interval  $\geq$  1.

P = Period of input (seconds).

TBE = Time base error (fractional).

TJE = Trigger jitter error (seconds).

W = Width being measured in seconds.

<sup>a</sup> Performance Requirement not checked in Service Manual.

<sup>b</sup> Reduces to  $\pm$  LSD for pulse inputs with transition times  $\leq$  20 ns.

<sup>c</sup> Reduces to  $\pm \frac{10 \text{ ns}}{\sqrt{G}}$  for pulse inputs with transition times  $\leq$  20 ns.

<sup>d</sup> Save as Resolution Error in nongated mode.

Table 1-1 (cont)

Characteristics	Performance Requirements														
Width (cont) Hysteresis Error (seconds)	$\left[ \frac{\text{sensitivity}}{2} \left( \frac{1}{\text{slew rate of trailing edge}} - \frac{1}{\text{slew rate of leading edge}} \right) \right]^{a,e}$ <p>Slew rate expressed in divisions on screen per second.</p>														
Sensitivity	See A and B TRIGGER sensitivities expressed in divisions on screen. <sup>a</sup>														
Delay Time	<table border="1" style="width: 100%; border-collapse: collapse; text-align: center;"> <thead> <tr> <th style="border-bottom: 1px solid black;">Ranges</th> <th style="border-bottom: 1px solid black;">Maximum Resolution<sup>a</sup></th> </tr> </thead> <tbody> <tr> <td>5 s</td> <td>1 μs</td> </tr> <tr> <td>1 s</td> <td>100 ns</td> </tr> <tr> <td style="border-bottom: 1px solid black;">100 ms</td> <td style="border-bottom: 1px solid black;">10 ns</td> </tr> <tr> <td>10 ms</td> <td>1 ns</td> </tr> <tr> <td>1 ms</td> <td>100 ps</td> </tr> <tr> <td style="border-bottom: 1px solid black;">100 μs</td> <td style="border-bottom: 1px solid black;">10 ps</td> </tr> </tbody> </table>	Ranges	Maximum Resolution <sup>a</sup>	5 s	1 μs	1 s	100 ns	100 ms	10 ns	10 ms	1 ns	1 ms	100 ps	100 μs	10 ps
Ranges	Maximum Resolution <sup>a</sup>														
5 s	1 μs														
1 s	100 ns														
100 ms	10 ns														
10 ms	1 ns														
1 ms	100 ps														
100 μs	10 ps														
Minimum Displayable Delay Time	500 ns. <sup>a</sup>														
Maximum Displayable Delay time	2.5 seconds. <sup>a</sup>														
B Triggered After Delay (seconds)  Resolution Error	$\left[ \pm \frac{1}{\sqrt{G}} (\text{TJE of leading edge} \pm \text{TJE of trailing edge}) \pm \frac{10 \text{ ns}}{\sqrt{G}} \right]^{a,f}$														
Accuracy	Resolution Error ± (TBE)Td ± 20 ns.														

G = Number of gate intervals in one measurement = (A sweep repetition rate) (0.25 second ± 0.01 second) ≥ 1.

TBE = Time base error (fractional).

Td = Delay time being measured in seconds.

TJE = Trigger jitter error (seconds).

<sup>a</sup> Performance Requirement not checked in Service Manual.

<sup>e</sup> Reduces to 5 ns for pulse inputs with transition times ≤ 20 ns and trigger level centered on waveform.

<sup>f</sup> Reduces to ±  $\frac{10 \text{ ns}}{\sqrt{G}}$  for pulse inputs with transition times ≤ 20 ns.

Table 1-1 (cont)

Characteristics	Performance Requirements	
Delta Time	Ranges	Maximum Resolution <sup>a</sup>
	5 s 1 s 100 ms	1 μs 100 ns 10 ns
Maximum Displayable Delta Time	10 ms 1 ms 100 μs	1 ns 100 ps 10 ps
	2.5 seconds. <sup>a</sup>	
B Runs After Delay (seconds)	$\pm \frac{2(\text{delay time jitter})}{\sqrt{G}} \pm \frac{10 \text{ ns}^a}{\sqrt{G}}$	
Accuracy	Resolution Error $\pm (TBE)T\Delta$ .	
B Triggered After Delay With Vertical MODE Switch in CH 1, CH 2, ADD, and CHOP (seconds)		
Resolution Error	$\left[ \pm \frac{1}{\sqrt{G}} (\text{TJE of leading edge} \pm \text{TJE of trailing edge}) \pm \frac{20 \text{ ns}}{\sqrt{G}} \right]^{a,f}$	
Accuracy	Resolution Error $\pm (TBE)T\Delta \pm 50 \text{ ps}$ .	
Resolution and Accuracy Definitions		
Trigger Jitter Error (TJE), Through Vertical Input Connectors (in seconds)	$\frac{\sqrt{(e_{n1})^2 + (e_{n2})^2}}{\text{Slew rate of triggering edge}}$ Slew rate expressed in divisions on screen per second.	
Value of $e_{n1}$ (divisions, RMS)	BW LIMIT on <sup>a</sup>	BW LIMIT off <sup>a</sup>
2 mV per Division	0.08	0.1
5 mV per Division to 5 V per Division	0.04	0.05

G = Number of gate intervals in one measurement = (A sweep repetition rate) (0.25 second  $\pm$  0.01 second)  $\geq$  1.

$e_{n1}$  = RMS noise of vertical system (in divisions on screen).

$e_{n2}$  = RMS noise voltage of input signal (in divisions on screen).

TBE = Time base error (fractional).

$T\Delta$  = Delta time being measured in seconds.

TJE = Trigger jitter error (seconds).

<sup>a</sup> Performance Requirement not checked in Service Manual.

<sup>f</sup> Reduces to  $\pm \frac{10 \text{ ns}}{\sqrt{G}}$  for pulse inputs with transition times  $\leq$  20 ns.

Table 1-1 (cont)

Characteristics	Performance Requirements	
<b>MULTIMETER</b>		
DC VOLTS	Ranges	Resolution <sup>a</sup>
	0.5 V 5 V 50 V 500 V	100 μV 1 mV 10 mV 100 mV
Accuracy		
+18°C to +28°C	±(0.1% of reading + 1 LSD).	
0°C to +18°C and +28°C to +40°C	±(0.2% of reading + 4 LSDs). <sup>a</sup>	
Normal Mode Rejection	≥ 50 dB from 48 Hz to 62 Hz.	
RMS Volts (AC Coupled)	Ranges	Resolution <sup>a</sup>
	0.5 V 5 V 50 V 500 V	100 μV 1 mV 10 mV 100 mV
Accuracy (20 Hz to 20 kHz)		
+18°C to +28°C	±(1.0% of reading + 6 LSDs).	
0°C to +18°C and +28°C to +40°C	±(1.5% of reading + 8 LSDs). <sup>a</sup>	
Common Mode Rejection	≥ 60 dB from 48 Hz to 62 Hz.	
Crest Factor	≤ 3.0 to maintain stated accuracy. <sup>a</sup>	
Resistance	Ranges	Resolution <sup>a</sup>
	50 Ω 500 Ω 5 kΩ 50 kΩ 500 kΩ 5 MΩ 50 MΩ 200 MΩ 2 GΩ	0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 10 MΩ

LSD = One count in least significant digit.

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements			
Resistance (cont)				
Overrange Indication (Resistance $\geq 2\text{ G}\Omega$ )	Display indicates <b>OPEN</b> . <sup>a</sup>			
Diode Detection (Fully Automatic)				
Detectable Forward Voltage Drop	0.15 volts to 2.0 volts. <sup>a</sup>			
Minimum Shunt Resistance	$\geq 2000\ \Omega$ per volt of forward drop. <sup>a</sup> Resistance shunts detectable device.			
Maximum Series Resistance	400 $\Omega$ per volt of forward drop. <sup>a</sup> Resistance in series with detectable device.			
Forward Drop Measurement Accuracy	$\pm(1\%$ of reading + LSD). <sup>a</sup>			
Accuracy	50 $\Omega$	500 $\Omega$ to 50 M $\Omega$	200 M $\Omega$	2 G $\Omega$
+ 18°C to + 28°C	$\pm(0.3\%$ of reading + 20 LSDs)	$\pm(0.15\%$ of reading + 2 LSDs)	$\pm(1.0\%$ of reading + 1 LSD)	$\pm(10.0\%$ of reading + 1 LSD)
0°C to + 18°C and + 28°C to + 40°C	$\pm(0.3\%$ of reading + 20 LSDs) <sup>a</sup>	$\pm(0.15\%$ of reading + 2 LSDs) <sup>a</sup>	$\pm(1.0\%$ of reading + 1 LSD) <sup>a</sup>	$\pm(10.0\%$ of reading + 1 LSD) <sup>a</sup>
Continuity	With less than 5.0 $\Omega$ $\pm 1\ \Omega$ measured, an audible tone will be generated. <sup>a</sup>			
Temperature				
Probe Tip Measurement Range	-62°C to + 240°C in one range. <sup>a</sup>			
Accuracy (At These Instrument Ambient Temperatures)				
+ 18°C to + 28°C	$\pm(2\%$ of reading $\pm 1.5^\circ\text{C}$ ) or $\pm(2\%$ of reading $\pm 2.7^\circ\text{F}$ ).			
0°C to + 18°C and + 28°C to + 50°C	$\pm(2\%$ of reading $\pm 2.0^\circ\text{C}$ ) or $\pm(2\%$ of reading $\pm 3.6^\circ\text{F}$ ).  No special calibration required for temperature (P6602).			

LSD = One count in least significant digit.

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements	
Multimeter Inputs	Isolated from the oscilloscope ground. <sup>a</sup>	
Input Impedance		
Resistance (DCV)	10 MΩ ±0.25%. <sup>a</sup>	
Capacitance (AC RMSV)	180 pF ±10%. <sup>a</sup>	
Maximum Safe Input Voltage		
+ (Positive) Input to Ground	500 V (dc + peak ac). <sup>a</sup>	
- (Negative) Input to Ground	500 V (dc + peak ac). <sup>a</sup>	
Positive to Negative Inputs	500 V (dc + peak ac). <sup>a</sup>	
	In all CTM functions and ranges.	
CH 1 Volts	Ranges	Resolution <sup>a</sup>
1X Probe	0.5 V 5 V 50 V	100 μV 1 mV 10 mV
10X Probe (P6121)	5 V 50 V 500 V	1 mV 10 mV 100 mV
	Ranges determined by CH 1 VOLTS/DIV switch. Multimeter automatically switches to 10 X ranges when a P6121 is used.	
Dc Volts		
Accuracy	1X Probe	10X Probe
+ 18°C to +28°C	±(0.30% of reading + 6 LSDs)	±(0.50% of reading + 6 LSDs)
0°C to +18°C and +28°C to +50°C	±(0.5% of reading + 20 LSDs) <sup>a</sup>	±(0.7% of reading + 20 LSDs) <sup>a</sup>
Normal Mode Rejection Ratio	≥ 30 dB from 48 Hz to 62 Hz. <sup>a</sup>	

LSD = One count in least significant digit.

<sup>a</sup> Performance Requirement not checked in Service Manual.

Table 1-1 (cont)

Characteristics	Performance Requirements		
CH 1 Volts (cont)			
AC RMS Volts			
Accuracy	1X PROBE		10X PROBE
	50 Hz to 100 Hz	100 Hz to 20 kHz	20 Hz to 20 kHz
+ 18°C to + 28°C	±(2.0% of reading + 6 LSDs)	±(1.0% of reading + 6 LSDs)	±(2.0% of reading + 6 LSDs)
0°C to + 13°C and + 28°C to + 50°C	±(2.25% of reading + 8 LSDs) <sup>a</sup>	±(1.25% of reading + 8 LSDs) <sup>a</sup>	±(2.25% of reading + 8 LSDs) <sup>a</sup>
	Probe compensation is adjusted ±0.1% of reading with CH 1 VOLTS/DIV switch set to 0.5 (X10), apply 4 V ±0.01%, 20 kHz sinewave.		
Crest Factor	≤ 3.0 to maintain stated accuracy. <sup>a</sup>		
<b>POWER SOURCE</b>			
Line Voltage Ranges	90 V to 250 V. <sup>a</sup>		
Line Frequency	48 Hz to 440 Hz. <sup>a</sup>		
Maximum Power Consumption	75 W (130 VA). <sup>a</sup>		
Line Fuse	1.25 A, 250 V, slow-blow. <sup>a</sup>		
<b>CATHODE-RAY TUBE</b>			
Display Area	80 by 100 mm. <sup>a</sup>		
Standard Phosphor	P31. <sup>a</sup>		
Nominal Accelerating Voltage	14 kV. <sup>a</sup>		

LSD = One count in least significant digit.

<sup>a</sup> Performance Requirement not checked in Service Manual.

**Table 1-2  
Environmental Characteristics**

Characteristics	Description
Environmental Requirements	Instrument meets or exceeds the environmental requirements of MIL-T-28800D for Type III, Class 5, Style D equipment as described below.
Temperature Operating AC RMSV, DCV, and $\Omega$ Modes All other Modes.	0°C to +40°C (+32°F to +104°F). 0°C to +50°C (+32°F to +122°F).
Nonoperating	-40°C to +71°C (-40°F to +160°F). Tested to MIL-T-28800D paragraphs 4.5.5.1.3 and 4.5.5.1.4, except in 4.5.5.1.3 steps 4 and 5 (0°C operating test) are performed ahead of step 2 (-40°C nonoperating test). Equipment shall remain off upon return to room ambient during step 6. Excessive condensation shall be removed before operating during step 7.
Altitude Operating Nonoperating	To 4,570 m (15,000 ft). Maximum operating temperature decreased 1°C per 1,000 ft above 5,000 ft. To 15,240 m (50,000 ft). Exceeds requirements of MIL-T-28800D paragraph 4.5.5.2.
Humidity (Operating and Nonoperating)	5 cycles (120 hours) referenced to MIL-T-28800D paragraph 4.5.5.1.2.2 for Type III, Class 5 instruments. Operating and non-operating at 95% +0° to -5% relative humidity. Operating at +30°C and +40°C for AC RMSV, DCV, and $\Omega$ Modes only and operating at 30°C and 50°C for all other modes of operation. Non-operating at 30°C to 60°C.
Radiated and Conducted Emission Requirements Per VDE 0871	Meets Class B.
Electrostatic Discharge	Withstands discharge of up to 20 kV. Test performed with probe containing a 500 pF capacitor with 1 K $\Omega$ series resistance charged to the test voltage. Conforms to Tektronix Standard 062-2862-00.
Vibration (Operating)	15 minutes along each of 3 major axis at a total displacement of 0.015 inch p-p (2.4 g's at 55Hz) with frequency varied from 10 Hz to 55 Hz to 10 Hz in 1-minute sweeps. Hold for 10 minutes at 55 Hz in each of the 3 major axis. All major resonances must be above 55 Hz. Meets requirements of MIL-T-28800D, paragraph 4.5.5.3.1.
Bench Handling Test (cabinet on and cabinet off)	Each edge lifted four inches and allowed to free fall onto a solid wooden bench surface. Meets requirements of MIL-T-28800D, paragraph 4.5.5.4.3.

Table 1-2 (cont)

Characteristics	Description
Shock (Operating and Non-operating)	<p>30 g's, half-sine, 11-ms duration, 3 shocks per axis each direction, for a total of 18 shocks.</p> <p>Meets requirements of MIL-T-28800D, paragraph 4.5.5.4.1, except limited to 30 g's.</p>
Transportation Packaged Vibration Test	<p>Meets the limits of the National Safe Transit Association test procedure 1A-B-1; excursion of 1 inch p-p at 4.63 Hz (1.1 g) for 30 minutes on the bottom and 30 minutes on the side (for a total of 60 minutes).</p>
Package Drop Test	<p>Meets the limits of the National Safe Transit Association test procedure 1A-B-2; 10 drops of 36 inches.</p>

**Table 1-3  
Physical Characteristics**

Characteristics	Description
Weight with Power Cord	
With Cover, Probes and Pouch	7.3 kg (16.0 lb).
Without Cover, Probes and Pouch	6.0 kg (13.3 lb).
Domestic Shipping Weight	10.1 kg (22.2 lb).
Height	
With Feet and Handles	137 mm (5.4 in).
Width	
With Handle	360 mm (14.2 in).
Without Handle	327 mm (12.9 in).
Depth	
With Front Cover	445 mm (17.5 in).
Without Front Cover	440 mm (17.3 in).
With Handle Extended	511 mm (20.1 in).




Figure 1-1. Maximum voltage versus frequency derating curve for the CH 1 OR X, CH 2 OR Y, and EXT INPUT OR Z connectors.




Figure 1-2. Physical dimensions of 2236A Oscilloscope.

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60  
61  
62  
63  
64  
65  
66  
67  
68  
69  
70  
71  
72  
73  
74  
75  
76  
77  
78  
79  
80  
81  
82  
83  
84  
85  
86  
87  
88  
89  
90  
91  
92  
93  
94  
95  
96  
97  
98  
99  
100

# OPERATING INSTRUCTIONS

## PREPARATION FOR USE

### SAFETY

Refer to the Operators Safety Summary at the front of this manual for power source, grounding, and other safety considerations pertaining to the use of this instrument. Before connecting the instrument to a power source, carefully read the following about line voltages, power cords, and fuses.

### LINE VOLTAGE

The instrument is capable of continuous operation using input voltages that range from 90 V to 250 V nominal at frequencies from 48 Hz to 440 Hz.

### POWER CORD

A detachable three-wire power cord with a three-contact plug is provided with each instrument to permit connection to both the power source and protective ground. The plug protective-ground contact connects (through the protective-ground conductor) to the accessible metal parts of the instrument. For electrical-shock protection, insert this plug only into a power outlet that has a securely grounded protective-ground contact. To secure the power cord to the instrument, use the power cord clamp as illustrated in Figure 2-1.

The instrument is shipped with the required power cord as ordered by the customer. Available power-cord information is illustrated in Figure 2-2, and part numbers are listed in Section 10 at the back of this manual. Contact your Tektronix representative, local Tektronix Field Office, or Tektronix product catalog for additional power-cord information.

### LINE FUSE

The instrument fuse holder is located on the rear panel (see Figure 2-1) and contains the line fuse. The following procedure can be used to verify that the proper fuse is installed or to install a replacement fuse.

1. Unplug the power cord from the power-input source (if applicable).




Figure 2-1. Power cord clamp and line fuse.

2. Press in and slightly rotate the fuse-holder cap counterclockwise to release it.
3. Pull the cap (with the attached fuse inside) out of the fuse holder.

4. Verify proper fuse value (1.25 A, 250 V, slow blow).
5. Reinstall the fuse (or replacement fuse) and the fuse-holder cap.

## INSTRUMENT COOLING







Always maintain adequate instrument cooling. The ventilation holes on both sides of the instrument cabinet and on the rear panel must remain free of obstruction.

## INSTRUMENT REPACKAGING

To ship an instrument, it is recommended that it be packaged in the original manner. The carton and packaging material in which the instrument was shipped should be saved and used for this purpose. The Accessory Pouch should be removed by a qualified service person before being shipped in the original carton.

If the original packaging is unfit for use or is not available, repackage the instrument as follows:

1. Obtain a corrugated cardboard shipping carton having inside dimensions at least six inches greater than the instrument dimensions and having a carton test strength of at least 275 pounds.
2. If the instrument is to be shipped to a Tektronix Service Center for service or repair, attach a tag to the instrument showing the following: owner of the instrument (with address), the name of a person who can be contacted, complete instrument type and serial number, and a description of the service required.
3. Wrap the instrument with polyethylene sheeting or equivalent to protect the outside finish and prevent entry of packing materials into the instrument.
4. Cushion the instrument on all sides by tightly packing dunnage or urethane foam between the carton and the instrument, allowing three inches on each side.
5. Seal the carton with shipping tape or with an industrial stapler.
6. Mark the address of the Tektronix Service Center and the return address on the carton in one or more prominent locations.

Plug Configuration	Option	Power Cord/ Plug Type	Line Voltage	Reference Standards <sup>b</sup>
	U.S. Std.	U.S. 120V	120V	ANSI C73.11 NEMA 5-15-P IEC 83 UL 198.6
	A1	EURO 220V	220V	CEE(7), II, IV, VII IEC 83 IEC 127
	A2	UK <sup>a</sup> 240V	240V	BS 1363 IEC 83 IEC 127
	A3	Australian 240V	240V	AS C112 IEC 127
	A4	North American 240V	240V	ANSI C73.20 NEMA 6-15-P IEC 83 UL 198.6
	A5	Switzerland 220V	220V	SEV IEC 127

<sup>a</sup> A 6A, type C fuse is also installed inside the plug of the Option A2 power cord.

<sup>b</sup> Reference Standards Abbreviations:

ANSI – American National Standards Institute  
AS – Standards Association of Australia  
BS – British Standards Institution  
CEE – International Commission on Rules for the Approval of Electrical Equipment  
IEC – International Electrotechnical Commission  
NEMA – National Electrical Manufacturer's Association  
SEV – Schweizerischer Elektrotechnischer Verein  
UL – Underwriters Laboratories Inc.

7685-19

Figure 2-2. Optional power cord data.

# CONTROLS, CONNECTORS AND INDICATORS

## DISPLAY AND POWER

Refer to Figure 2-3 for location of items 1 through 8.

- ① **Internal Graticule**—Eliminates parallax viewing error between trace and graticule lines. Rise-time amplitude and measurement points are indicated at the left edge of the graticule.
- ② **POWER Switch**—Turns instrument power on and off. Press in for ON; press again for OFF.
- ③ **Power Indicator**—An LED that illuminates when the instrument is operating.
- ④ **FOCUS Control**—Adjusts for optimum display definition.
- ⑤ **SCALE ILLUM Control**—Adjusts the light level of the graticule illumination.
- ⑥ **BEAM FIND Switch**—When held in, compresses the display to within the graticule area and provides a visible viewing intensity to aid in locating off-screen displays.
- ⑦ **TRACE ROTATION Control**—Screwdriver adjustment used to align the crt trace with horizontal graticule lines.
- ⑧ **A and B INTENSITY Controls**—Determines the brightness of the A and B Sweep traces.

## VERTICAL

Refer to Figure 2-4 for location of items 9 through 18.

- ⑨ **CH 1 VOLTS/DIV and CH 2 VOLTS/DIV Switches**—Used to select the vertical deflection factor in a 1-2-5 sequence. To obtain a calibrated deflection factor, the VOLTS/DIV Variable control must be in the calibrated (CAL) detent (fully clockwise). In CH 1 V CTM mode, the CH1 VOLTS/DIV switch selects the voltage range for the Multimeter.

**1X**—Indicates the deflection factor selected when using either a 1X probe or a coaxial cable.

**10X PROBE**—Indicates the deflection factor selected when using a 10X probe.




Figure 2-3. Power and display controls and power-on indicator.




Figure 2-4. Vertical and CAL controls and connectors.

- ⑩ **VOLTS/DIV Variable Controls**—When rotated counterclockwise out of their calibrated detent positions, these controls provide continuously variable, uncalibrated deflection factors between the calibrated settings of the VOLTS/DIV switches. In CH 1 V CTM mode, the CH 1 VOLTS/DIV Variable control has no influence on the Channel 1 volts measurements.
- ⑪ **POSITION Controls**—Used to vertically position the display on the crt. When the SEC/DIV switch is set to X-Y, the Channel 2 POSITION control moves the display vertically (Y-axis), and the Horizontal POSITION control moves the display horizontally (X-axis).
- ⑫ **A/B SWP SEP**—Vertically positions the B Sweep trace with respect to the A Sweep when ALT Horizontal MODE is selected.
- ⑬ **CH 1 OR X & DMM and CH 2 OR Y Input Connectors**—Provide for application of external signals to the instrument deflection system for display or to the Multimeter (CH 1 OR X & DMM input connectors only). In normal deflection mode (SEC/DIV switch not set to X-Y), signals from both connectors provide vertical deflection for the display. In the X-Y mode, the signal connected to the CH 1 OR X & DMM input connector provides horizontal deflection (X-axis), and the signal connected to the CH 2 OR Y input connector provides vertical deflection (Y-axis).

**14 Input Coupling (AC/RMSV AC–GND–DC/DCV and AC–GND–DC Switches**—Three-position switches that select the method of coupling the input signals to the instrument deflection system. When in CH 1 V CTM mode the Channel 1 Input Coupling switch selects the type of voltage measurement the Multimeter will perform on the Channel 1 input signal (RMSV AC or DCV).

**AC**—Input signal is capacitively coupled to the vertical amplifier. The dc component of the input signal is blocked. Low-frequency limit (–3 dB point) is about 10 Hz.

**GND**—The input of the vertical amplifier is grounded to provide a zero (ground) reference-voltage display (does not ground the input signal). This switch position allows precharging the input coupling capacitor.

**DC**—All frequency components of the input signal are coupled to the vertical deflection systems.

**15 Vertical MODE Switches**—Two three-position switches and two button switches are used to select the mode of operation for the vertical deflection system.

**CH 1**—Selects only the Channel 1 input signal for display.

**BOTH**—Selects both Channel 1 and Channel 2 input signals for display. The CH 1–BOTH–CH 2 switch must be in the BOTH position for either ADD, ALT, or CHOP operation.

**CH 2**—Selects only the Channel 2 input signal for display.

**ADD**—Displays the algebraic sum of the Channel 1 and Channel 2 input signals.

**ALT**—Alternately displays Channel 1 and Channel 2 input signals. The alternation occurs during retrace at the end of each sweep. This mode is useful for viewing both input signals at sweep speeds from 0.05  $\mu$ s per division to 0.2 ms per division.


**CHOP**—The display switches between the Channel 1 and Channel 2 input signals during the sweep. The switching rate is about 500 kHz.

This mode is useful for viewing both Channel 1 and Channel 2 input signals at sweep speeds from 0.5 ms per division to 0.5 s per division.

**BW LIMIT**—Limits the bandwidth of the vertical deflection system and the A Trigger system to about 20 MHz when the button is pressed in. Button must be pressed in a second time to release it and regain full 100-MHz bandwidth operation. Provides a method for reducing interference from high-frequency signals when viewing low-frequency signals.

**TRIG VIEW**—Press and hold the button in to display a sample of the signal present in the A Trigger amplifier (for all A & B SOURCE switch settings). All other signals displays are removed while the TRIG VIEW button is held in.

**16 INVERT Switch**—Inverts the Channel 2 display when button is pressed in. Button must be pressed in a second time to release it and regain a noninverted display.

**17 CAL  Connector**—Provides a 0.5 V, negative-going square-wave voltage at 1 kHz for compensating voltage probes and checking the operation of the oscilloscope's vertical system. It is not intended to verify the accuracy of the vertical and the horizontal deflection systems.

**18 GND Connector**—Provides direct connection to the instrument chassis ground.

## HORIZONTAL

Refer to Figure 2-5 for location of items 19 through 23.

**19 A and B SEC/DIV Switches**—Used to select the sweep speeds for the A and B Sweep generators in a 1–2–5 sequence. To obtain calibrated sweep speeds, the A and B SEC/DIV Variable control must be in the calibrated detent (fully clockwise).

**A SEC/DIV**—The calibrated sweep speed is shown between the two black lines on the clear plastic skirt. This switch also selects the delay time for delayed-sweep operation when used in conjunction with the B DELAY TIME POSITION control.




Figure 2-5. Horizontal controls.

**B SEC/DIV**—The B Sweep is set by pulling out the DLY'D SWEEP KNOB and rotating it clockwise to a setting opposite the white line scribed on the knob. The B Sweep circuit is used only for delayed-sweep operation.

**20 A and B SEC/DIV Variable Control and X0 Magnifier Switch**—Provides continuously variable, uncalibrated A Sweep speeds to at least 2.5 times the calibrated setting. It extends the slowest sweep speed to at least 1.25 s per division.

To expand the crt display by a factor of 10, pull out the X10 Magnifier control (SEC/DIV Variable control knob). The display portion of the sweep will be 10 times faster than the A and B SEC/DIV switch settings. This allows a maximum sweep speed of 5 ns per division. Push in the SEC/DIV Variable knob to regain the X1 (normal) sweep speed.

**21 Horizontal MODE Switch**—Determines the mode of operation for the horizontal deflection system, and for CTM frequency, period, width, and totalize measurements.

**A**—Horizontal deflection is provided by the A Sweep generator at a sweep speed determined by the A SEC/DIV switch setting. Nongated frequency, period, width, and totalize measurements are made on the A Trigger signal.

**ALT**—Alternates the horizontal displays between the A Sweep (with an intensified zone) and the B Delayed Sweep. The A Sweep speed is determined by the setting of the A SEC/DIV switch. The B Sweep speed and the length of the intensified zone on the A Sweep are both determined by the B SEC/DIV switch setting. Gated frequency, period, width, and totalize measurements are made on the B Trigger signal. The gate interval is defined by the length of the intensified zone and is valid only when B Sweep is triggered.

**B**—Horizontal deflection is provided by the B Sweep generator at a sweep speed determined by the B SEC/DIV switch setting. The start of the B Sweep is delayed from the start of the A Sweep by a time determined by the settings of both the A SEC/DIV switch and the B DELAY TIME POSITION control. Gated frequency, period, width, and totalize measurements are made on the B Trigger signal.

**22 POSITION Control**—Horizontally positions the A Sweep display, B Sweep display, and X-axis in X-Y mode.

**23 B DELAY TIME POSITION Control**—Selects the amount of delay time between the start of the A sweep and the start of the B Sweep. In  $\Delta$  TIME mode, the B DELAY TIME POSITION control operates in conjunction with the  $\Delta$  TIME POSITION control. The B DELAY TIME POSITION control moves both intensified zones (reference and comparisons), while the  $\Delta$  TIME POSITION control moves only one intensified zone (time-measurement point). Time difference between the start of the two intensified zones is displayed on the readout. In DLY TIME mode, only one intensified zone is displayed which is controlled by the B DELAY TIME POSITION control. The time difference between the start of the A Sweep and the start of the intensified zone is displayed on the readout.

## TRIGGER

Refer to Figure 2-6 for location of items 24 through 34.




Figure 2-6. Trigger controls, connector, and indicator.

- 24 **A TRIGGER Mode Switches**—Determine the A Sweep triggering mode.

**NORM**—Sweep is initiated when an adequate trigger signal is applied. In the absence of a trigger signal, no baseline trace will be present.

**P-P AUTO/TV LINE**—Permits triggering on trigger signals having adequate amplitude and a repetition rate of about 20 Hz or faster. In the absence of a proper trigger signal, an

autotrigger is generated, and the sweep free runs. The range of the A TRIGGER LEVEL control is restricted to the peak-to-peak range of the trigger signal. P-P AUTO is the usual trigger mode selection to obtain stable displays of TV Line information.

**TV FIELD**—Permits stable triggering on a television field (vertical sync) signal when the P-P AUTO and the NORM Trigger buttons are pressed in together. In the absence of an adequate trigger signal, the sweep free-runs. The instrument otherwise behaves as in P-P AUTO.

**SGL SWP**—Arms the A Trigger circuit for a single-sweep display or to reset the CTM when in TOTALIZE mode. In CH 1 V mode, pressing the SGL SWP button will enter or cancel relative reference mode. Triggering requirements are the same as in NORM trigger mode, except only one sweep is displayed for each trigger signal. After the completion of a triggered sweep, pressing in the SGL SWP button rearms the trigger circuitry to accept the next triggering event. This mode is useful for displaying and photographing either nonrepetitive signals or signals that cause unstable conventional displays (e.g., signals that vary in amplitude, shape, or time).

- 25 **RESET/READY Indicator**—A dual-function LED indicator. In P-P AUTO and NORM Trigger modes, the LED is turned on when triggering occurs. In SGL SWP Trigger mode, the LED turns on when the A Trigger circuit is armed, awaiting a triggering event, and turns off again after the single sweep event occurs.

- 26 **A TRIGGER LEVEL Control**—Selects the voltage level on the A Trigger signal that produces triggering.

- 27 **A TRIGGER SLOPE Switch**—Selects either the positive slope (button out) or negative slope (button in) of the trigger signal to start the A Sweep. In WIDTH mode, the A TRIGGER SLOPE switch selects the positive displayed or the negative displayed half cycles of the trigger signals to be measured by the CTM.

- 28 **A & B SOURCE Switch**—Selects the source of the trigger signal for both the A and the B Trigger Generator circuits.

The A Trigger source will be a sample of the ac line voltage regardless of the setting of the A & B SOURCE switch when the A COUPL switch is set to A LINE SOURCE.

**VERT MODE** – Trigger signals for both A & B are obtained alternately from the Channel 1 and Channel 2 input signals in ALT Vertical MODE. In the CHOP or ADD Vertical MODE the trigger signal is the sum of the Channel 1 and Channel 2 input signals. See Table 2-1 for VERT MODE trigger source.

**CH 1** – The signal applied to the CH 1 OR X & DMM input connector is the source of the A & B trigger signal.

**CH 2** – The signal applied to the CH 2 OR Y input connector is the source of the A & B trigger signal. The polarity CH 2 Trigger signal may be inverted by the Channel 2 INVERT switch so the displayed slope agrees with the Trigger SLOPE switch.

**A EXT** – Signals applied to the EXT INPUT connector are routed to the A Trigger circuit. The B Trigger source will be as described for VERT MODE.

Table 2-1  
Vertical MODE Trigger Source

VERT MODE	Trigger Source
CH 1	CH 1 OR X AND DMM input signal.
CH 2	CH 2 OR Y input signal.
BOTH and ADD	Algebraic sum of CH 1 OR X and CH 2 OR Y input signals.
BOTH and CHOP	Algebraic sum of CH 1 OR X and CH 2 OR Y input signals.
BOTH and ALT	Alternates between Channel 1 and Channel 2 on every other sweep (i.e., CH 1 OR X input signal triggers the sweep that displays Channel 1, and CH 2 OR Y input signal triggers the sweep that displays Channel 2.)

29 **A COUPL Switch** – Selects the method of coupling the input trigger signal to the A Trigger circuit.

**NORM** – All frequency components of the trigger signals are coupled to the A Trigger circuit.

**HF REJ** – Attenuates the high-frequency triggering signal components above 40 kHz of the trigger signal.

**LF REJ** – Attenuates low-frequency triggering signal components below 40 kHz of the trigger signal.

**A LINE SOURCE** – Routes a sample of the ac power source waveform to the A Trigger circuit regardless of the setting of the A & B SOURCE switch.

30 **A EXT COUPL Switch** – Selects the method of coupling the external signal applied to the EXT INPUT connector to the A Trigger circuit.

**AC** – Input signal is capacitively coupled, and blocks the dc component of the signal.

**DC** – Couples dc and all frequency components of the external trigger signal.

**DC/10** – Attenuates the external signal by a factor of 10. Couples dc and all frequency components of the external trigger signal.

31 **EXT INPUT Connector** – Provides for connection of external signals to the A Trigger circuit.

32 **B TRIGGER SLOPE Switch** – Selects either the positive slope (button out) or the negative slope (button in) of the B Trigger signal (internal source only) that starts the B sweep. In WIDTH mode, the B TRIGGER SLOPE switch selects the positive displayed or the negative displayed half cycles of the trigger signals to be measured by the CTM.

33 **B TRIGGER LEVEL Control** – Selects the amplitude point on the B Trigger signal where triggering occurs in triggerable after delay mode. The fully clockwise position of the B TRIGGER LEVEL Control selects the runs after delay mode of operation for the B Trigger circuitry. Out of the cw position, B Sweep is triggerable after the delay time.

- 34 **VAR HOLDOFF Control**—Varies the holdoff time over a 10 to 1 range. Variable Holdoff starts at the end of the A Sweep. This control improves the ability to trigger on aperiodic signals (such as complex digital waveforms).

### COUNTER, TIMER, AND MULTIMETER

Refer to Figure 2-7 for location of items 35 through 38.

- 35 **UPPER FUNCTIONS/LOWER FUNCTIONS Switch**—Push-push switch that determines which set of CTM functions is activated by the Function Select switches.
- 36 **Function Select Switches**—Five button switches along with the UPPER FUNCTIONS/LOWER FUNCTIONS switch select the CTM functions. With all the switch buttons out, the CTM is disabled and the readout is blank.

With the UPPER FUNCTIONS/LOWER FUNCTIONS switch in the IN position the following CTM functions can be selected.

**FREQ**—Measures the frequency of the trigger signal from the output of the A Trigger circuit (instrument in A Horizontal MODE) or the B Trigger circuit (instrument in ALT or B Horizontal MODE).

**PER**—Measures the period of the trigger signal from the output of the A Trigger circuit (instrument in A Horizontal MODE) or the B

Trigger circuit (instrument in ALT or B Horizontal MODE).

**TOTALIZE**—Press in both FREQ and PER buttons. Counts trigger events in the A Trigger circuit (instrument in A Horizontal MODE) or the B Trigger circuit (instrument in ALT or B Horizontal MODE). The displayed count can be reset to zero by switching between A and ALT Horizontal MODE or by pressing in momentarily the SGL SWP RESET button.

**WIDTH**—Measures the width of the trigger signal from the output of the A Trigger circuit (instrument in A Horizontal MODE) or the B Trigger circuit (instrument in ALT or B Horizontal MODE). With the trigger slope switch in positive position, the CTM measures the positive displayed half cycles of the trigger signals; when in negative position the CTM measures the negative displayed half cycles of the trigger signals.

**DCV**—Measures dc voltage from 0 to 499.9 volts when applied to the Multimeter inputs. The word **OUCH** is displayed if 500 V or more is applied to the Multimeter inputs.

**$\Omega$  /  $\rightarrow$  (Resistance/Semiconductor)**—Measures resistance up to 1.99 G $\Omega$  or indicates that the device-under-test is a forward-biased semiconductor when applied to the Multimeter inputs.

With the UPPER FUNCTIONS/LOWER FUNCTIONS switch in the OUT position the following CTM functions can be selected.




Figure 2-7. Counter, timer, and multimeter controls and indicator.

**Δ TIME** – Provides two intensified zones on the CRT trace for differential time measurements. The location of the first intensified zone is determined by the B DELAY TIME POSITION control, and location of the second intensified zone is determined by the Δ TIME POSITION control. The time difference between the start of the two intensified zones is displayed on the readout.

**DLY TIME** – Measures and displays on the readout the time difference between the start of the B Sweep (B DLY'D SWEEP) and the start of the A Sweep in ALT and B Horizontal MODE.

**CH 1 V** – Measures dc voltage or true ac rms voltage signals applied to the CH 1 OR X & DMM vertical input connector. For dc voltage measurements set the Channel 1 Input Coupling switch to DCV measurements, and for true ac rms voltage measurements set the Channel 1 Input coupling switch to the AC RMSV position. There is nonautoranging; CH 1 VOLTS/DIV switch selects the range for the CTM voltage measurements. The readout will indicate the measuring mode the CTM is in by displaying an **Ac** or **dc** on the left side of the voltage reading. In overrange condition, the CTM readout will display the word **OUCH**.

In CH 1 V mode, the CTM has the capability of storing a relative reference value, such that the displayed value is the input voltage minus the stored reference value. When either Channel 1 AC RMSV or DCV volts are first entered, the stored reference value is zero, and the Channel 1 input voltage is displayed directly. To set the reference value, press in the SGL SWP button momentarily, and the value on the readout becomes the stored reference value. To regain normal voltage measurement press again the SGL SWP RESET button momentarily or change the CH 1 VOLTS/DIV switch position.

**AC RMSV** – Measures the true ac rms value of the input signal applied to the Multimeter inputs, from 0 V to 349.9 V rms. The word **OUCH** is displayed if 350 V or more is applied to the Multimeter input connector.

**TEMP** – Measures temperature when the optional P6602 temperature probe is connected

to the MULTIMETER INPUTS connectors. Temperatures are measured in degrees Celsius °C when the TEMP button is pressed in, or degrees Fahrenheit °F when both TEMP and Δ TIME buttons are pressed in. Temperature readings following by °C or °F are displayed on the CTM readout.

**☞) < 5Ω (Continuity)** – Press in both AC RMSV and TEMP buttons for continuity measurements. For resistance readings below 5.0 Ω the word **Short** is displayed and an audible tone will be generated. For resistance readings greater than or equal to 5.0 Ω, the word **OPEN** is displayed.

**Self-Test Routine** – The CTM can be entered into a Self-Test routine by setting the UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN position and pressing in the FREQ, PER, and WIDTH buttons at the same time. The message **SELF tEST** will be displayed on the readout to indicate that the routine is in self test mode. To exit from the routine press in any of the CTM front-panel buttons to regain normal measurement mode. Repeated pressing of the SGL SWP button will cause the CTM to advance through the test menu. To exit from the test menu, continue pressing the SGL SWP button until **END S-t** message is displayed on the readout and then press one of the CTM front panel buttons. The self test routine should be performed by a qualified service person.

③⑦ **Δ TIME POSITION Control** – Used in conjunction with the B DELAY TIME POSITION control in TIME mode. The control determines the time difference between the starts of the two intensified zones.

③⑧ **Readout** – Consists of a nine-digit vacuum fluorescent unit which is used to display measurements selected by the CTM. No polarity indication is displayed for positive values. Negative polarity indication is automatic for negative values. In Counter and Timer modes the decimal point is always placed in one of the three most significant digits, and the exponent is always an integer multiple of three. Resistance measurements are also displayed with an exponent. All frequency measurements are in Hertz, and time measurements are in seconds.

### RIGHT SIDE PANEL

Refer to Figure 2-8 for location of item 39.

- 39 **Multimeter Connectors**—Two banana like jacks provide positive (red) and negative (black) inputs for voltages, resistance, and temperature measurements.




Figure 2-8. Multimeter right side panel connectors.

### REAR PANEL

Refer to Figure 2-9 for location of item 40.

- 40 **EXT Z-AXIS Connector**—Provides a means of connecting external signals to the Z-Axis amplifier to intensity modulate the crt. Applied signals do not affect display waveshape. signals with fast rise times and fall times provide the best intensity change, and a 5 V p-p signal will produce noticeable modulation. The Z-axis signals must be time-related to the display to obtain a stable presentation on the crt.




Figure 2-9. Rear-panel connector.

## OPERATING CONSIDERATIONS

### GRATICULE

The graticule is internally marked on the faceplate of the crt to enable accurate measurements without parallax error (see Figure 2-10). It is marked with eight vertical and ten horizontal major divisions. Each major division is divided into five subdivisions. The vertical deflection factors and horizontal timing are calibrated to the graticule so that accurate measurements can be made directly from the crt. Also, percentage markers for the measurement of rise and fall times are located on the left side of the graticule.




Figure 2-10. Graticule measurement markings.

### GROUNDING

The most reliable signal measurements are made when this instrument and the unit under test are connected by a common reference (ground lead), in addition to the signal lead or probe. The probe's ground lead provides the best grounding method for signal interconnection and

ensures the maximum amount of signal-lead shielding in the probe cable. A separate ground lead can also be connected from the unit under test to the oscilloscope GND connector located on the front panel.

For floating Multimeter measurements the – (negative) connector (located on the right side panel) should be connected to the lower impedance point of the unit-under-test being measured, to minimize loading.

### SIGNAL CONNECTIONS

Generally, probes offer the most convenient means of connecting an input signal to the instrument. They are shielded to prevent pickup of electromagnetic interference, and the supplied 10X probe offers a high input impedance that minimizes circuit loading. This allows the circuit under test to operate with a minimum of change from its normal condition as measurements are being made.

Coaxial cables may also be used to connect signals to the input connectors, but they may have considerable effect on the accuracy of a displayed waveform. To maintain the original frequency characteristics of an applied signal, only high-quality, low-loss coaxial cables should be used. Coaxial cables should be terminated at both ends in their characteristic impedance. If this is not possible, use suitable impedance-matching devices.

### INPUT COUPLING CAPACITOR PRECHARGING

When the Input Coupling switch is set to GND, the input signal is connected to ground through the input coupling capacitor in series with a 1 MΩ resistor to form a precharging network. This network allows the input coupling capacitor to charge to the average dc-voltage level of the signal applied to the probe. Thus any large voltage transients that may accidentally be generated will not be applied to the amplifier input when the Input Coupling switch is moved from GND to AC. The precharging network also provides a measure of protection to the external circuitry by reducing the current levels that can be drawn from the external circuitry during capacitor charging.

The following procedure should be used whenever the probe tip is connected to a signal source having a different dc level than that previously applied, especially if the dc-level difference is more than 10 times the VOLTS/DIV switch setting:

1. Set the Input Coupling switch to GND.
2. Insert the probe tip into the oscilloscope GND connector and wait several seconds for the input coupling capacitor to discharge.
3. Connect the probe tip to the signal source and wait several seconds for the input coupling capacitor to charge.
4. Set the Input Coupling switch to AC. The display will remain on the screen, and the ac component of the signal can be measured in the normal manner.

### CTM MEASUREMENT CONSIDERATIONS

Nongated measurements are performed from the A Trigger system on signals selected by the A & B SOURCE switch. Signals may be applied through CH 1 OR X & DMM, CH 2 OR Y, or the EXT INPUT connectors. Nongated measurements may also be performed on the instrument power source by setting the A COUPL switch to A LINE SOURCE position. The A TRIGGER LEVEL control, A TRIGGER Mode, A TRIGGER SLOPE, A & B SOURCE, A COUPL, A EXT COUPL, and BW LIMIT switches are effective in conditioning the input signal.

Trigger status messages **no A trig** and **no b trig** are not displayed in any mode of the width function so that measurement results can be displayed indefinitely. This allows single pulse measurements.

Gated measurements are performed from the B Trigger system on input signals being applied to either CH 1, or CH 2 input connectors. The gate interval (set by adjusting the intensified zone with the B DELAY TIME POSITION control and the B SEC/DIV switch) must be shorter than the A Sweep duration, such that the intensified zone ends before A Sweep ends. The A & B SOURCE switch and the B TRIGGER LEVEL control and SLOPE switch are effective in conditioning the input signal.

Noise may be coupled to the trigger circuits along with the signal to be measured. Noise may originate from the operating environment, the signal source, or by improper connections. If the noise is of sufficient amplitude, it can result in inaccurate measurements due to false triggering. The instrument 20 MHz Bandwidth Limiter may be helpful in removing or reducing the noise in the A Trigger system.

Channel 1 volts measurements should be made with the signal fully displayed within the graticule area through the appropriate use of the CH 1 VOLTS/DIV switch and Variable control. This prevents the Channel 1 input circuit from being overdriven and distorting the voltage measurements.

# OPERATOR'S CHECKS AND ADJUSTMENTS

## INTRODUCTION

To verify the operation and accuracy of this instrument before making measurements, perform the following check and adjustment procedures. Adjustments beyond the scope of Operator's Adjustments are in the Adjustment Procedure, Section 5 of this manual.

Before proceeding with these instructions, refer to Preparation for Use in this section for first-time startup considerations.

Verify that the POWER switch is OFF (button out). Then plug the power cord into the power-source outlet:

## BASELINE TRACE

First, obtain a baseline trace using the following procedure:

### Display

A and B INTENSITY	Fully counter-clockwise
FOCUS	Midrange

### Vertical (Both Channels)

POSITION	Midrange
A/B SWP SEP	Midrange
MODE	CH 1
BW LIMIT	Off (button out)
VOLTS/DIV	10 mV
VOLTS/DIV Variable	CAL detent
INVERT	Off (button out)
Input Coupling	AC

### Horizontal

POSITION	Midrange
MODE	A
A and B SEC/DIV	0.5 ms
SEC/DIV Variable	CAL detent
X10 Magnifier	Off (knob in)
B DELAY TIME POSITION	Fully counter-clockwise

## B TRIGGER

SLOPE	Positive (button out)
LEVEL	RUNS AFTER DLY (fully clockwise)

## A TRIGGER

VAR HOLDOFF Mode	NORM P-P AUTO
SLOPE	Positive (button out)
LEVEL	Midrange
A & B SOURCE	VERT MODE
A COUPL	NORM
A EXT COUPL	AC

## CTM

UPPER FUNCTIONS/LOWER FUNCTIONS	OUT
Function Select	All buttons out
Δ TIME POSITION	Midrange

2. Press in the POWER switch button (ON) and allow the instrument to warm up (20 minutes is recommended for maximum accuracy).
3. Adjust the A INTENSITY control for desired display brightness.
4. Adjust the Vertical and Horizontal POSITION controls as needed to center the trace on the screen.

## TRACE ROTATION

Normally, the resulting trace will be parallel to the center horizontal graticule line, and the Trace Rotation adjustment would not be required. If adjustment is needed, perform the following procedure:

1. Preset instrument controls and obtain a baseline trace.
2. Use the Channel 1 POSITION control to move the baseline trace to the center horizontal graticule line.
3. If the resulting trace is not parallel to the center horizontal graticule line, use a small, flat-bladed screwdriver to adjust the TRACE ROTATION control and align the trace with the center horizontal graticule line.

## PROBE COMPENSATION

Misadjustment of probe compensation is one source of measurement error. Most attenuator probes are equipped with a compensation adjustment. To ensure optimum measurement accuracy, always compensate the oscilloscope probes before making measurements. Probe compensation is accomplished as follows:

1. Preset instrument controls and obtain a baseline trace.
2. Connect the two 10X probes (supplied with the instrument) to the CH 1 and CH 2 input connectors.
3. Connect the Channel 1 probe to the CAL output connector.
4. Use the the Channel 1 POSITION control to vertically center the 5-division display. Adjust the A TRIGGER LEVEL control to obtain a stable display.
5. Check the waveform display for overshoot and rolloff (see Figure 2-11). If necessary adjust the probe compensation for flat tops on the waveforms. Refer to the instructions supplied with probe for details of compensation adjustment.




Figure 2-11. Probe compensation.

### NOTE

*Channel 1 ac rms volts measurement accuracy above 1 kHz depends directly upon the precision with which the preceding adjustment is*

*performed. An overshoot of 0.2 division (or 4% of a 5-division display) will add approximately 4% to the displayed reading. Compensation related errors can be reduced 1/2% or less by carefully adjusting the probe compensation to the CAL square-wave signal.*

*For best measurement accuracy above 1 kHz with a 10X probe, refer the X10 probe adjustment to a qualified service personnel to perform Channel 1 RMSV AC probe adjustment in the Adjustment Procedure, Section 5 of this manual.*

6. Disconnect the Channel 1 probe from the CAL connector.
7. Connect the Channel 2 probe to the Cal connector.
8. Set the Vertical MODE to CH 2 and vertically center the 5-division display using the Channel 2 POSITION control.
9. Check the waveform display for overshoot and rolloff (see Figure 2-11). If necessary adjust the probe compensation for flat tops on the waveforms. Refer to the instructions supplied with probe for details of compensation adjustment.

## VERTICAL DEFLECTION CHECK

The CAL signal can be used as a convenient way of checking the instrument vertical deflection system with the following checks:

1. Preset the instrument controls and obtain a baseline trace.
2. Connect the two 10X probes (supplied with the instrument) to the CH 1 OR X and CH 2 OR Y input connectors.
3. Set both VOLTS/DIV switches to 0.1 V 10X PROBE setting and set both Input Coupling switches to DC.
4. Select CH 1 Vertical MODE and connect the Channel 1 probe to the CAL connector.
5. Using the 1 kHz CAL square-wave signal as the input, obtain a 5-division display of the signal.
6. Set the A SEC/DIV switch to display several cycles of the CAL signal. Use the Channel 1 POSITION control to vertically center the display.

7. Check for a vertical display amplitude of approximately 5 divisions.
8. Select CH 2 Vertical MODE and connect the Channel 2 probe to the CAL connector.
9. Use the Channel 2 POSITION control to vertically center the display and repeat step 7 for the Channel 2 probe.
10. Disconnect the probes from the instrument.

## POWER-UP CHECKS

The Counter, Timer, and Multimeter (CTM) contains Power-Up Checks that are performed each time the instrument is turned on. These checks consists of ROM checksum, counter, and DMM in this order. Each of these checks must pass before proceeding to the next check. If a check fails, an error message will appear in the readout (see Table 2-2). Refer the error message to a qualified service person.

Table 2-2  
Power-Up Checks

Check	Error Message
ROM Checksum	FAIL-ro
Counter	FAIL-ctr
DMM	FAIL-d

## STATUS AND ERROR MESSAGES

The Counter, Timer, and Multimeter (CTM) system will display several types of status and error messages on the readout.

**Control button Error**—When the buttons are in an illegal mode, the readout will display Control button Error by scrolling it across the readout. The error message will repeat itself until a correction is made.

**O’FLO**—When measuring the width of an input signal that exceeds five seconds in duration, the overflow message **O’FLO** will be displayed on the readout.

**no ALt H**—With the instrument in A Horizontal MODE (non-delay), the readout will display a **no**

**ALt H** message when either the  $\Delta$  TIME or DLY TIME mode is selected.

**no A trig**—With the  $\Delta$  TIME mode and either the ALT or B Horizontal MODE selected, the readout will display **no A trig** message when either the A Sweep is not running, or is too slow. This message is also displayed with the instrument in A Horizontal MODE and either the **FREQ** or **PER** mode selected, and the CTM is not receiving a trigger signal.

**no b trig**—When either the  $\Delta$  TIME or DLY TIME mode and either ALT or B Horizontal MODE is selected with the A Sweep triggered and the B Trigger circuit not receiving a trigger signal in delay mode, the readout will display **no b trig** message. This message is also displayed with the instrument in either ALT or B Horizontal MODE and either **FREQ** or **PER** mode selected, and the CTM is not receiving a trigger signal.

**OPEN**—The readout will display **OPEN** with the CTM in either Ohms or Continuity mode (AC RMSV and TEMP buttons pressed in) under the following conditions; in  $\Omega$  mode one or both test leads disconnected or when the resistance exceeds 1.99 G $\Omega$ , in Continuity mode when the resistance equals or exceeds 5.0  $\Omega$ .

**diode**—In Resistance/Semiconductor mode the readout will momentarily display **diode**, and an audible tone will be generated if the device being measured is a forward-biased semiconductor junction. After about one second, it will display the forward voltage drop of the device.

**Fd**—In Ohms mode **Fd** will be displayed on the left side of the readout to indicate that a forward voltage drop is being displayed.

**Probe-?**—When the temperature probe exceeds its internal resistance limits the readout will display **Probe-?** to indicate that the temperature probe is either faulty or disconnected from the Multimeter input connectors.

**Short**—In Continuity mode the readout will display **Short** when the measured resistance is less than 5.0  $\Omega$ .

**OUCH**—The message **OUCH** is displayed when the input voltage to the multimeter inputs is 500 volts or more in DCV mode, and 350 volts or more in AC RMSV mode. In CH 1 V mode, the message **OUCH** is displayed if overranged.

**dc**—In CH 1 V mode with Channel 1 Input Coupling switch in GND or DCV position, **dc** will be displayed on the left side of the readout.

**Ac**—In CH 1 V mode with Channel 1 Input Coupling switch in AC RMSV position, **Ac** will be displayed on the left side of the readout.

**r**—In CH 1 V Relative Reference mode, **r** will be displayed on the right side of the right side of the readout.

**no dELTA**—When in  $\Delta$  TIME mode with the measurement intensified zone (moved by rotating the  $\Delta$  TIME position control) dialed off the right end of the trace, the readout will display **no dELTA**.

# BASIC APPLICATIONS

## INTRODUCTION

The information in this part is designed to enhance operator understanding and to assist in developing efficient techniques for making specific measurements. Recommended methods for making basic measurements with your instrument are described in the procedures contained in this section.

When a procedure first calls for presetting instrument controls and obtaining a baseline trace, refer to the Operator's Checks and Adjustments part in this section and perform steps 1 through 4 under Baseline Trace.

## INDEX TO BASIC APPLICATION PROCEDURES

VOLTAGE MEASUREMENTS	
AC Peak-to-Peak Voltage .....	2-18
Instantaneous Voltage .....	2-19
Algebraic Addition .....	2-20
Common-Mode Rejection .....	2-20
Amplitude Comparison .....	2-21
TELEVISION DISPLAYS	
TV Line Signal .....	2-21
TV Field Signal .....	2-22
DELAYED-SWEEP	
Magnification .....	2-22
Magnified Sweep Runs After Delay .....	2-23
Pulse Jitter Time Measurement .....	2-23
Triggered Magnified Sweep .....	2-24
TIME AND FREQUENCY MEASUREMENTS	
Period .....	2-24
Frequency .....	2-24
Width .....	2-24
Time Interval On Signal Waveform .....	2-25
Rise Time .....	2-25
Time Difference Between Pulses On Two Time-Related Signals (Short Method) .....	2-26
Time Difference Between Points On Two Time-Related Signals (General Method) .....	2-27
Phase Difference .....	2-28
Gated Measurements .....	2-28
MULTIMETER MEASUREMENTS	
Resistance and Semiconductor Junctions ..	2-29

Determining Semiconductor	
Leakage Current .....	2-30
Continuity .....	2-31
AC RMS Volts .....	2-31
DC Volts .....	2-31
Channel 1 Volts .....	2-31
Temperature .....	2-32

## VOLTAGE MEASUREMENTS

### AC Peak-to-Peak Voltage

To make a peak-to-peak voltage measurement, use the following procedure:

#### NOTE

*This procedure may also be used to make voltage measurements between any two points on the waveform.*

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the ac signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set the appropriate VOLTS/DIV switch to display about five divisions of the waveform, ensuring that the VOLTS/DIV Variable control is in the CAL detent.
4. Adjust the A TRIGGER LEVEL control to obtain a stable display.
5. Set the A SEC/DIV switch to a position that displays several cycles of the waveform.
6. Vertically position the display so that the negative peak of the waveform coincides with one of the horizontal graticule lines (see Figure 2-12, Point A).
7. Horizontally position the display so that one of the positive peaks coincides with the center vertical graticule line (see Figure 2-12, Point B).
8. Measure the vertical deflection from peak-to-peak (see Figure 2-12, Point A to Point B).

**NOTE**

*If the amplitude measurement is critical or if the trace is thick (as a result of hum or noise on the signal), a more accurate value can be obtained by measuring from the top of a peak to the top of a valley. This will eliminate trace thickness from the measurement.*




Figure 2-12. Peak-to-peak waveform voltage.

- Calculate the peak-to-peak voltage, using the following formula:

$$\text{VOLTS (p-p)} = \text{vertical deflection (divisions)} \times \text{VOLTS/DIV switch setting indicated by 1X (or 10X PROBE when 10X probe is used)}$$

**EXAMPLE:** The measured peak-to-peak vertical deflection is 4.6 divisions (see Figure 2-12) with a VOLTS/DIV switch setting of 5 V (at 10X PROBE indicator), using a 10X probe.

Substituting the given values:

$$\text{Volts (p-p)} = 4.6 \text{ div} \times 5 \text{ V/div} = 23 \text{ V.}$$

**Instantaneous Voltage**

To measure instantaneous voltage level at a given point on a waveform, referred to ground, use the following procedure:

- Preset the instrument controls and obtain a baseline trace.

- Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
- Verify that the VOLTS/DIV Variable control is in the CAL detent and set the Input Coupling switch to GND.
- Vertically position the baseline trace to the center horizontal graticule line. This establishes the ground reference location.

**NOTE**

*If measurements are to be made relative to a voltage level other than ground, set the Input Coupling switch to DC instead, and apply the reference voltage to the input connector. Then position the trace to the reference (horizontal graticule) line.*

- Set the COUPLING switch to DC. Points on the waveform above the ground reference location are positive. Those points below are negative.

**NOTE**

*If using Channel 2, ensure that the Channel 2 Invert mode is not selected (INVERT button out).*

- If necessary, repeat Step 4 using a different reference line which allows the waveform in Step 5 to be displayed on screen.
- Adjust the A TRIGGER LEVEL control to obtain a stable display.
- Set the A SEC/DIV switch to a position that displays several cycles of the signal.
- Measure the divisions of vertical deflection between the ground reference line and the point on the waveform at which the level is to be determined (see Figure 2-13).
- Calculate the instantaneous voltage, using the following formula:

$$\text{Instantaneous Voltage} = \text{vertical deflection (divisions)} \times \text{polarity (+ or -)} \times \text{VOLTS/DIV switch setting indicated by 1X (or 10X PROBE when 10X probe is used)}$$

EXAMPLE: The measured vertical deflection from the reference line is 4.6 divisions (see Figure 2-13). The waveform point is above the reference line, A 10X attenuator probe is being used, and the VOLTS/DIV switch is set to 2 V (at 10X PROBE setting).

Substituting the given values:

$$\text{Instantaneous Voltage} = 4.6 \text{ div} \times (+1) \times 2 \text{ V/div} = 9.2 \text{ V.}$$




Figure 2-13. Instantaneous voltage measurement.

### Algebraic Addition

With the Vertical MODE switches set to BOTH and ADD, the waveform displayed is the algebraic sum of the signals applied to the Channel 1 and Channel 2 inputs (CH 1 + CH 2). If the Channel 2 INVERT mode is selected (INVERT button in) the waveform displayed is the difference between the signals applied to the Channel 1 and Channel 2 inputs (CH 1 minus CH 2). When both VOLTS/DIV switches are set to the same deflection factor, the deflection factor in the ADD mode is equal to the deflection factor indicated by either VOLTS/DIV switch.

The following general precautions should be observed when using the ADD mode.

- a. Do not exceed the input voltage rating of the oscilloscope.

- b. Do not apply signals that exceed the equivalent of about eight times the VOLTS/DIV switch settings, since large voltages may distort the display. For example, with a VOLTS/DIV switch setting of 0.5 V, the voltage applied to that channel should not exceed about 4 volts.
- c. Use Channel 1 and Channel 2 POSITION control settings which most nearly position the signal on each channel to midscreen, when viewed in either CH 1 or CH 2 Vertical MODE. This ensures the greatest dynamic range for ADD mode of operation.
- d. To attain similar response from each channel, set both the Channel 1 and Channel 2 Input Coupling switches to the same position.

### Common-Mode Rejection

The following procedure shows how to eliminate unwanted ac input-power frequency components. Similar methods could be used either to eliminate other unwanted frequency components or to provide a dc offset.

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal containing the unwanted line-frequency components to the CH 1 input connector.
3. Apply a line-frequency signal to the CH 2 input connector. To maximize cancellation, the signal applied to Channel 2 must be in phase with the unwanted line-frequency component on the Channel 1 input.
4. Select BOTH and ALT Vertical MODE and set both VOLTS/DIV switches to produce displays of approximately four or five divisions in amplitude.
5. Adjust the CH 2 VOLTS/DIV switch and CH 2 VOLTS/DIV Variable control so that the Channel 2 display is approximately the same amplitude as the undesired portion of the Channel 1 display (see Figure 2-14A).
6. Select ADD Vertical MODE and press in the INVERT button, and slightly readjust the CH 2 VOLTS/DIV Variable control for maximum cancellation of the undesired signal component (see Figure 2-14B).




Figure 2-14. Common-mode rejection.

### Amplitude Comparison (Ratio)

In some applications it may be necessary to establish a set of deflection factors other than those indicated by the VOLTS/DIV switch settings. This is useful for comparing unknown signals to a reference signal of known amplitude. To accomplish this, a reference signal of known amplitude is first set to an exact number of vertical divisions by adjusting the VOLTS/DIV switch and the variable control. Unknown signal can then be quickly and accurately compared with the reference signal without disturbing the setting of the VOLTS/DIV Variable control. This procedure is as follows:

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the reference signal to either vertical channel input and set the Vertical MODE switch to display the channel used.
3. Set the amplitude of the reference signal to five vertical divisions by adjusting the VOLTS/DIV Variable control.

4. Disconnect the reference signal and apply the unknown signal to be measured to the same channel input. Adjust the vertical position of the waveform so that its bottom edge just touches the 0% line on the screen.
5. Horizontally position the waveform so that its top most features cross the center vertical graticule line (see Figure 2-15).
6. Read the percent ratio directly from the graduations of the center line graticule line, referring to the 0% and 100% percentage marks on the left edge of the graticule (1 minor division equals 4% for a 5-division display).




Figure 2-15. Voltage ratios.

## TELEVISION DISPLAYS

### TV Line Signal

The following procedure is used to display a TV Line signal.

1. Preset instrument controls and set A TRIGGER mode to P-P AUTO/TV LINE.
2. Apply the TV signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set the appropriate VOLTS/DIV switch to display 0.35 division or more of composite sync signal.
4. Set the A SEC/DIV switch to 10  $\mu$ s.

5. For positive-going TV signal sync pulses, set the A TRIGGER SLOPE switch to positive (button out); for negative-going TV signal sync pulses, set the A TRIGGER SLOPE switch to negative (button in).

**NOTE**

*To examine a TV Line signal in more detail, either the X10 Magnifier or the Delayed-Sweep Magnification feature may be used.*

**TV Field Signal**

The television feature of the instrument can also be used to display TV Field signals.

1. Preset the instrument controls and obtain a baseline trace.
2. Select TV FIELD A TRIGGER mode (pushed both P-P AUTO and NORM buttons in) and set the A SEC/DIV switch to 2 ms.
3. To display a signal field, connect the TV signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
4. Set the appropriate VOLTS/DIV switch to display 0.35 divisions or more of composite sync signal.
5. For positive-going TV signal sync pulses, set the A TRIGGER SLOPE switch to positive (button out); for negative-going TV signal sync pulses, set the A TRIGGER SLOPE switch to negative (button in).
6. To change the TV field that is displayed, momentarily interrupt the trigger signal by setting the Input Coupling switch to GND and then back to AC until the desired field is displayed.

**NOTE**

*To examine a TV Field signal in more detail either the X10 Magnifier or the Delayed-Sweep Magnification feature may be used.*

7. To display a selected horizontal line, first trigger the sweep on a vertical (field) sync pulse, then use the Magnified Sweep Runs After Delay procedure in this part (steps 5 through 7) to magnify the selected horizontal line for a closer examination. This procedure is useful for examining Vertical Interval Test Signals (VITS).

8. To display either Field 1 or Field 2 individually, connect the TV signal to both CH 1 and CH 2 input connectors and select BOTH and ALT Vertical MODE.
9. Set the A SEC/DIV switch to 0.5 ms or faster sweep speed (displays of less than one full field). This will synchronize Channel 1 display to one field and Channel 2 to the other field.

**DELAYED-SWEEP MAGNIFICATION**

The delayed-sweep feature of the instrument can be used to provide higher apparent magnification than is provided by the X10 Magnifier switch. Apparent magnification occurs as a result of displaying a selected portion of the A trace at a faster sweep speed (B Sweep speed). The A SEC/DIV switch setting determines how often the B trace will be displayed. Since the B Sweep can occur only once for each A Sweep time duration sets the amount of time elapse between succeeding B Sweeps.

The intensified zone is an indication of both the location and length of the B sweep interval within the A Sweep interval. Positioning of the intensified zone, (i.e., setting the amount of time between start of the A Sweep and start of the B Sweep) is accomplished with the B DELAY TIME POSITION control. With either ALT or B Horizontal MODE selected and B TRIGGER LEVEL control set fully clockwise (RUNS AFTER DLY), the B DELAY TIME POSITION control provides continuously variable positioning of the start of the B Sweep. The range of this control is sufficient to place the B Sweep interval at most locations within the A Sweep interval. When ALT Horizontal Mode is selected, the B SEC/DIV switch setting determines the B Sweep speed and concurrently sets the length of the intensified zone on the A trace.

**NOTE**

*Gated FREQ, PER, WIDTH, or TOTALIZE functions alter the length of the intensified zone to indicate the gate interval established by the CTM section, when the B sweep is triggered. Therefore, in these cases the B sweep duration does not match that of the intensified zone.*

Using delayed-sweep magnification may produce a display with some slight horizontal movement (pulse jitter). Pulse jitter includes not only the inherent uncertainty of triggering the delayed sweep at exactly the same trigger point each time, but also jitter that may be present in the input signal. If pulse jitter needs to be

measured, use the Pulse Jitter Time Measurement procedure which follows the discussion of Magnified Sweep Runs After Delay.

### Magnified Sweep Runs After Delay

The following procedure explains how to operate the B Sweep in a nontriggered mode and to determine the resulting apparent magnification factor.

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set the appropriate VOLTS/DIV to produce a display of about 2 or 3 divisions in amplitude and center the display.
4. Set the A SEC/DIV switch to a sweep speed which displays at least one complete waveform cycle.
5. Select ALT Horizontal MODE. Adjust both the appropriate vertical POSITION control and the A/B SWP SEP control to display the A trace above the B trace.
6. Adjust the B DELAY TIME POSITION control to position the start of the intensified zone to the portion of the display to be magnified (see Figure 2-16).
7. Set the B SEC/DIV switch to a setting which intensifies the full portion of the A trace to be magnified. The intensified zone will be displayed as the B trace (see Figure 2-16). The B Horizontal MODE may also be used to magnify the intensified portion of the A Sweep.
8. The apparent sweep magnification can be calculated from the following formula:

$$\text{Apparent Delayed Sweep Magnification} = \frac{\text{A SEC/DIV switch setting}}{\text{B SEC/DIV switch setting}}$$




Figure 2-16. Delayed-sweep magnification.

EXAMPLE: Determine the apparent magnification of a display with an A SEC/DIV switch setting of 0.1 ms and a B SEC/DIV switch setting of 1 μs.

Substituting the given values:

$$\text{Apparent Magnification} = \frac{1 \times 10^{-4} \text{ s}}{1 \times 10^{-6} \text{ s}} = 10^2 = 100$$

### Pulse Jitter Time Measurement

To measure pulse jitter time:

1. Perform steps 1 through 7 of the preceding Magnified Sweep Runs After Delay procedure.
2. Referring to Figure 2-17, measure the difference between Point C and Point D in divisions and calculate the pulse jitter time using the following formula:

$$\text{Pulse Jitter Time} = \text{Horizontal difference (divisions)} \times \text{B SEC/DIV switch setting}$$




Figure 2-17. Pulse jitter.

### Triggered Magnified Sweep

The following procedure explains how to operate the B Sweep in a triggered mode and to determine the resulting apparent magnification factor. Operating the B Sweep in a triggered mode provides a more stable display, since the delayed display is triggered at the same point each time.

1. Perform steps 1 through 7 of the preceding Magnified Sweep Runs After Delay procedure.
2. Adjust the B TRIGGER LEVEL control so the intensified zone on the A trace is stable.

#### NOTE

*The intensified zone seen in the ALT Horizontal MODE display will move from trigger point to trigger point as the B DELAY TIME POSITION control is rotated.*

3. The apparent magnification factor can be calculated from the formula shown in step 8 of the Magnified Sweep Runs After Delay procedure.

## TIME AND FREQUENCY MEASUREMENTS

### Period

To measure period between repetitive pulses use the following procedure:

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Adjust the A TRIGGER LEVEL control to obtain a stable display.
4. Set the A SEC/DIV switch to a position that displays the point of interest within the graticule area.
5. Set UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in the PER button.
6. Read the period between repetitive pulses on the readout.

### Frequency

For frequency measurements, use the same method as previously described in Period procedure, except that in step 5, the FREQ button is pressed in along with the UPPER FUNCTIONS/LOWER FUNCTIONS switch. The readout will then display the frequency measurement.

### Width

When in WIDTH mode, the CTM measures the time interval between a point on the first slope of the selected polarity and the same point on the following slope of the opposite polarity. For waveforms with either slow rise time or fall time, width measurement can be made more accurately by adjusting the A TRIGGER LEVEL control so that the start of the trace is at the 50% point of the waveform. The A Trigger threshold affects the width measurement and should be set carefully.

#### NOTE

*The no A trig message is not used in this function, allowing indefinite display of measurement results.*

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set A TRIGGER Mode to NORM.
4. To measure positive half cycles of the input signal, set the A TRIGGER SLOPE switch to positive (button

out); to measure negative half cycles of the input signal, set the A TRIGGER SLOPE switch to negative (button in).

5. Adjust the A TRIGGER LEVEL control to trigger about the midpoint on the waveform.
6. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in the WIDTH button.
7. Read the waveform width on the readout.

### Time Interval On Single Waveform

The built-in  $\Delta$  Time function provides tremendous flexibility for making general timing measurements. An internal circuit generates two different delay times alternately which results in two intensified zones on the A Sweep. The  $\Delta$  TIME POSITION control adjusts the separation of the two intensified zones, and in conjunction with the B DELAY TIME POSITION control, allows the start of each B Sweep to be positioned at any desired waveform location. In RUNS AFTER DLY mode, the time interval is measured and displayed on the readout. This is a nontriggered  $\Delta$  time measurement.

In triggered  $\Delta$  Time measurement mode, the time interval to be measured is defined by similar pulse edges (i.e. positive or negative slopes). The B Sweep is triggered on those edges alternately and the time between the trigger events is measured and displayed on the readout. The intensified zones relocate to correspond with the starts of their respective B Sweeps. Because this mode eliminates oscilloscope delay time jitter from the measurement, over the starting (reference) point.

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set the appropriate VOLTS/DIV switch for a convenient amplitude display of the waveform.
4. Set the A SEC/DIV switch to display the measurement points of interest on the waveform.
5. Select the ALT Horizontal MODE and adjust both the appropriate vertical POSITION control and A/B SWP SEP control to display the A trace above the B trace.
6. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in the  $\Delta$  TIME button.

7. Set the B SEC/DIV switch to the fastest sweep that provides a usable (visible) intensified zone.
8. Adjust the B DELAY TIME POSITION control to move the first (left) intensified zone to cover the starting (reference) point.
9. Adjust the  $\Delta$  TIME POSITION control to move the second intensified zone to cover the ending (measurement) point.

### NOTE

*To perform a nontriggered measurement proceed to step 10; to perform a triggered measurement proceed to step 11.*

10. Adjust both B DELAY TIME POSITION and  $\Delta$  TIME POSITION controls until the points of interest on the two B traces intersect at any convenient vertical graticule line. Read the time interval on the readout.
11. Adjust the B TRIGGER LEVEL control counterclockwise to midrange position to trigger at the desired threshold on the waveform. Read the time interval on the readout.

### Rise Time

The following technique describes how to measure rise time between the 10% and 90% points of the low to high transition of the selected waveform using a nontriggered  $\Delta$  time measurement. Fall time is measured between the 90% and 10% points of the high to low transition of the waveform.

1. Preset the instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Set the appropriate VOLTS/DIV switch and variable control for an exact 5-division display.
4. Vertically position the trace so that the zero reference of the waveform touches the 0% graticule line and the top of the waveform touches the 100% graticule line.
5. Set the A SEC/DIV switch to spread out the transition of interest as much as possible.

6. Select ALT Horizontal MODE and set the B SEC/DIV switch to the fastest sweep speed that still allows the intensified zone to overlap the transition of interest.
7. Set UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in the  $\Delta$  TIME button.
8. Select B Horizontal MODE and adjust the B INTENSITY control as desired.
9. Rotate the  $\Delta$  TIME POSITION control to its fully counterclockwise position.
10. Adjust the B DELAY TIME POSITION control until the displayed waveform intersects the center vertical graticule line at the 10% point (see Figure 2-18, Point A).
11. Adjust the  $\Delta$  TIME POSITION control so that as the waveform splits into two, the left most one (corresponding to the longer of the two delay times) intersects the center vertical graticule line at the 90% point (see Figure 2-18, Point B).
12. Read the rise time on the readout.




Figure 2-18. Rise time.

**NOTE**

*To measure fall time, the procedure is the same except that the right-most transition display is positioned (by the B DELAY TIME POSITION control) so that it intersects the center vertical graticule line at the 90% point. The left-most transition is positioned (by the  $\Delta$  TIME POSITION control) so that it intersects the center vertical graticule line at the 10% point.*

**Time Difference Between Pulses On Two Time-Related Signals (Short Method)**

Time difference between two time related pulses can often be measured using the triggered delay time technique. This is accomplished by triggering the A Sweep externally on the reference pulses, and the B Sweep on the pulses whose timing relationship to the reference is to be determined. The measurement is performed from the event triggering the A Sweep to the one triggering the B Sweep. This is possible when the time between the pulses to be measured exceeds 500 ns.

1. Preset the instrument controls and obtain a baseline trace.
2. Set the A & B SOURCE switch to A EXT, UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT, and press in the DLY TIME button.
3. Apply the reference signal to EXT INPUT connector and the comparison signal to CH 1 or CH 2 input connector.
4. Press and hold in the TRIG VIEW button, adjust the A TRIGGER LEVEL control to obtain a stable display, triggered on the appropriate edge of the reference signal.
5. Select CH 1 or CH 2 Vertical MODE and set the appropriate VOLTS/DIV switch to produce a display amplitude of 3 divisions.
6. Select ALT Horizontal MODE and rotate the B TRIGGER LEVEL control to its fully clockwise position. Position the intensified zone to cover the pulse to be measured and adjust the length of the zone with the B TIME/DIV switch.
7. Select the appropriate B TRIGGER SLOPE, and rotate the B TRIGGER LEVEL control counterclockwise to the midrange position, and trigger on the transition to be measured.
8. Read the time from the reference edge to the edge displayed by the B Sweep on the display.

### Time Difference Between Points On Two Time-Related Signals (General Method)

Time difference between pulses on two time-related signals can also be measured in a manner similar to that previously described for measuring time difference between two time related pulses on single waveforms using the delta time measurement procedure.

Increased timing accuracy may be obtained by externally triggering the A Sweep, while internally triggering the B Sweep alternately from Channel 1 and Channel 2 (triggered delta time). The following procedure describes how to measure from a point of interest on the Channel 1 display to a point of interest on the Channel 2 display, with or without the benefit of triggered delta time.

1. Preset the instrument controls and obtain a baseline trace. For nontriggered mode, set the A & B SOURCE switch to CH 1, and for triggered mode set A & B SOURCE switch to A EXT.
2. Using probes or cables having equal time delays, apply the reference signal to the CH 1 input connector and apply the comparison signal to the CH 2 input connector. For triggered measurement, the signal to EXT INPUT connector must also be applied to the CH 1 input connector.
3. Adjust the A TRIGGER LEVEL control to obtain a stable display.
4. Select BOTH and ALT Vertical MODE, and set both VOLTS/DIV switches to produce a display amplitude of 3 divisions.
5. Select ALT Horizontal MODE. The first intensified zone (reference) will appear on the Channel 1 trace and the second intensified zone (comparison) will appear on the Channel 2 trace.
6. Adjust Channel 1 POSITION, Channel 2 POSITION, and A/B SWP SEP controls to display the A traces above the B traces.
7. Set the A SEC/DIV switch to display the measurement points of interest within the graticule area.
8. Set UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in the  $\Delta$  TIME button.
9. For the most accurate measurement, set the B SEC/DIV switch to the fastest sweep speed that provides a usable (visible) intensified zone.

10. Adjust the B DELAY TIME POSITION control to move the reference zone to the reference signal on the CH 1 A trace (see Figure 2-19).
11. Adjust the  $\Delta$  TIME POSITION control to move the comparison zone to the comparison signal on the CH 2 A trace (see Figure 2-19).
12. To perform a nontriggered measurement proceed to step 13, and to perform a triggered measurement proceed to step 14.
13. Adjust both B DELAY TIME POSITION and  $\Delta$  TIME POSITION controls until the points of interest on the two B traces intersect at any convenient vertical graticule line. Read the time difference on the readout.
14. Adjust both B TRIGGER LEVEL controls counter-clockwise to midrange position to trigger at the desired threshold of the waveform.




Figure 2-19. Time difference between points on two time-related signals in nontriggered mode.

### Phase Difference

The phase difference between two signals of equal frequency is determined in a manner similar to that described in the preceding procedure for measuring time difference between two time-related pulses. To

measure the phase difference between two time-related signals, use the following procedure:

1. Perform steps 1 through 14 of the preceding Time Difference Between Two Time-related Pulses procedure.
2. Record the time difference on the readout.
3. Set the A & B SOURCE switch to CH 1.
4. Select A Horizontal MODE and set UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in the PER button.
5. Read the period of the two signals on the readout, and calculate the phase difference using the following formula:

$$\text{Phase Difference} = 360 \times \frac{\text{time difference}}{\text{period}}$$

### Gated Measurements

Gated measurements of frequency, period, width, and totalize are performed through the B Trigger circuitry. When the B Sweep is properly triggered, the intensified zone duration indicates the gate interval over which the CTM samples the displayed waveform, rather than the duration of B Sweep. The intensified zone blinks at the measurement cycle rate (except when in totalize mode) to indicate that the B Sweep is triggered, and that the CTM is receiving the B Trigger signal.

In gated frequency and period mode, the intensified zone will completely cover one period of the input signal (two consecutive trigger edges) if the B Sweep duration is less than that period. The CTM extends the intensified zone to cover this interval when the B Sweep is triggered. If the B Sweep duration is actually longer than one period of the input signal, all consecutive periods that fall within the B Sweep duration will be included in the gated measurement. In frequency mode the CTM will display the reciprocal of the average of all periods with the blinking intensified zone. In period mode the CTM will display the arithmetic average of all periods with the blinking intensified zone.

In gated width mode, the intensified will cover one pulse width on the input signal, if the B Sweep duration is less than that width. If the B Sweep duration covers more than one pulse width, then the CTM will display the arithmetic

average of all the pulse widths inside the blinking intensified zone. Gated width does not display the **no b trig** status message.

The accuracy in gated frequency, period, and width measurements improves as the number of samples within the gate interval increases. It is advantageous to lengthen the intensified zone to include as many trigger events as possible within the gating interval.

In gated totalize mode, one event is counted for each period of input signal inside the intensified zone (the stop trigger edge is not counted).

The following procedure describes how gated measurements are accomplished.

1. Preset instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the Vertical MODE switch to display the channel used.
3. Adjust the A TRIGGER LEVEL control to obtain a stable display.
4. Set the A SEC/DIV switch to a position that displays the point of interest within the graticule area.
5. Select ALT Horizontal MODE, rotate the B TRIGGER LEVEL control to its fully clockwise position and adjust both the appropriate vertical POSITION control and the A/B SWP SEP control to display the A trace above the B trace.
6. Set the B SEC/DIV switch to the desired sweep speed and adjust the B DELAY TIME POSITION control to move the intensified zone to that portion of the display to be measured.
7. To select the desired gated measurement function (frequency, period, width, or totalize), set UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in **FREQ**, **PER**, **WIDTH**, or **TOTALIZE** (**FREQ** and **PER**) buttons.
8. Adjust the B TRIGGER LEVEL control for a stable display.
9. Observe that the intensified zone moves, covering the interval that the CTM is sampling.
10. Read the measurement on the readout.

## MULTIMETER MEASUREMENTS

### Resistance and Semiconductor Junctions

Before making resistance measurements, ensure that the readout displays the message OPEN. If the device under test is a linear resistance, the CTM will display the resistance on the readout. If the resistance is a nonlinear device (semiconductor), the CTM will automatically enter into the junction test mode and display the message **diode** on the readout and at the same time generate an audible tone. After a brief moment the CTM will display the forward voltage drop of the semiconductor on the readout. The red lead sources the measuring voltage (+2.5 V in series with 1 kΩ). To measure resistance or forward voltage drop of a semiconductor use, the following procedure:

1. Set UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in the ohms button.
2. Connect the multimeter leads across the unknown resistance or across the semiconductor as indicated in Table 2-3. The diode detection circuitry will be activated only when the readout displays the message OPEN before the multimeter leads are connected across the device under test.
3. Read the resistance value or the forward voltage drop of the device on the readout.

### Determining Semiconductor Leakage Current

The Multimeter performs resistance measurements with divider technique. A reference voltage of about +2.5 volts is applied through an internal precision resistor (reference resistance) to the Multimeter + (positive) connector (see Figure 2-20).

The unknown resistance is calculated from the following formula:

$$\text{Unknown Resistance} = \frac{\text{reference resistance}}{\text{reference resistance} + 2.5 \text{ volts}} \times \frac{V_{\text{terminal}}}{-V_{\text{terminal}}}$$

A typical semiconductor will have a reverse leakage current vs reverse voltage drop as shown in Figure 2-21. Observe that the typical semiconductor leakage current is relatively constant as the reverse voltage drop increases.

To measure the reverse-leakage current, perform a normal resistance measurement with the Multimeter + (positive) connector connected to the cathode of the device (reverse biased).




Figure 2-20. Resistance measurement technique.

Table 2-3  
Multimeter Connections to  
Forward Bias Semiconductors

Semiconductor	Red Lead	Black Lead	Typical Forward Drop (volts)
Silicon Diode	Anode	Cathode	0.65
Silicon NPN Transistor	Base	Collector or Emitter	0.70
Silicon PNP Transistor	Collector or Emitter	Base	0.70
N Channel FET	Gate	Drain or Source	0.80
LED	Anode	Cathode	1.5 to 2.0




Figure 2-21. Reverse voltage drop to reverse leakage current characteristics for typical semiconductor.

Calculate the reverse-leakage, using the following formula:

$$\text{Reverse-Leakage Current} = \frac{\text{reference voltage (+ 2.5 volts)}}{\frac{\text{reference resistance}}{\text{resistance displayed}} + 1}$$

(see Table 2-4) on the readout

Table 2-4  
Reference Resistance Values

Resistance Displayed On Readout	Reference Resistance
0 to 50 Ω	1 kΩ
50 Ω to 500 Ω	1 kΩ
50 Ω to 5 kΩ	10 kΩ
5 kΩ to 50 kΩ	100 kΩ
50 kΩ to 500 kΩ	1 MΩ
Greater than 500 kΩ	10 MΩ

### Continuity

Continuity checks are made by connecting the test leads across the device under test and measuring the unknown resistance. If the resistance is greater than or equal to 5 Ω the readout will display the message OPEN;

if the resistance is less than 5 Ω an audible tone is generated and the readout will display the message Short.

1. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in both AC RMSV and TEMP buttons.
2. Connect the test leads across the device under test.
3. Read the readout and listen for an audible tone.

### AC RMS Volts



If the readout displays the message **OUCH** (input voltage equals or exceeds 350 V), immediately disconnect the test leads from the unit-under-test to prevent possible instrument damage.

1. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in the AC RMSV button.
2. Connect the - (negative) lead to the reference point (usually a ground or test point) and connect the + (positive) lead to the unknown voltage to be measured.
3. Read the rms value of the input signal on the readout.

### DC Volts



If the readout displays the message **OUCH** (input voltage equals or exceeds 500 V), immediately disconnect the test leads from the unit-under-test to prevent possible instrument damage.

1. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to IN and press in the DCV button.
2. Connect the - (negative) lead to the reference point (usually a ground or test point) and connect the + (positive) lead to the unknown voltage to be measured.
3. Read the dc voltage on the readout.

## Channel 1 Volts

The instrument has the capability to measure the dc and true ac coupled rms values of the input signal through CH 1 OR X & DMM connector. Maximum voltage measurement when using either a 1X probe or cable is 49.99 V (dc or ac rms). Maximum voltage measurement with a 10X probe is 499.9 V dc and 349.9 V rms ac. When an overrange condition exists the readout will display the message **OUCH**. The procedure for measuring voltage through the CH 1 OR X & DMM input connector is as follows:

1. Preset instrument controls and obtain a baseline trace.
2. Connect either a probe (1X or 10X) or a cable to CH 1 OR X & DMM input connector.
3. Set the Channel 1 input Coupling switch to either RMSV AC position for ac rms voltage measurement or to DCV position for dc voltage measurement.
4. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT and press in the CH1 button.
5. Connect the probe ground clip to a reference point (usually a ground or test point) and connect the probe tip to the unknown voltage to be measured.
6. Observe the readout. If necessary, select the next lower CH 1 VOLTS/DIV position to obtain maximum resolution without overranging.

## NOTE

*CH 1 AC RMS voltage measurements are influenced above 1 KHz by compensation when a X10 probe is used. Below 100 Hz the CH 1 AC coupling time constant degrades direct measurements. Use of a 10X probe increases the time constant and extends the useful range to 20 Hz. CH 1 dc voltage measurements are made without the benefit of a two-pole low pass input filter. Dc voltage measurements made through the side inputs utilize such a filter and therefore offer better normal mode rejection (rejection of input signal noise).*

## Temperature

1. Connect the P6602 temperature probe to the Multimeter input connector.
2. Set the UPPER FUNCTIONS/LOWER FUNCTIONS switch to OUT.
3. Press in the TEMP button for temperature measurement in Celsius or press both the TEMP and TIME buttons for temperature measurement in Fahrenheit.
4. Read the temperature on the readout.

