INSTRUCTION

MODIFICATION INSERT

Serial Number

561B/R561B MOD 272C

This insert is provided as a supplement to the instruction manual furnished with this modified instrument. The information given in this insert supersedes that given in the manual.

This insert material Copyright © 1973 by Tektronix, Inc., Beaverton, Oregon. Printed in the United States of America. All rights reserved. Contents of this insert may not be reproduced in any form without permission of the copyright owner.

U.S.A. and foreign Tektronix products covered by U.S. and foreign patents and/or patents pending.

561B/R561B MOD 272C

561B/R561B MOD 272C

This manual insert describes the features of MOD 272C as it pertains to the 561B/R561B Oscilloscope. It provides a special 10 X 10 division internal graticule, deletion of the front panel and knobs, and changes in the accessories.

CALIBRATION

Calibrate using the standard procedure with the exception of Calibrator and Current Loop.

			PARTS LIST			
			Electrical			
			Transistors			
Q151 Q159 Q162	Delete Delete Delete					
			Switches			
			DWITCHES			
SW150	Delete					
]	Electron Tube			
V249	Change	037-403	35-00	T5611-31-9.14		
			Mechanical			
			rechanical			
Connecto	or, BNC			Delete	1	131-0274-00
	arrent loop			Delete	1	214-0335-00
Spacer,	current loop			Delete	1	361-0059-00
Switch,	Calibrator			Delete	1	262-0839-00
*Panel,				Delete	1	333-1081-00
	gray, FOCUS			Delete	1	366-0153-00
	gray, SCALE ILLUM			Delete	1	366-0236-00
	gray, INTENSITY			Delete	1	366-0236-00
	gray, CALIBRATOR			Delete	1	366-0322-01
**Panel	front			Delete	1	333-1078-01

Delete

Delete

Delete

Delete.

Delete

1

1

1

2

366-0220-00

366-0220-00

366-0220-00

366-0220-00

390-0046-02

**Knob, gray, INTENSITY

**Knob, gray, FOCUS

**Cabinet, side

**Knob, gray, SCALE ILLUM

**Knob, gray, ASTIGMATISM

561B/R561B MOD 272C

Standard Accessories

Cord, patch
Jack, BNC post
*Hardware Kit
Adapter, 3-2 wire

012-0087-00
012-0092-00
016-0131-00
103-0013-00

*R561B only

**561B only