

Part No. 070-6815-00 Product Group 47

# 067–1341–00 Risetime Limiter Calibration Fixture


## **Description**

The 067-1341-00 Risetime Limiter Calibration Fixture is used with the 067-0608-00 Overdrive Recovery Calibration Fixture when testing the overdrive recovery capability of the 11A33 Differential Comparator. The 067-1341-00 has no replaceable parts.

# **Operation**

The Risetime Limiter Calibration Fixture limits input signal transition time by adding 267  $\Omega$  of resistance in series with the 50  $\Omega$  output impedance of the Overdrive Recovery Calibration Fixture. The input time constant, resulting from 317  $\Omega$  of combined calibration fixture impedance and 15 pF of 11A33 input capacitance, limits the transition time of input signals to approximately 1 V/ns.

The output of the 067-1341-00 Risetime Limiter Calibration Fixture is connected directly to the input of the 11A33; the output of the 067-0608-00 Overdrive Recovery Calibration Fixture is connected to the input of the Risetime Limiter Calibration Fixture. The following diagram shows the interconnection of the 067-0608-00, 067-1341-00, and 11A33.



Interconnection of the Overdrive Recovery Calibration Fixture, Risetime Limiter Calibration Fixture, and 11A33 Differential Comparator.

Instruction Sheet

Checkout Procedure The Checkout Procedure checks the internal resistance of the 067-1341-00 Risetime Limiter.

## Test Equipment

Test equipment required to perform the Checkout Procedure is listed in the following table.

Description	Minimum Specification and Use
Digital Multimeter	0.1% Accuracy; 5-digit display.
	Used to measure the internal resistance of the 067-1341-00.
Test Leads	One lead with a male BNC connector; one with a female BNC; the remaining end of each is compatible with the Digital Multimeter connectors.
	Used to connect the Digital Multi- meter connectors to the 067-1341-00 connectors.

### **Procedure**

- 1. Connect the Digital Multimeter to the input and output connectors of the 067-1341-00 Risetime Limiter with the test leads.
- 2. Set the Digital Multimeter to the range providing maximum accuracy for measurement of 267  $\Omega$ .
- 3. Check the Digital Multimeter for a reading of 267  $\Omega \pm 1\%$ .