

INSTRUCTION MANUAL

FILE COPY

Serial Number _____

MODIFICATION INSERT

TYPE 310A
MOD 815E

This manual insert describes the features of MOD 815E as installed in the Type 310A Oscilloscope. The instrument is modified by replacing the EXT TRIG IN and the CAL OUT binding posts with BNC connectors.

CALIBRATION

This modification does not affect the operation of the instrument. Therefore, it can be calibrated as directed in the Instruction Manual.

PARTS LIST

The following changes should be made to the appropriate parts list for this modified instrument. When ordering replacement parts, specify instrument type, serial number, and MOD number. Include the part number and description of the desired item.

CONNECTOR, BNC, 1/2" "D" mtg. Change 2 131-0126-00

This insert material Copyright © 1969 by Tektronix, Inc., Beaverton, Oregon. Printed in the United States of America. All rights reserved. Contents of this insert may not be reproduced in any form without permission of the copyright owner.

TYPE 310A

MOD 815E

Tektronix, Inc.

S.W. Millikan Way • P. O. Box 500 • Beaverton, Oregon 97005 • Phone 644-0161 • Cables: Tektronix

06099

be