LOW-LEVEL DIFFERENTIAL AC PREAMP TYPE E

INSTRUCTION MANUAL




TEKTRONIX, INC.

MANUFACTURERS OF CATHODE-RAY AND VIDEO TEST INSTRUMENTS

Sunset Highway and Barnes Road • P. O. Box 831 • Portland 7, Oregon, U. S. A.

Phone: CYpress 2-2611 • Cables: Tektronix



GENERAL DESCRIPTION

General

The Type E Plug-in Unit is a high-gain adjustable bandwidth, ac-coupled preamplifier. Calibrated vertical sensitivities to 50 microvolts per centimeter in conjunction with differential input extended the capabilities of the 530- and 540-Series Oscilloscopes into a new region of usefulness.

Characteristics

SIGNAL INPUT

Both grids of the push-pull input stage and ground are brought out to the 3-pin INPUT connector. Thus, single-ended or differential input may be used by merely making the proper grid and ground connections.

SENSITIVITY

.05 millivolts/cm to 10 millivolts/cm in eight calibrated positions. Continuously variable between the fixed positions and to 25 millivolts/cm.

FREQUENCY RESPONSE

Adjustable high- and low-frequency response having the following 3-db points:

High: 50 cps, 250 cps, 1 kc, 10 kc, 20 to 60 kc Low: 0.06 cps, 0.2 cps, 0.8 cps, 8 cps, 80 cps

PHYSICAL CHARACTERISTICS

Construction, aluminum alloy chassis; finish, photo-etched, anodized panel; weight, 4½ lbs.

Functions of Controls and Connectors

INPUT	3-pin connector for signal input connections. Pin 1, ground. Pin 2, positive input
	produces negative output. Pin 3, positive input produces positive output.

VERTICAL	Potentiometer to move the trace vertically.
POSITION	

MILLIVOLTS/CM Eight-position switch to select the calibrated vertical deflect	ction sensitivities.
---	----------------------

VARIABLE	Control concentric with the MILLIVOLTS/CM switch for varying the deflec-
	tion sensitivity between the fixed steps and to 25 millivolts/cm.

DIFF. BAL.	Screwdriver-adjust control used to obtain maximum in-phase signal rejection
	when using differential input.

"VARIABLE"	Screwdriver-adjust control to balance the attenuator stage dc levels so the trace
BALANCE	does not shift position as the VARIABLE control is rotated.

GAIN ADJUST	Screwdriver-adjust control used to set the gain of the preamplifier so the crt
_	deflection will agree with the MILLIVOLTS/CM sensitivities.

TRACE	This push button switch, when operated, discharges the last set of coupling capaci-
RESTORER	tors and returns the trace to the face of the crt.

APPROXIMATE FREQUENCY	Concentric pass band.	switches	to sel	ect the	upper	and	lower	3 db	points	of	the	amplifier
RESPONSE												



OPERATING INSTRUCTIONS

NOTE: Remove the shipping blocks from the shock mounts before operating this preamp.

General

Allow the Type E Plug-In Unit to warm up for several minutes before attempting to use it. This is a very high gain unit and the initial drift will probably keep the trace off the crt screen for a short time. This initial drift can be minimized by setting the red FREQUENCY RESPONSE knob to 80. If, after the preamp has stabilized, a switching transient or bad connection causes the trace to be deflected by operating the TRACE RESTORER push button.

Signal Input Connections

This preamp is designed for single-ended or differential input. Each of the grids of the balanced input stage is brought out to the front panel to a 3-pin connector labeled INPUT. Pins 2 and 3 are the signal connections and pin 1 is the chassis ground. Matching female connectors are available through the Tektronix Customer Service Department. Order by Tektronix Part Number 012-022. The input impedance of either grid is 10 megohms shunted by 50 $\mu\mu$ f. The grid-to-grid impedance is 20 megohms shunted by 50 $\mu\mu$ f.

There are no blocking capacitors in the input connections. If signals having a dc component greater than two volts are connected to the preamp, an external blocking capacitor must be used.

Single-Ended Input

Either of the two grids can be used for single-ended input. A positive-going signal on pin 3 of the INPUT connector will produce a normal display on the crt screen. Signals connected to pin 2 will produce an inverted display. In either case, the unused signal grid must be grounded. Normally this should be done by connecting the unused pin directly to pin 1 at the INPUT connector. However, if there is hum introduced in the connecting cable, this can sometimes be reduced by grounding the unused grid at the input to the cable.

Differential Input

In many applications the desired signal is superimposed on an undesired signal such as

line-frequency hum, etc. By use of differential input you can, in many cases, improve the ratio of the desired to undesired signal. If you can make a connection to the pin-3 input with both the desired and undesired signals and to the pin-2 input with only the undesired signal, the undesired signal will be attenuated by outphasing. If the DIFF. BAL. control is properly adjusted and the undesired signals are exactly in phase, the undesired signal will be reduced by a ratio of about 50,000 to 1. The maximum common-mode signal amplitude that can be satisfactorily attenuated is about two volts, peak to peak.

By using the same connection you can display push-pull signals. Again you should observe the proper input polarity to avoid an inverted display.

Frequency Response

In addition to the differential rejection of unwanted signals, many times an undesired signal can be attenuated by varying the frequency response of the unit. An example of this would be a low-frequency sinusoidal waveform that is modulated by high-frequency noise. By reducing the high-frequency response the noise can, in many cases, be almost completely eliminated without distorting the desired signal. Likewise you might filter out undesired line-frequency hum by restricting the low-frequency response of the unit. Be careful not to distort non-sinusoidal waveforms by a too severe restriction of the amplifier's pass band.

Deflection Sensitivity and Gain Adjust

The MILLIVOLTS/CM switch provides eight fixed deflection sensitivities. The maximum frequency response is limited in the three most sensitive positions as indicated, decreasing to 20 kc at .05 millivolts/cm. The VARIABLE control which is concentric with the MILLI-VOLTS/CM switch, must be turned full right to the CALIBRATED position if you are to read signal amplitudes from the ruled graticule.

The gain of the unit should be checked periodically and, if necessary, readjusted to compensate for tube aging. Be sure the VARIABLE control is turned full right to the CALIBRAT-ED position and display a signal of known amplitude such as the oscilloscope's calibrator waveform. Set the GAIN ADJUST control so the displayed signal amplitude agrees with the deflection sensitivity indicated by the MILLI-VOLTS/CM switch.



"Variable" Balance

As tubes age, their characteristics change, particularly their static characteristics. The VARIABLE gain control is connected between the cathodes of the tubes in the gain-control stage. Any difference in dc level between the two cathodes will cause the trace to shift vertically as the VARIABLE control is rotated. The "VARIABLE" BALANCE control is used to set the two cathodes to the same dc level to

remedy this condition. Rotate the VARIABLE control back and forth and adjust the "VARIABLE" BALANCE to the point where the trace does not shift position. If the "VARIABLE" BALANCE control is very far out of adjustment you will find it easier to restrict the low-frequency response of the unit and make a rough adjustment and then increase the low-frequency response to maximum and make a final fine adjustment.

CIRCUIT DESCRIPTION

General

The Type E Plug-In Unit has a maximum pass band of a fraction of a cycle to 60 kc. It consists of three stages of amplification with output cathode followers to drive the interconnecting plug. Voltage-regulator tubes further stablize the regulated-supply voltages applied to the input-amplifier and gain-control stages.

Input Amplifier

The input amplifier, V3015, is a cathode-coupled phase-inverter amplifier. The DIFF. BAL. control in the cathodes provides a means of balancing this stage for maximum rejection of in-phase signals on the grids. C3155, in one plate lead, balances the stray capacitance to ground in the plate leads for better high-frequency differential balance.

Gain-Control Stage

The gain-control stage, V3205 and V3265, provides additional gain and gain control. The MILLIVOLTS/CM control selects fixed resistors for the plate loads of this stage to adjust the gain in fixed steps. The VARIABLE control, R3225, varies the gain of the stage between the fixed steps, and the GAIN ADJUST control sets the amplifier gain to agree with the front panel. The "VARIABLE" BALANCE control allows this stage to be balanced so there is no shift as the VARIABLE control is rotated.

The LOW frequency response switch selects coupling capacitors and grid resistors to select the desired low-frequency 3-db point.

Positioning Amplifier

The positioning amplifier stage, V3805, provides additional gain and a means of positioning the trace vertically. The VERTICAL POSITION control, R3805, varies the current through each half of the balanced amplifier. Since the ampifier is dc coupled from this point on, this determines the average vertical position of the trace on the screen. The VERT. POS. RANGE control allows the balance to be adjusted so that the trace is centered when the VERTICAL POSITION control is centered.

The HIGH frequency response switch selects shunt capacitors to limit the high-frequency response of the preamp. In the 20-60 KC position the switch contacts remain open for maximum frequency response.

The output cathode follower, V2905, provides the necessary low impedance to drive the interconnecting plug and oscilloscope input amplifier.

Regulated Supplies

Regulated supplies in the oscilloscope supply all voltages used in the preamp including the heater voltage. Voltage-regulator tubes, V4005 and V4015 provide additional decoupling between the preamp and the oscilloscope.



MAINTENANCE

Replacement of Components

Tektronix will supply replacement components at current net prices. However, since most of the components are standard electronic and radio parts we suggest you get them from your local dealer if you can. Be sure to consult your instruction manual first to see what tolerances are required.

We specially select some of the components, whose values must fall within prescribed limits, by sorting through our regular stocks. The components so selected will have standard RETMA color-code marks showing the values and tolerances of the stock they were selected from, but they will not in general be replaceable from dealers stocks.

Such selected parts, as well as the parts we manufacture at Tektronix, are identified in the parts lists either by notes or by our own stock numbers. Order these parts from the Tektronix factory in Portland, Oregon.

Parts-Ordering Information

You will find a serial number on the frontispiece of this manual. This is the serial number of the instrument the manual was prepared for. Be sure the manual number matches the number of the instrument when you order parts.

A Tektronix instruction manual usually contains hand-made changes to diagrams and parts lists, and sometimes text. These changes are in general only appropriate to the instrument whose serial number appears on the manual frontispiece. The hand-made changes show changes to the instrument that have been made after the printing of the manual.

We make some of the instrument changes during the factory test procedure. Our technicians hand-tailor the circuits, if it seems appropriate, to provide the widest possible latitude of operation. Other changes are made to include the latest circuit improvements as they are developed in our engineering department, or when improved components become available. In any event, the changes are to your benefit. We have tried to give you the best instrument we can.

Soldering Precaution

The solder used on the ceramic terminals in this instrument must contain a small percentage of silver. If for any reason you resolder, be sure the solder you use contains silver. Silver-bearing solder is used in printed-circuit techniques, and is therefore available from all solder manufacturers. Repeated use of ordinary tin-lead solder will dissolve the fused bond of silver that makes the solder adhere to the porcelain, especially if the soldering iron is quite hot.

Trouble Shooting

Most troubles are caused by tube failure so this is the first thing to suspect in trouble shooting. If the trace cannot be centered vertically check to see if the unit is firmly plugged in. Be sure there has been adequate time for the tubes to warm up completely. Then, with the TRACE RESTORER button depressed, turn the VERT. POS. RANGE control through its range. If the trace cannot be centered, try changing V3805 and V3905.

If the trace can be centered with the TRACE RESTORER button depressed, but drifts off the screen upon its release, turn the MILLI-VOLTS/CM switch to 10. If this allows the trace to be centered try changing V3205 and V3265 or possibly V3015. If not, try changing V3805 or check the coupling capacitors. No simple formula can be given for finding all troubles, but the Circuit Description section of this manual may help.

Adjustments

Allow the unit to warm up for at least five minutes before making any adjustments. All adjustments can be made with the preamp plugged into an oscilloscope that has its panels removed or if the Type EP53 Plug-in Extension is used.

Positioning Adjustment

Center the VERTICAL POSÍTION control and, with the oscilloscope sweep running recurrently, adjust the VERT. POS. RANGE control to center the trace.

Gain-Adjustment

Connect a 20-millivolt signal from the oscilloscope calibrator to the INPUT connector, pin 3. Ground pin 2. Set the MILLIVOLTS/CM control to 5 and turn the VARIABLE control clockwise to the CALIBRATED position. Adjust the front-panel GAIN ADJUST control to obtain four centimeters of deflection.



"Variable" Balance Adjustment

With the oscilloscope sweep free running and no signal connected to the input, rotate the VARIABLE control, concentric with the MILLIVOLTS/CM knob, back and forth. If necessary, adjust the "VARIABLE" BALANCE control until there is no deflection of the trace as the VARIABLE control is rotated.

Differential-Balance Adjustment

Three adjustments are made to provide maximum rejection of common-mode signals.

a. Connect a 2-volt signal from the oscilloscope calibrator to pins 2 and 3 of the IN-PUT connector.

> PINI- FROUND PINZ - BLACK PINZ - RED

- b. Set the MILLIVOLTS/CM control to .05.
- c. Preset the DIFF. BAL. control to the center of its range.
- d. Adjust C3095 and C3155 just above the HEATER BAL. control for minimum amplitude of the spikes.
- e. Turn the HEATER BAL. control about one-eighth turn in either direction and wait until the drifting has stopped. If the balance is improved, continue moving the heater balance in the same direction. If the balance deteriorated move the heater balance in the opposite direction. Continue this pocess until optimum balance is achieved.
- f. Fine Balance is accomplished by varying the DIFF, BAL, control.




MODIFICATION NOTICE

CLIP-MOUNTED CERAMIC STRIPS

YOUR INSTRUMENT MAY BE EQUIPPED WITH CLIP-MOUNTED CERAMIC STRIPS. IF YOU FIND IT NECESSARY TO ORDER THESE STRIPS FOR REPLACEMENT, BE SURE TO CONSULT THIS SHEET. INCLUDE A DESCRIPTION OF THE PART, PART NUMBER, INSTRUMENT TYPE AND SERIAL NUMBER.

CERAMIC STRIP PARTS LIST

	PART NUMBER
STUD, CLIP, MOLDED NYLON	355046
SPACER, MOLDED NYLON, 5/32" HEIGHT	361-007
SPACER, MOLDED NYLON, 1/4" HEIGHT	361008
SPACER, MOLDED NYLON, 3/8" HEIGHT	361009
CERAMIC STRIP, 7/16" BY 3 NOTCHES	124092
CERAMIC STRIP, 7/16" BY 5 NOTCHES	124093
CERAMIC STRIP, 7/16" BY 7 NOTCHES	124094
CERAMIC STRIP, 7/16" BY 9 NOTCHES	124-4095
CERAMIC STRIP, 7/16" BY 11 NOTCHES	124-106
CERAMIC STRIP, 3/4" BY 1 NOTCH	124-100
CERAMIC STRIP, 3/4" BY 2 NOTCHES	124-086
CERAMIC STRIP, 3/4" BY 3 NOTCHES	124-087
CERAMIC STRIP, 3/4" BY 4 NOTCHES	124088
CERAMIC STRIP, 3/4" BY 7 NOTCHES	124089
CERAMIC STRIP, 3/4" BY 9 NOTCHES	124-090
CERAMIC STRIP. 3/4" BY 11 NOTCHES	124091



CERAMIC STRIPS AND MOUNTINGS USED IN / / TEKTRONIX EQUIPMENT.

ABBREVIATIONS

Cer.	ceramic	m	milli or 10 ⁻³
Comp.	composition	Ω	ohm
EMC	electrolytic, metal cased	Poly.	polystyrene
EMT	electrolytic, metal tubular	Prec.	precision
f	farad	PT.	paper tubular
h	henry	Tub.	tubular
k	kilohm or 10 ^a ohms	v	working volts dc
meg	megohm or 106 ohms	Var.	variable
μ	micro or 10 ⁻⁴	w	watt
μμ	micromicro or 10 ⁻¹²	ww	wire wound
••	GMV guaranteed m	inimum va	

LOW-LEVEL DIFFERENTIAL PREAMP


Capacitors 1						
C3095 1.5-7 μμf C3155 1.5-7 μμf C3165A,B .5 μf	Cer. Var. Cer. Var. Special	500 v 500 v	±1/4%	281005 281005 295060		
C3605A .5 μf C3605B .1 μf C3605C .01 μf C3605D .001 μf	Special — ma Tektronix. Of from factory.	Order replac		291010		
C3705A .5 μf C3705B .1 μf C3705C .01 μf C3705D .001 μf	Special — ma Tektronix. Of from factory.	Order replac		291010		
C3805 200 μμf C3815 .0022 μf C3825 .01 μf C3835 .047 μf C3845 22 μμf	Mica Fixed PT Fixed PT Fixed PT Fixed Cer. Fixed	500 v 400 v 400 v 400 v	10% 20% 2%+20% 20%	283511 285543 285510 285519 281511		
C4005 .1 μf C4015 .1 μf	PT Fixed PT Fixed	400 v 400 v	20% 20%	285526 285526		
		Re	sistors			
R3015 10 meg R3025 200 Ω R3035 300 k R3045 10 meg R3085 41.5 k	2 w Var. ½ w Fixed	Comp. Comp. Prec. Comp. Prec.	10% 20% DIFF. BAL. 1% 10% 1%	302106 311063 309125 302106 310512		
R3095 22 me; R3155 41.5 k R3165 22 me; R3205 270 Ω R3215 56.5 k	5 w Fixed	Comp. Prec. Comp. Comp. Prec.	10% 1% 10% 10% 1%	302226 310512 302226 302271 309040		
R3225 2 k R3235 10 k R3245 56.5 k R3265 500 Ω R3295 27 k	2 w Var. 2 w Var. ½ w Fixed 2 w Var. ½ w Fixed	Comp. Comp. Prec. Comp. Comp.	20% GAIN ADJUST 20% VARIABLE 1% 20% VAR. BAL. 10%	311008 311016 309040 311005 302273		
R3405 80 k R3415 26.67 k R3425 8.89 k R3435 4.21 k R3445 2.05 k	½ w Fixed ½ w Fixed ½ w Fixed ½ w Fixed ½ w Fixed	Prec. Prec. Prec. Prec. Prec.	1% 1% 1% 1% 1%	309108 309107 309106 309105 309104		



					Resistors	(Continu	ed)			Order I by Nu	
	R3455 R3465 R3505 R3515 R3605	402 Ω 41.5 k 41.5 k	1/2 W 1/2 W 5 W 5 W 1/2 W	Fixed Fixed Fixed Fixed Fixed	Prec. Prec. Prec. Prec. Prec.	1% 1% 1% 1% 1%	,			30 31 3 3	09103 09102 10512 10512 09087
	R3615 R3705 R3715 R3755 R3805	5 meg 2 meg	½ w ½ w ½ w ½ w ½ w 2 w	Fixed Fixed Fixed Fixed Var.	Prec. Prec. Prec. Comp. Comp.	1% 1% 1% 10% 20%	Vert. Pe	os.		3(3(3)	090 23 09087 0 9023 02226 11018
	R3815 R3825 R3835 R3845 R3855	56.5 k 56.5 k 20 k	½ W ½ W ½ W ½ W 2 W	Fixed Fixed Fixed Var. Fixed	Comp. Prec. Prec. Comp. Comp.	10% 1% 1% 20% 10%	Ver. P	os.	Range	30 30 3	02335 09040 09040 11018 02105
	R3865 R3875 R3905 R3915 R4005	240 k 82 k	1/2 W 1/2 W 1/2 W 1/2 W 1/2 W	Fixed Fixed Fixed Fixed Fixed	Comp. Comp. Comp. WW	5% 5% 10% 10% 5%				30 30 30	01244 01244 02823 02823 08061
*	R4015 R4035 R4055 R4065 R4075	330 Ω	5 w 5 w ½ w ½ w ½ w	Fixed Fixed Fixed Fixed Var.	WW WW Comp. Comp.	5% 5% 10% 10% 20%	Heater 1	Bal.		30 30 30	08066 08045 02331 02331 11091
1					Swi	tches					
]	SW3205 SW3305* SW3755 SW3805*	2 wafer 3 wafer double 1 2 wafer	5 pole si	position position ngle thro position	rotary rotary w push bu rotary	LI itton TI	V/CM F RESPO RACE R F RESP	EST	ORER	260123 26 260102 26 260135 260102 26	62088 —
J					Vacuui	n Tubes					
	V3015 V3205 V3265 V3805 V3905	12AX7 5879 5879 12AU7 12AU7	Gai Gai Pos		l Stage	} Check	ked		- ·	1.	54043 57052 54041 54041
J	V4005 V4015	OB2 OB2		tage Reg tage Reg					* * * * * * * * * * * * * * * * * * *		54075 540 7 5

^{*}SW3305 and SW3805 shafts are concentric. Furnished as a unit.





G

TYPE E LOW LEVEL DIFFERENTIAL AC PREAMP

ABBREVIATIONS USED IN OUR PARTS LISTS

Cer.	ceramic	m	milli	
Comp.	composition	Ω	ohm	
EMC	electrolytic, metal cased	Poly.	polystyrene	
EMT	electrolytic, metal tubular	Prec.	precision	
f	farad	PT	paper tubular	
h	henry	Tub.	tubular	
k	thousands of ohms	V	working volts dc	
meg	megohms	Var.	variable	
μ	micro	w	watt	
μμ	micromicro	WW	wire wound	
	GMV guaranteed minim	d minimum value		

ABBREVIATIONS USED IN OUR CIRCUIT DIAGRAMS

Resistance values are in ohms. The symbol k stands for thousands. A resistor marked 2.7 k has a resistance of 2,700 ohms. The symbol M stands for million. For example, a resistor marked 5.6 M has a resistance of 5.6 megohms.

Unless otherwise specified on the circuit diagram, capacitance values marked with the number 1 and numbers greater than 1 are in $\mu\mu$ f. For example, a capacitor marked 3.3 would have a capacitance of 3.3 micromicrofgrads. Capacitance values marked with a number less than 1 are in μ f. For example, a capacitor marked .47 would have a capacitance of .47 microfarads.

Inductance values marked in mh are in millihenrys. Inductance values marked in μh are in microhenrys.

Your instrument WARRANTY appears on the reverse side of this sheet.

SERIAL NO. 3421

IMPORTANT

Include the INSTRUMENT TYPE and the above SERIAL NUMBER in any correspondence regarding this instrument. The above serial number must match the instrument serial number if parts are to be ordered from the manual. Your help in this will enable us to answer your questions or fill your order with the least delay possible.



WARRANTY

All Tektronix instruments are fully guaranteed against defective materials and workmanship for one year. Should replacement parts be required, whether at no charge under warranty or at established net prices, notify us promptly, including sufficient details to identify the required parts. We will ship them prepaid (via air if requested) as soon as possible, usually within 24 hours.

Tektronix transformers, manufactured in our own plant, carry an indefinite warranty.

All price revision and design modification privileges reserved.